

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** AECO02  
**Lab Code:** ACE **SDG NO.:** Q3363  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

**Client Sample No.:** CCAL05 **Date Analyzed:** 10/17/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP075862.D **Time Analyzed:** 20:38

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                      |        | FROM      | TO     |                    |                   |       |
| Aroclor-1016-1       | 5.803  | 5.708     | 5.908  | 514.790            | 500.000           | 3.0   |
| Aroclor-1016-2       | 5.825  | 5.729     | 5.929  | 517.700            | 500.000           | 3.5   |
| Aroclor-1016-3       | 5.888  | 5.792     | 5.992  | 501.680            | 500.000           | 0.3   |
| Aroclor-1016-4       | 5.985  | 5.889     | 6.089  | 536.970            | 500.000           | 7.4   |
| Aroclor-1016-5       | 6.277  | 6.181     | 6.381  | 525.610            | 500.000           | 5.1   |
| Aroclor-1260-1       | 7.395  | 7.299     | 7.499  | 489.380            | 500.000           | -2.1  |
| Aroclor-1260-2       | 7.648  | 7.551     | 7.751  | 467.900            | 500.000           | -6.4  |
| Aroclor-1260-3       | 8.005  | 7.909     | 8.109  | 513.780            | 500.000           | 2.8   |
| Aroclor-1260-4       | 8.234  | 8.137     | 8.337  | 510.550            | 500.000           | 2.1   |
| Aroclor-1260-5       | 8.559  | 8.464     | 8.664  | 519.560            | 500.000           | 3.9   |
| Decachlorobiphenyl   | 10.430 | 10.334    | 10.534 | 42.350             | 50.000            | -15.3 |
| Tetrachloro-m-xylene | 4.652  | 4.555     | 4.755  | 48.990             | 50.000            | -2.0  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** AECO02  
**Lab Code:** ACE **SDG NO.:** Q3363  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

**Client Sample No.:** CCAL05 **Date Analyzed:** 10/17/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP075862.D **Time Analyzed:** 20:38

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.885 | 4.793     | 4.993 | 590.830            | 500.000           | 18.2  |
| Aroclor-1016-2       | 4.942 | 4.851     | 5.051 | 541.470            | 500.000           | 8.3   |
| Aroclor-1016-3       | 5.063 | 4.971     | 5.171 | 568.040            | 500.000           | 13.6  |
| Aroclor-1016-4       | 5.103 | 5.011     | 5.211 | 592.080            | 500.000           | 18.4  |
| Aroclor-1016-5       | 5.317 | 5.225     | 5.425 | 542.830            | 500.000           | 8.6   |
| Aroclor-1260-1       | 6.345 | 6.255     | 6.455 | 420.570            | 500.000           | -15.9 |
| Aroclor-1260-2       | 6.532 | 6.441     | 6.641 | 450.650            | 500.000           | -9.9  |
| Aroclor-1260-3       | 6.685 | 6.595     | 6.795 | 453.010            | 500.000           | -9.4  |
| Aroclor-1260-4       | 7.154 | 7.065     | 7.265 | 405.010            | 500.000           | -19.0 |
| Aroclor-1260-5       | 7.394 | 7.304     | 7.504 | 427.800            | 500.000           | -14.4 |
| Decachlorobiphenyl   | 8.790 | 8.703     | 8.903 | 44.620             | 50.000            | -10.8 |
| Tetrachloro-m-xylene | 3.791 | 3.696     | 3.896 | 49.220             | 50.000            | -1.6  |