

Surrogate Summary

SDG No.: Q3363

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114312.D	PIBLK-PO114312.D	Tetrachloro-m-xyl	1	20	18.6	93		60	140
		Decachlorobiphen	1	20	19.2	96		60	140
		Tetrachloro-m-xyl	2	20	18.9	94		60	140
		Decachlorobiphen	2	20	19.0	95		60	140
I.BLK-PO114452.D	PIBLK-PO114452.D	Tetrachloro-m-xyl	1	20	21.7	108		60	140
		Decachlorobiphen	1	20	20.5	103		60	140
		Tetrachloro-m-xyl	2	20	21.2	106		60	140
		Decachlorobiphen	2	20	20.1	101		60	140
PB170141BL	PB170141BL	Tetrachloro-m-xyl	1	20	23.9	119		32	144
		Decachlorobiphen	1	20	22.4	112		32	175
		Tetrachloro-m-xyl	2	20	23.0	115		32	144
		Decachlorobiphen	2	20	21.5	108		32	175
PB170141BS	PB170141BS	Tetrachloro-m-xyl	1	20	24.6	123		32	144
		Decachlorobiphen	1	20	23.4	117		32	175
		Tetrachloro-m-xyl	2	20	22.5	112		32	144
		Decachlorobiphen	2	20	22.5	113		32	175
Q3363-01	PR132-SO1-085011-20251015	Tetrachloro-m-xyl	1	20	21.2	106		32	144
		Decachlorobiphen	1	20	20.1	101		32	175
		Tetrachloro-m-xyl	2	20	23.7	119		32	144
		Decachlorobiphen	2	20	18.9	94		32	175
Q3363-03	PR132-SO3-028010-20251015	Tetrachloro-m-xyl	1	20	18.8	94		32	144
		Decachlorobiphen	1	20	16.4	82		32	175
		Tetrachloro-m-xyl	2	20	19.5	98		32	144
		Decachlorobiphen	2	20	15.2	76		32	175
Q3369-01MS	MH-22MS	Tetrachloro-m-xyl	1	20	20.8	104		32	144
		Decachlorobiphen	1	20	16.4	82		32	175
		Tetrachloro-m-xyl	2	20	20.7	104		32	144
		Decachlorobiphen	2	20	14.9	74		32	175
Q3369-01MSD	MH-22MSD	Tetrachloro-m-xyl	1	20	20.0	100		32	144
		Decachlorobiphen	1	20	16.4	82		32	175
		Tetrachloro-m-xyl	2	20	19.5	97		32	144
		Decachlorobiphen	2	20	14.3	71		32	175
I.BLK-PO114465.D	PIBLK-PO114465.D	Tetrachloro-m-xyl	1	20	21.8	109		60	140
		Decachlorobiphen	1	20	20.6	103		60	140
		Tetrachloro-m-xyl	2	20	21.8	109		60	140
		Decachlorobiphen	2	20	19.5	97		60	140
I.BLK-PP075246.D	PIBLK-PP075246.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	20.0	100		60	140
I.BLK-PP075839.D	PIBLK-PP075839.D	Tetrachloro-m-xyl	1	20	21.9	109		60	140

Surrogate Summary

SDG No.: Q3363

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075839.D	PIBLK-PP075839.D	Decachlorobiphen	1	20	20.4	102		60	140
		Tetrachloro-m-xyl	2	20	19.0	95		60	140
Q3363-04	PR132-SO3-010013-20251015	Decachlorobiphen	2	20	21.4	107		60	140
		Tetrachloro-m-xyl	1	20	18.9	95		32	144
		Decachlorobiphen	1	20	14.9	74		32	175
		Tetrachloro-m-xyl	2	20	22.2	111		32	144
		Decachlorobiphen	2	20	15.4	77		32	175
Q3363-05	COMP01-20251015	Tetrachloro-m-xyl	1	20	20.9	104		32	144
		Decachlorobiphen	1	20	11.4	57		32	175
		Tetrachloro-m-xyl	2	20	21.8	109		32	144
		Decachlorobiphen	2	20	11.5	58		32	175
I.BLK-PP075854.D	PIBLK-PP075854.D	Tetrachloro-m-xyl	1	20	21.3	107		60	140
		Decachlorobiphen	1	20	19.1	96		60	140
		Tetrachloro-m-xyl	2	20	18.5	93		60	140
		Decachlorobiphen	2	20	20.1	100		60	140
Q3363-02	PR132-SO1-039085-20251015	Tetrachloro-m-xyl	1	20	13.2	66		32	144
		Decachlorobiphen	1	20	11.1	56		32	175
		Tetrachloro-m-xyl	2	20	22.9	114		32	144
		Decachlorobiphen	2	20	9.25	46		32	175
I.BLK-PP075866.D	PIBLK-PP075866.D	Tetrachloro-m-xyl	1	20	20.9	105		60	140
		Decachlorobiphen	1	20	18.9	95		60	140
		Tetrachloro-m-xyl	2	20	18.1	91		60	140
		Decachlorobiphen	2	20	17.9	90		60	140