

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090670.D **Time Analyzed:** 20:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	53.400	50.000	6.8
4,4'-DDE	6.191	6.090	6.290	53.210	50.000	6.4
4,4'-DDT	7.016	6.915	7.115	52.590	50.000	5.2
Aldrin	5.264	5.164	5.364	53.680	50.000	7.4
alpha-BHC	3.994	3.894	4.094	54.470	50.000	8.9
alpha-Chlordane	6.021	5.921	6.121	51.170	50.000	2.3
beta-BHC	4.512	4.412	4.612	51.910	50.000	3.8
Decachlorobiphenyl	9.066	8.964	9.164	51.560	50.000	3.1
delta-BHC	4.760	4.660	4.860	53.950	50.000	7.9
Dieldrin	6.341	6.240	6.440	53.290	50.000	6.6
Endosulfan I	6.068	5.968	6.168	52.600	50.000	5.2
Endosulfan II	6.781	6.681	6.881	53.170	50.000	6.3
Endosulfan sulfate	7.145	7.045	7.245	52.300	50.000	4.6
Endrin	6.568	6.468	6.668	53.340	50.000	6.7
Endrin aldehyde	6.910	6.810	7.010	52.750	50.000	5.5
Endrin ketone	7.625	7.525	7.725	52.370	50.000	4.7
gamma-BHC (Lindane)	4.324	4.225	4.425	53.930	50.000	7.9
gamma-Chlordane	5.940	5.840	6.040	52.910	50.000	5.8
Heptachlor	4.923	4.823	5.023	51.150	50.000	2.3
Heptachlor epoxide	5.685	5.584	5.784	53.260	50.000	6.5
Methoxychlor	7.489	7.388	7.588	50.870	50.000	1.7
Tetrachloro-m-xylene	3.544	3.445	3.645	51.790	50.000	3.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090670.D **Time Analyzed:** 20:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.921	5.821	6.021	52.390	50.000	4.8
4,4'-DDE	5.365	5.266	5.466	52.220	50.000	4.4
4,4'-DDT	6.174	6.074	6.274	54.440	50.000	8.9
Aldrin	4.360	4.261	4.461	52.220	50.000	4.4
alpha-BHC	3.386	3.287	3.487	52.250	50.000	4.5
alpha-Chlordane	5.181	5.081	5.281	52.420	50.000	4.8
beta-BHC	4.018	3.919	4.119	51.770	50.000	3.5
Decachlorobiphenyl	8.062	7.962	8.162	53.450	50.000	6.9
delta-BHC	4.255	4.155	4.355	52.490	50.000	5.0
Dieldrin	5.503	5.404	5.604	52.550	50.000	5.1
Endosulfan I	5.238	5.138	5.338	52.240	50.000	4.5
Endosulfan II	6.072	5.972	6.172	52.390	50.000	4.8
Endosulfan sulfate	6.473	6.374	6.574	52.790	50.000	5.6
Endrin	5.780	5.681	5.881	52.830	50.000	5.7
Endrin aldehyde	6.250	6.150	6.350	53.080	50.000	6.2
Endrin ketone	6.983	6.883	7.083	52.900	50.000	5.8
gamma-BHC (Lindane)	3.722	3.623	3.823	52.320	50.000	4.6
gamma-Chlordane	5.116	5.017	5.217	52.450	50.000	4.9
Heptachlor	4.075	3.975	4.175	52.480	50.000	5.0
Heptachlor epoxide	4.863	4.764	4.964	51.240	50.000	2.5
Methoxychlor	6.745	6.645	6.845	53.840	50.000	7.7
Tetrachloro-m-xylene	2.874	2.775	2.975	51.420	50.000	2.8