

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.066	8.970	9.170	22.860	20.000	14.3
Tetrachloro-m-xylene	3.544	3.490	3.590	21.710	20.000	8.6
alpha-BHC	3.994	3.940	4.040	9.050	10.000	-9.5
beta-BHC	4.512	4.460	4.560	10.070	10.000	0.7
gamma-BHC (Lindane)	4.324	4.270	4.370	9.330	10.000	-6.7
Endrin	6.568	6.500	6.640	49.200	50.000	-1.6
4,4'-DDT	7.016	6.950	7.090	100.270	100.000	0.3
Methoxychlor	7.489	7.420	7.560	236.810	250.000	-5.3

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.062	7.960	8.160	21.400	20.000	7.0
Tetrachloro-m-xylene	2.875	2.820	2.930	21.220	20.000	6.1
alpha-BHC	3.386	3.340	3.440	10.300	10.000	3.0
beta-BHC	4.019	3.970	4.070	10.460	10.000	4.6
gamma-BHC (Lindane)	3.723	3.670	3.770	10.340	10.000	3.4
Endrin	5.781	5.710	5.850	47.940	50.000	-4.1
4,4'-DDT	6.174	6.100	6.240	95.670	100.000	-4.3
Methoxychlor	6.745	6.670	6.820	201.830	250.000	-19.3