

DATA PACKAGE
GC SEMI-VOLATILES

PROJECT NAME : AE1-CTY 3.2.Z-PR-DOCK-DES-ODC - 60693795-1719206

AECOM
605 3rd Avenue
29th Floor
New York, NY - 10158
Phone No: 212-973-2900

ORDER ID : Q3363
ATTENTION : Rob Forstner



Laboratory Certification ID # 20012



1) PESTICIDE-TCL Data	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For Pesticide-TCL	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) MS/MSD Summary	16
9.3) LCS/LCSD Summary	20
9.4) Method Blank Summary	21
10) Sample Data	22
10.1) PR132-SO1-085011-20251015	23
10.2) PR132-SO1-039085-20251015	27
10.3) PR132-SO3-028010-20251015	34
10.4) PR132-SO3-010013-20251015	41
10.5) COMP01-20251015	45
11) Calibration Data Summary	49
11.1) Initial Calibration Data	50
11.1.1) PD101725	50
11.2) Continued Calibration Data	152
11.2.1) PD090670.D	152
11.2.2) PD090688.D	170
11.2.3) PEM Files	188
11.3) RESCHK Data	208
11.4) Analytical Seq	216
12) Compound Detection Summary	218
13) QC Sample Data	226
13.1) Method Blank Data	227
13.2) PIBLK Data	231
13.3) LCS Data	243
13.4) MS Data	258
13.5) MSD Data	273

Table Of Contents for Q3363

14) Manual Integration	288
15) Analytical Runlogs	291
16) Percent Solid	296
17) Extraction Logs	300
17.1) PB170142.pdf	300
17.2) PB170142IC.pdf	302
18) Standard Prep Logs	303
19) Shipping Document	357
19.1) Chain Of Custody	358
19.2) Lab Certificate	359
19.3) Internal COC	360

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Cover Page

Order ID : Q3363

Project ID : AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

Client : AECOM

Lab Sample Number

Q3363-01
Q3363-02
Q3363-03
Q3363-04
Q3363-05

Client Sample Number

PR132-SO1-085011-20251015
PR132-SO1-039085-20251015
PR132-SO3-028010-20251015
PR132-SO3-010013-20251015
COMP01-20251015

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/28/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

AECOM

Project Name: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

Project # N/A

Order ID # Q3363

Test Name: Pesticide-TCL

A. Number of Samples and Date of Receipt:

5 Solid samples were received on 10/15/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Pesticide-TCL. This data package contains results for Pesticide-TCL.

C. Analytical Techniques:

The analysis was performed on instrument ECD_D. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df,; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of Pesticide-TCLs was based on method 8081B and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the requirements for all compounds.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3363

MATRIX: Solid

METHOD: 8081B/3541

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified.			✓
2.	Standard Summary Submitted.			✓
3.	Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD.			✓
	The Initial Calibration met the requirements.			
	The Continuous Calibration met the requirements.			
4.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5.	Surrogate Recoveries Meet Criteria			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
6.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The MS recoveries met the requirements for all compounds.			
	The MSD recoveries met the requirements for all compounds.			
	The Blank Spike met requirements for all compounds.			
	The RPD were met for all analysis.			
7.	Retention Time Shift Meet Criteria (if applicable)			✓
	Comments:			
8.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
9.	Analysis Holding Time Met			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

The soil samples results are based on a dry weight basis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3363

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/28/2025

LAB CHRONICLE

OrderID:	Q3363	OrderDate:	10/16/2025 10:56:42 AM
Client:	AECOM	Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC -
Contact:	Rob Forstner	Location:	DOCK-Des-ODC - Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3363-01	PR132-SO1-085011-2 0251015	SOIL			10/15/25			10/15/25
			Pesticide-TCL	8081B		10/17/25	10/17/25	
Q3363-02	PR132-SO1-039085-2 0251015	SOIL			10/15/25			10/15/25
			Pesticide-TCL	8081B		10/17/25	10/17/25	
Q3363-03	PR132-SO3-028010-2 0251015	SOIL			10/15/25			10/15/25
			Pesticide-TCL	8081B		10/17/25	10/17/25	
Q3363-04	PR132-SO3-010013-2 0251015	SOIL			10/15/25			10/15/25
			Pesticide-TCL	8081B		10/17/25	10/17/25	
Q3363-05	COMP01-20251015	SOIL			10/15/25			10/15/25
			Pesticide-TCL	8081B		10/17/25	10/18/25	

Hit Summary Sheet
SW-846

SDG No.: Q3363

Order ID: Q3363

Client: AECOM

Project ID: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : PR132-SO1-039085-20251015								
Q3363-02	PR132-SO1-039085-2	SOIL	4,4-DDE	9.70	P	0.25	3.00	ug/kg
Q3363-02	PR132-SO1-039085-2	SOIL	Endrin	1.40	JP	0.25	3.00	ug/kg
Q3363-02	PR132-SO1-039085-2	SOIL	4,4-DDD	5.70		0.26	3.00	ug/kg
Q3363-02	PR132-SO1-039085-2	SOIL	Endosulfan Sulfate	0.62	JP	0.23	3.00	ug/kg
Q3363-02	PR132-SO1-039085-2	SOIL	alpha-Chlordane	0.51	J	0.21	3.00	ug/kg
Total Concentration:				17.930				
Client ID : PR132-SO3-028010-20251015								
Q3363-03	PR132-SO3-028010-2	SOIL	Dieldrin	0.95	JP	0.25	3.10	ug/kg
Q3363-03	PR132-SO3-028010-2	SOIL	4,4-DDE	4.40		0.25	3.10	ug/kg
Q3363-03	PR132-SO3-028010-2	SOIL	Endrin	0.47	JP	0.25	3.10	ug/kg
Q3363-03	PR132-SO3-028010-2	SOIL	4,4-DDD	6.80		0.27	3.10	ug/kg
Q3363-03	PR132-SO3-028010-2	SOIL	alpha-Chlordane	0.59	J	0.22	3.10	ug/kg
Q3363-03	PR132-SO3-028010-2	SOIL	gamma-Chlordane	1.20	JP	0.27	3.10	ug/kg
Total Concentration:				14.410				



QC SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Surrogate Summary

SDG No.: Q3363

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		57	171
		Tetrachloro-m-xyl	1	20	19.4	97		61	148
		Decachlorobiphen	2	20	20.3	102		57	171
		Tetrachloro-m-xyl	2	20	20.0	100		61	148
I.BLK-PD090668.D	PIBLK-PD090668.D	Decachlorobiphen	1	20	22.8	114		57	171
		Tetrachloro-m-xyl	1	20	20.4	102		61	148
		Decachlorobiphen	2	20	23.1	116		57	171
		Tetrachloro-m-xyl	2	20	21.1	106		61	148
PB170142BL	PB170142BL	Decachlorobiphen	1	20	20.5	103		20	144
		Tetrachloro-m-xyl	1	20	18.6	93		19	148
		Decachlorobiphen	2	20	20.5	103		20	144
		Tetrachloro-m-xyl	2	20	19.0	95		19	148
PB170142BS	PB170142BS	Decachlorobiphen	1	20	20.0	100		20	144
		Tetrachloro-m-xyl	1	20	18.2	91		19	148
		Decachlorobiphen	2	20	20.6	103		20	144
		Tetrachloro-m-xyl	2	20	19.0	95		19	148
Q3363-01	PR132-SO1-085011-20251015	Decachlorobiphen	1	20	16.6	83		20	144
		Tetrachloro-m-xyl	1	20	16.1	81		19	148
		Decachlorobiphen	2	20	11.5	57		20	144
		Tetrachloro-m-xyl	2	20	16.1	81		19	148
Q3363-02	PR132-SO1-039085-20251015	Decachlorobiphen	1	20	6.91	35		20	144
		Tetrachloro-m-xyl	1	20	6.63	33		19	148
		Decachlorobiphen	2	20	5.08	25		20	144
		Tetrachloro-m-xyl	2	20	6.57	33		19	148
Q3363-03	PR132-SO3-028010-20251015	Decachlorobiphen	1	20	12.8	64		20	144
		Tetrachloro-m-xyl	1	20	13.7	68		19	148
		Decachlorobiphen	2	20	9.99	50		20	144
		Tetrachloro-m-xyl	2	20	14.0	70		19	148
Q3363-04	PR132-SO3-010013-20251015	Decachlorobiphen	1	20	11.7	58		20	144
		Tetrachloro-m-xyl	1	20	11.9	60		19	148
		Decachlorobiphen	2	20	8.54	43		20	144
		Tetrachloro-m-xyl	2	20	12.6	63		19	148
Q3363-04MS	PR132-SO3-010013-20251015M	Decachlorobiphen	1	20	12.7	63		20	144
		Tetrachloro-m-xyl	1	20	13.0	65		19	148
		Decachlorobiphen	2	20	9.66	48		20	144
		Tetrachloro-m-xyl	2	20	13.7	69		19	148
Q3363-04MSD	PR132-SO3-010013-20251015M	Decachlorobiphen	1	20	13.2	66		20	144
		Tetrachloro-m-xyl	1	20	13.5	68		19	148
		Decachlorobiphen	2	20	9.75	49		20	144
		Tetrachloro-m-xyl	2	20	13.8	69		19	148
Q3363-05	COMP01-20251015	Decachlorobiphen	1	20	10.1	51		20	144

Surrogate Summary

SDG No.: Q3363

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3363-05	COMP01-20251015	Tetrachloro-m-xyl	1	20	14.7	73		19	148
		Decachlorobiphen	2	20	7.24	36		20	144
		Tetrachloro-m-xyl	2	20	15.1	75		19	148
I.BLK-PD090687.D	PIBLK-PD090687.D	Decachlorobiphen	1	20	20.7	103		57	171
		Tetrachloro-m-xyl	1	20	21.0	105		61	148
		Decachlorobiphen	2	20	17.8	89		57	171
		Tetrachloro-m-xyl	2	20	21.0	105		61	148

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3363</u>	Analytical Method:	<u>8081B</u>
Client:	<u>AECOM</u>	DataFile :	<u>PD090682.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result				Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3363-04MS (Column 1)	Client Sample ID:		PR132-SO3-010013-202510								
	alpha-BHC	32.3	0	29.0	ug/kg	90				60	144	
	beta-BHC	32.3	0	24.9	ug/kg	77				54	143	
	delta-BHC	32.3	0	29.0	ug/kg	90				29	151	
	gamma-BHC (Lindane)	32.3	0	28.5	ug/kg	88				61	140	
	Heptachlor	32.3	0	27.7	ug/kg	86				63	135	
	Aldrin	32.3	0	28.9	ug/kg	89				49	139	
	Heptachlor epoxide	32.3	0	28.0	ug/kg	87				41	156	
	Endosulfan I	32.3	0	27.5	ug/kg	85				56	142	
	Dieldrin	32.3	0	27.1	ug/kg	84				47	161	
	4,4'-DDE	32.3	0	25.3	ug/kg	78				55	136	
	Endrin	32.3	0	27.8	ug/kg	86				57	139	
	Endosulfan II	32.3	0	26.8	ug/kg	83				40	163	
	4,4'-DDD	32.3	0	27.0	ug/kg	84				47	163	
	Endosulfan sulfate	32.3	0	26.3	ug/kg	81				62	139	
	4,4'-DDT	32.3	0	27.1	ug/kg	84				51	146	
	Methoxychlor	32.3	0	26.9	ug/kg	83				54	136	
	Endrin ketone	32.3	0	26.8	ug/kg	83				60	129	
	Endrin aldehyde	32.3	0	27.0	ug/kg	84				59	132	
	alpha-Chlordane	32.3	0	26.6	ug/kg	82				39	166	
	gamma-Chlordane	32.3	0	27.1	ug/kg	84				44	175	
Lab Sample ID:	Q3363-04MS (Column 2)	Client Sample ID:		PR132-SO3-010013-202510								
	alpha-BHC	32.3	0	25.7	ug/kg	80				60	144	
	beta-BHC	32.3	0	24.7	ug/kg	76				54	143	
	delta-BHC	32.3	0	25.6	ug/kg	79				29	151	
	gamma-BHC (Lindane)	32.3	0	25.5	ug/kg	79				61	140	
	Heptachlor	32.3	0	26.3	ug/kg	81				63	135	
	Aldrin	32.3	0	25.2	ug/kg	78				49	139	
	Heptachlor epoxide	32.3	0	25.8	ug/kg	80				41	156	
	Endosulfan I	32.3	0	22.4	ug/kg	69				56	142	
	Dieldrin	32.3	0	25.3	ug/kg	78				47	161	
	4,4'-DDE	32.3	0	23.1	ug/kg	72				55	136	
	Endrin	32.3	0	24.5	ug/kg	76				57	139	
	Endosulfan II	32.3	0	24.1	ug/kg	75				40	163	
	4,4'-DDD	32.3	0	22.1	ug/kg	68				47	163	
	Endosulfan sulfate	32.3	0	23.4	ug/kg	72				62	139	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3363</u>	Analytical Method:	<u>8081B</u>
Client:	<u>AECOM</u>	DataFile :	<u>PD090682.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits		RPD
		Result	Result							High		
4,4'-DDT	32.3	0	23.9	ug/kg	74				51	146		
Methoxychlor	32.3	0	23.6	ug/kg	73				54	136		
Endrin ketone	32.3	0	21.7	ug/kg	67				60	129		
Endrin aldehyde	32.3	0	23.6	ug/kg	73				59	132		
alpha-Chlordane	32.3	0	25.2	ug/kg	78				39	166		
gamma-Chlordane	32.3	0	26.0	ug/kg	80				44	175		

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3363</u>	Analytical Method:	<u>8081B</u>
Client:	<u>AECOM</u>	DataFile :	<u>PD090683.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Low	Limits		RPD
			Result				Qual	RPD	Qual		High		
Lab Sample ID:	Q3363-04MSD (Column 1)	Client Sample ID:		PR132-SO3-010013-202510									
	alpha-BHC	32.28	0	30.2	ug/kg	94		4		60	144	20	
	beta-BHC	32.28	0	25.8	ug/kg	80		4		54	143	20	
	delta-BHC	32.28	0	30.0	ug/kg	93		3		29	151	20	
	gamma-BHC (Lindane)	32.28	0	29.6	ug/kg	92		4		61	140	20	
	Heptachlor	32.28	0	28.7	ug/kg	89		3		63	135	20	
	Aldrin	32.28	0	29.8	ug/kg	92		3		49	139	20	
	Heptachlor epoxide	32.28	0	29.0	ug/kg	90		3		41	156	20	
	Endosulfan I	32.28	0	28.8	ug/kg	89		5		56	142	20	
	Dieldrin	32.28	0	28.3	ug/kg	88		5		47	161	20	
	4,4'-DDE	32.28	0	26.6	ug/kg	82		5		55	136	20	
	Endrin	32.28	0	29.1	ug/kg	90		5		57	139	20	
	Endosulfan II	32.28	0	28.0	ug/kg	87		5		40	163	20	
	4,4'-DDD	32.28	0	28.4	ug/kg	88		5		47	163	20	
	Endosulfan sulfate	32.28	0	27.5	ug/kg	85		5		62	139	20	
	4,4'-DDT	32.28	0	28.2	ug/kg	87		4		51	146	20	
	Methoxychlor	32.28	0	27.7	ug/kg	86		4		54	136	20	
	Endrin ketone	32.28	0	28.4	ug/kg	88		6		60	129	20	
	Endrin aldehyde	32.28	0	28.2	ug/kg	87		4		59	132	20	
	alpha-Chlordane	32.28	0	27.6	ug/kg	86		5		39	166	20	
	gamma-Chlordane	32.28	0	28.3	ug/kg	88		5		44	175	20	
Lab Sample ID:	Q3363-04MSD (Column 2)	Client Sample ID:		PR132-SO3-010013-202510									
	alpha-BHC	32.28	0	26.0	ug/kg	81		1		60	144	20	
	beta-BHC	32.28	0	25.0	ug/kg	77		1		54	143	20	
	delta-BHC	32.28	0	25.8	ug/kg	80		1		29	151	20	
	gamma-BHC (Lindane)	32.28	0	25.7	ug/kg	80		1		61	140	20	
	Heptachlor	32.28	0	26.7	ug/kg	83		2		63	135	20	
	Aldrin	32.28	0	25.7	ug/kg	80		3		49	139	20	
	Heptachlor epoxide	32.28	0	25.6	ug/kg	79		1		41	156	20	
	Endosulfan I	32.28	0	22.8	ug/kg	71		3		56	142	20	
	Dieldrin	32.28	0	25.9	ug/kg	80		3		47	161	20	
	4,4'-DDE	32.28	0	23.5	ug/kg	73		1		55	136	20	
	Endrin	32.28	0	24.9	ug/kg	77		1		57	139	20	
	Endosulfan II	32.28	0	24.4	ug/kg	76		1		40	163	20	
	4,4'-DDD	32.28	0	22.4	ug/kg	69		1		47	163	20	
	Endosulfan sulfate	32.28	0	23.5	ug/kg	73		1		62	139	20	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3363</u>	Analytical Method:	<u>8081B</u>
Client:	<u>AECOM</u>	DataFile :	<u>PD090683.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
		Result	Result							High	RPD
4,4'-DDT	32.28	0	24.3	ug/kg	75		1		51	146	20
Methoxychlor	32.28	0	24.2	ug/kg	75		3		54	136	20
Endrin ketone	32.28	0	21.7	ug/kg	67		0		60	129	20
Endrin aldehyde	32.28	0	24.0	ug/kg	74		1		59	132	20
alpha-Chlordane	32.28	0	25.5	ug/kg	79		1		39	166	20
gamma-Chlordane	32.28	0	26.4	ug/kg	82		2		44	175	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3363</u>	Analytical Method:	<u>8081B</u>
Client:	<u>AECOM</u>	Datafile :	<u>PD090672.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170142BS (Column 1)	alpha-BHC	16.65	15.4	ug/kg	92				84	123	
	beta-BHC	16.65	15.0	ug/kg	90				82	123	
	delta-BHC	16.65	15.4	ug/kg	92				83	126	
	gamma-BHC (Lindane)	16.65	15.4	ug/kg	92				83	125	
	Heptachlor	16.65	14.7	ug/kg	88				83	122	
	Aldrin	16.65	15.6	ug/kg	94				82	124	
	Heptachlor epoxide	16.65	15.6	ug/kg	94				83	120	
	Endosulfan I	16.65	15.5	ug/kg	93				81	124	
	Dieldrin	16.65	15.6	ug/kg	94				85	121	
	4,4'-DDE	16.65	15.5	ug/kg	93				81	123	
	Endrin	16.65	15.7	ug/kg	94				76	130	
	Endosulfan II	16.65	15.6	ug/kg	94				80	125	
	4,4'-DDD	16.65	15.6	ug/kg	94				80	131	
	Endosulfan sulfate	16.65	15.4	ug/kg	92				81	122	
	4,4'-DDT	16.65	15.4	ug/kg	92				70	129	
	Methoxychlor	16.65	15.0	ug/kg	90				60	119	
	Endrin ketone	16.65	15.5	ug/kg	93				77	132	
	Endrin aldehyde	16.65	15.4	ug/kg	92				79	124	
	alpha-Chlordane	16.65	16.0	ug/kg	96				84	120	
	gamma-Chlordane	16.65	15.4	ug/kg	92				83	122	
PB170142BS (Column 2)	alpha-BHC	16.65	15.5	ug/kg	93				84	123	
	beta-BHC	16.65	15.5	ug/kg	93				82	123	
	delta-BHC	16.65	15.6	ug/kg	94				83	126	
	gamma-BHC (Lindane)	16.65	15.6	ug/kg	94				83	125	
	Heptachlor	16.65	15.7	ug/kg	94				83	122	
	Aldrin	16.65	15.7	ug/kg	94				82	124	
	Heptachlor epoxide	16.65	15.6	ug/kg	94				83	120	
	Endosulfan I	16.65	15.8	ug/kg	95				81	124	
	Dieldrin	16.65	15.8	ug/kg	95				85	121	
	4,4'-DDE	16.65	15.7	ug/kg	94				81	123	
	Endrin	16.65	16.0	ug/kg	96				76	130	
	Endosulfan II	16.65	15.8	ug/kg	95				80	125	
	4,4'-DDD	16.65	15.8	ug/kg	95				80	131	
	Endosulfan sulfate	16.65	15.9	ug/kg	95				81	122	
	4,4'-DDT	16.65	16.2	ug/kg	97				70	129	
	Methoxychlor	16.65	15.9	ug/kg	95				60	119	
	Endrin ketone	16.65	15.9	ug/kg	95				77	132	
	Endrin aldehyde	16.65	15.8	ug/kg	95				79	124	
	alpha-Chlordane	16.65	15.7	ug/kg	94				84	120	
	gamma-Chlordane	16.65	15.7	ug/kg	94				83	122	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170142BL

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: PB170142BL

Lab File ID: PD090671.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 10/17/2025

Date Analyzed (1): 10/17/2025

Date Analyzed (2): 10/17/2025

Time Analyzed (1): 20:31

Time Analyzed (2): 20:31

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB170142BS	PB170142BS	PD090672.D	10/17/2025	10/17/2025
PR132-S01-085011-20251015	Q3363-01	PD090678.D	10/17/2025	10/17/2025
PR132-S01-039085-20251015	Q3363-02	PD090679.D	10/17/2025	10/17/2025
PR132-S03-028010-20251015	Q3363-03	PD090680.D	10/17/2025	10/17/2025
PR132-S03-010013-20251015	Q3363-04	PD090681.D	10/17/2025	10/17/2025
PR132-S03-010013-20251015MS	Q3363-04MS	PD090682.D	10/17/2025	10/17/2025
PR132-S03-010013-20251015MSD	Q3363-04MSD	PD090683.D	10/17/2025	10/17/2025
COMP01-20251015	Q3363-05	PD090684.D	10/18/2025	10/18/2025

COMMENTS: _____



SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM		Date Collected:	10/15/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206		Date Received:	10/15/25
Client Sample ID:	PR132-SO1-085011-20251015		SDG No.:	Q3363
Lab Sample ID:	Q3363-01		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	50.1
Sample Wt/Vol:	30.06 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541B	Prep Date	10/17/25	Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	3.40	U	1	0.26	3.40	ug/kg	10/17/25 22:48	PB170142
319-85-7	beta-BHC	3.40	U	1	0.36	3.40	ug/kg	10/17/25 22:48	PB170142
319-86-8	delta-BHC	3.40	U	1	0.78	3.40	ug/kg	10/17/25 22:48	PB170142
58-89-9	gamma-BHC (Lindane)	3.40	U	1	0.28	3.40	ug/kg	10/17/25 22:48	PB170142
76-44-8	Heptachlor	3.40	U	1	0.24	3.40	ug/kg	10/17/25 22:48	PB170142
309-00-2	Aldrin	3.40	U	1	0.24	3.40	ug/kg	10/17/25 22:48	PB170142
1024-57-3	Heptachlor epoxide	3.40	U	1	0.38	3.40	ug/kg	10/17/25 22:48	PB170142
959-98-8	Endosulfan I	3.40	U	1	0.28	3.40	ug/kg	10/17/25 22:48	PB170142
60-57-1	Dieldrin	3.40	U	1	0.28	3.40	ug/kg	10/17/25 22:48	PB170142
72-55-9	4,4-DDE	3.40	U	1	0.28	3.40	ug/kg	10/17/25 22:48	PB170142
72-20-8	Endrin	3.40	U	1	0.28	3.40	ug/kg	10/17/25 22:48	PB170142
33213-65-9	Endosulfan II	3.40	U	1	0.58	3.40	ug/kg	10/17/25 22:48	PB170142
72-54-8	4,4-DDD	3.40	U	1	0.30	3.40	ug/kg	10/17/25 22:48	PB170142
1031-07-8	Endosulfan Sulfate	3.40	U	1	0.26	3.40	ug/kg	10/17/25 22:48	PB170142
50-29-3	4,4-DDT	3.40	U	1	0.28	3.40	ug/kg	10/17/25 22:48	PB170142
72-43-5	Methoxychlor	3.40	U	1	0.74	3.40	ug/kg	10/17/25 22:48	PB170142
53494-70-5	Endrin ketone	3.40	U	1	0.38	3.40	ug/kg	10/17/25 22:48	PB170142
7421-93-4	Endrin aldehyde	3.40	U	1	0.74	3.40	ug/kg	10/17/25 22:48	PB170142
5103-71-9	alpha-Chlordane	3.40	U	1	0.24	3.40	ug/kg	10/17/25 22:48	PB170142
5103-74-2	gamma-Chlordane	3.40	U	1	0.30	3.40	ug/kg	10/17/25 22:48	PB170142
8001-35-2	Toxaphene	65.7	U	1	10.8	65.7	ug/kg	10/17/25 22:48	PB170142
SURROGATES									
2051-24-3	Decachlorobiphenyl	16.6			20 - 144	83%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	16.1			19 - 148	81%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090678.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 22:48
Operator : AR\AJ
Sample : Q3363-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PR132-SO1-085011-20251015

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:32:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.543	2.874	50287033	479.5E6	16.148m	16.103
28) SA Decachlor...	9.066	8.061	77442561	347.1E6	16.555	11.459 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090678.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 22:48
Operator : AR\AJ
Sample : Q3363-01
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PR132-SO1-085011-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:32:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: -0.002 min
Response: 50287033
Conc: 16.15 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO1-085011-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 479505133
Conc: 16.10 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 77442561
Conc: 16.55 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 347097316
Conc: 11.46 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM		Date Collected:	10/15/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206		Date Received:	10/15/25
Client Sample ID:	PR132-SO1-039085-20251015		SDG No.:	Q3363
Lab Sample ID:	Q3363-02		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	56.6
Sample Wt/Vol:	30.04 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541B	Prep Date	10/17/25	Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	3.00	U	1	0.23	3.00	ug/kg	10/17/25 23:02	PB170142
319-85-7	beta-BHC	3.00	U	1	0.32	3.00	ug/kg	10/17/25 23:02	PB170142
319-86-8	delta-BHC	3.00	U	1	0.69	3.00	ug/kg	10/17/25 23:02	PB170142
58-89-9	gamma-BHC (Lindane)	3.00	U	1	0.25	3.00	ug/kg	10/17/25 23:02	PB170142
76-44-8	Heptachlor	3.00	U	1	0.21	3.00	ug/kg	10/17/25 23:02	PB170142
309-00-2	Aldrin	3.00	U	1	0.21	3.00	ug/kg	10/17/25 23:02	PB170142
1024-57-3	Heptachlor epoxide	3.00	U	1	0.34	3.00	ug/kg	10/17/25 23:02	PB170142
959-98-8	Endosulfan I	3.00	U	1	0.25	3.00	ug/kg	10/17/25 23:02	PB170142
60-57-1	Dieldrin	3.00	U	1	0.25	3.00	ug/kg	10/17/25 23:02	PB170142
72-55-9	4,4-DDE	9.70	P	1	0.25	3.00	ug/kg	10/17/25 23:02	PB170142
72-20-8	Endrin	1.40	JP	1	0.25	3.00	ug/kg	10/17/25 23:02	PB170142
33213-65-9	Endosulfan II	3.00	U	1	0.51	3.00	ug/kg	10/17/25 23:02	PB170142
72-54-8	4,4-DDD	5.70		1	0.26	3.00	ug/kg	10/17/25 23:02	PB170142
1031-07-8	Endosulfan Sulfate	0.62	JP	1	0.23	3.00	ug/kg	10/17/25 23:02	PB170142
50-29-3	4,4-DDT	3.00	U	1	0.25	3.00	ug/kg	10/17/25 23:02	PB170142
72-43-5	Methoxychlor	3.00	U	1	0.65	3.00	ug/kg	10/17/25 23:02	PB170142
53494-70-5	Endrin ketone	3.00	U	1	0.34	3.00	ug/kg	10/17/25 23:02	PB170142
7421-93-4	Endrin aldehyde	3.00	U	1	0.65	3.00	ug/kg	10/17/25 23:02	PB170142
5103-71-9	alpha-Chlordane	0.51	J	1	0.21	3.00	ug/kg	10/17/25 23:02	PB170142
5103-74-2	gamma-Chlordane	3.00	U	1	0.26	3.00	ug/kg	10/17/25 23:02	PB170142
8001-35-2	Toxaphene	58.2	U	1	9.50	58.2	ug/kg	10/17/25 23:02	PB170142
SURROGATES									
2051-24-3	Decachlorobiphenyl	6.91			20 - 144	35%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	6.63			19 - 148	33%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090679.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 23:02
 Operator : AR\AJ
 Sample : Q3363-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PR132-SO1-039085-20251015

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 10/21/2025
 Supervised By : mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:32:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.873	20655640	195.5E6	6.633	6.567m
2) SA Decachlor...	9.066	8.062	32338431	154.0E6	6.913	5.083 #
Target Compounds						
11) B alpha-Chl...	6.019	5.183	5192212	37446983	0.840	0.868m
12) B 4,4'-DDE	6.189	5.365	87802210	452.7E6	16.484	10.327 #
14) MA Endrin	6.567	5.779	11983717	23923050	2.419m	0.591m#
16) A 4,4'-DDD	6.699	5.920	39153955	239.0E6	9.658m	6.560 #
19) B Endosulfa...	7.134	6.471	3029335	38108319	0.653	1.063 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 23:02
Operator : AR\AJ
Sample : Q3363-02
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PR132-SO1-039085-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:32:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 20655640
Conc: 6.63 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO1-039085-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.002 min
Response: 195547540
Conc: 6.57 ng/ml m



#11 alpha-Chlordane

R.T.: 6.019 min
Delta R.T.: -0.001 min
Response: 5192212
Conc: 0.84 ng/ml



#11 alpha-Chlordane

R.T.: 5.183 min
Delta R.T.: 0.002 min
Response: 37446983
Conc: 0.87 ng/ml m

Response_ Signal: PD090679.D\ECD1A.ch

#12 4,4'-DDE

R.T.: 6.189 min
Delta R.T.: 0.000 min
Response: 87802210
Conc: 16.48 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO1-039085-20251015

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Time

Response_ Signal: PD090679.D\ECD2B.ch

#12 4,4'-DDE

R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 452716393
Conc: 10.33 ng/ml

Time

Response_ Signal: PD090679.D\ECD1A.ch

#14 Endrin

R.T.: 6.567 min
Delta R.T.: -0.001 min
Response: 11983717
Conc: 2.42 ng/ml m


Time

Response_ Signal: PD090679.D\ECD2B.ch

#14 Endrin

R.T.: 5.779 min
Delta R.T.: -0.001 min
Response: 23923050
Conc: 0.59 ng/ml m

Time




#16 4,4'-DDD

R.T.: 6.699 min
Delta R.T.: -0.001 min
Response: 39153955
Conc: 9.66 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO1-039085-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#16 4,4'-DDD

R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 239019753
Conc: 6.56 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.134 min
Delta R.T.: -0.010 min
Response: 3029335
Conc: 0.65 ng/ml



#19 Endosulfan Sulfate

R.T.: 6.471 min
Delta R.T.: -0.002 min
Response: 38108319
Conc: 1.06 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 32338431
Conc: 6.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO1-039085-20251015

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 153958665
Conc: 5.08 ng/ml

Report of Analysis

Client: AECOM	Date Collected: 10/15/25	
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received: 10/15/25	
Client Sample ID: PR132-SO3-028010-20251015	SDG No.: Q3363	
Lab Sample ID: Q3363-03	Matrix: SOIL	
Analytical Method: 8081B	% Solid: 54.8	
Sample Wt/Vol: 30.08 g	Test: Pesticide-TCL	Final Vol: 10000 uL
Prep Method: SW3541B	Prep Date: 10/17/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	3.10	U	1	0.24	3.10	ug/kg	10/17/25 23:15	PB170142
319-85-7	beta-BHC	3.10	U	1	0.33	3.10	ug/kg	10/17/25 23:15	PB170142
319-86-8	delta-BHC	3.10	U	1	0.71	3.10	ug/kg	10/17/25 23:15	PB170142
58-89-9	gamma-BHC (Lindane)	3.10	U	1	0.25	3.10	ug/kg	10/17/25 23:15	PB170142
76-44-8	Heptachlor	3.10	U	1	0.22	3.10	ug/kg	10/17/25 23:15	PB170142
309-00-2	Aldrin	3.10	U	1	0.22	3.10	ug/kg	10/17/25 23:15	PB170142
1024-57-3	Heptachlor epoxide	3.10	U	1	0.35	3.10	ug/kg	10/17/25 23:15	PB170142
959-98-8	Endosulfan I	3.10	U	1	0.25	3.10	ug/kg	10/17/25 23:15	PB170142
60-57-1	Dieldrin	0.95	JP	1	0.25	3.10	ug/kg	10/17/25 23:15	PB170142
72-55-9	4,4-DDE	4.40		1	0.25	3.10	ug/kg	10/17/25 23:15	PB170142
72-20-8	Endrin	0.47	JP	1	0.25	3.10	ug/kg	10/17/25 23:15	PB170142
33213-65-9	Endosulfan II	3.10	U	1	0.53	3.10	ug/kg	10/17/25 23:15	PB170142
72-54-8	4,4-DDD	6.80		1	0.27	3.10	ug/kg	10/17/25 23:15	PB170142
1031-07-8	Endosulfan Sulfate	3.10	U	1	0.24	3.10	ug/kg	10/17/25 23:15	PB170142
50-29-3	4,4-DDT	3.10	U	1	0.25	3.10	ug/kg	10/17/25 23:15	PB170142
72-43-5	Methoxychlor	3.10	U	1	0.67	3.10	ug/kg	10/17/25 23:15	PB170142
53494-70-5	Endrin ketone	3.10	U	1	0.35	3.10	ug/kg	10/17/25 23:15	PB170142
7421-93-4	Endrin aldehyde	3.10	U	1	0.67	3.10	ug/kg	10/17/25 23:15	PB170142
5103-71-9	alpha-Chlordane	0.59	J	1	0.22	3.10	ug/kg	10/17/25 23:15	PB170142
5103-74-2	gamma-Chlordane	1.20	JP	1	0.27	3.10	ug/kg	10/17/25 23:15	PB170142
8001-35-2	Toxaphene	60.1	U	1	9.80	60.1	ug/kg	10/17/25 23:15	PB170142
SURROGATES									
2051-24-3	Decachlorobiphenyl	12.8			20 - 144	64%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	14.0			19 - 148	70%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3363-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090680.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 23:15
 Operator : AR\AJ
 Sample : Q3363-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PR132-SO3-028010-20251015

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 10/21/2025
 Supervised By : mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:32:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.543	2.874	42520913	418.2E6	13.654m	14.043
28) SA Decachlor...	9.065	8.061	59985682	302.6E6	12.823	9.989
Target Compounds						
10) B gamma-Chl...	5.937	5.116	3656406	92238013	0.616m	2.052 #
11) B alpha-Chl...	6.015	5.182	5986982	38565056	0.969m	0.894m
12) B 4,4'-DDE	6.189	5.364	38421379	241.1E6	7.213	5.499
13) MA Dieldrin	6.339	5.501	3484951	69491918	0.598m	1.573 #
14) MA Endrin	6.567	5.779	3881126	20020062	0.783m	0.494m#
16) A 4,4'-DDD	6.698	5.919	45250337	367.9E6	11.161m	10.098m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090680.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 23:15
Operator : AR\AJ
Sample : Q3363-03
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PR132-SO3-028010-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:32:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: -0.002 min
Response: 42520913
Conc: 13.65 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-028010-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 418172098
Conc: 14.04 ng/ml




#10 gamma-Chlordane

R.T.: 5.937 min
Delta R.T.: -0.003 min
Response: 3656406
Conc: 0.62 ng/ml m



#10 gamma-Chlordane

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 92238013
Conc: 2.05 ng/ml




#11 alpha-Chlordane

R.T.: 6.015 min
Delta R.T.: -0.005 min
Response: 5986982
Conc: 0.97 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-028010-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#11 alpha-Chlordane

R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 38565056
Conc: 0.89 ng/ml m




#12 4,4'-DDE

R.T.: 6.189 min
Delta R.T.: 0.000 min
Response: 38421379
Conc: 7.21 ng/ml



#12 4,4'-DDE

R.T.: 5.364 min
Delta R.T.: -0.002 min
Response: 241055255
Conc: 5.50 ng/ml




#13 Dieldrin

R.T.: 6.339 min
Delta R.T.: -0.002 min
Response: 3484951
Conc: 0.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-028010-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#13 Dieldrin

R.T.: 5.501 min
Delta R.T.: -0.003 min
Response: 69491918
Conc: 1.57 ng/ml



#14 Endrin

R.T.: 6.567 min
Delta R.T.: -0.001 min
Response: 3881126
Conc: 0.78 ng/ml m



#14 Endrin

R.T.: 5.779 min
Delta R.T.: -0.002 min
Response: 20020062
Conc: 0.49 ng/ml m

Response_ Signal: PD090680.D\ECD1A.ch

#16 4,4'-DDD

R.T.: 6.698 min
Delta R.T.: -0.002 min
Response: 45250337
Conc: 11.16 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-028010-20251015

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Time Response_ Signal: PD090680.D\ECD2B.ch

#16 4,4'-DDD

R.T.: 5.919 min
Delta R.T.: -0.002 min
Response: 367949305
Conc: 10.10 ng/ml m

Time Response_ Signal: PD090680.D\ECD1A.ch

#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 59985682
Conc: 12.82 ng/ml

Time Response_ Signal: PD090680.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 302581500
Conc: 9.99 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM		Date Collected:	10/15/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206		Date Received:	10/15/25
Client Sample ID:	PR132-SO3-010013-20251015		SDG No.:	Q3363
Lab Sample ID:	Q3363-04		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	51.5
Sample Wt/Vol:	30.03 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541B	Prep Date	10/17/25	Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	3.30	U	1	0.25	3.30	ug/kg	10/17/25 23:29	PB170142
319-85-7	beta-BHC	3.30	U	1	0.35	3.30	ug/kg	10/17/25 23:29	PB170142
319-86-8	delta-BHC	3.30	U	1	0.76	3.30	ug/kg	10/17/25 23:29	PB170142
58-89-9	gamma-BHC (Lindane)	3.30	U	1	0.27	3.30	ug/kg	10/17/25 23:29	PB170142
76-44-8	Heptachlor	3.30	U	1	0.23	3.30	ug/kg	10/17/25 23:29	PB170142
309-00-2	Aldrin	3.30	U	1	0.23	3.30	ug/kg	10/17/25 23:29	PB170142
1024-57-3	Heptachlor epoxide	3.30	U	1	0.37	3.30	ug/kg	10/17/25 23:29	PB170142
959-98-8	Endosulfan I	3.30	U	1	0.27	3.30	ug/kg	10/17/25 23:29	PB170142
60-57-1	Dieldrin	3.30	U	1	0.27	3.30	ug/kg	10/17/25 23:29	PB170142
72-55-9	4,4-DDE	3.30	U	1	0.27	3.30	ug/kg	10/17/25 23:29	PB170142
72-20-8	Endrin	3.30	U	1	0.27	3.30	ug/kg	10/17/25 23:29	PB170142
33213-65-9	Endosulfan II	3.30	U	1	0.56	3.30	ug/kg	10/17/25 23:29	PB170142
72-54-8	4,4-DDD	3.30	U	1	0.29	3.30	ug/kg	10/17/25 23:29	PB170142
1031-07-8	Endosulfan Sulfate	3.30	U	1	0.25	3.30	ug/kg	10/17/25 23:29	PB170142
50-29-3	4,4-DDT	3.30	U	1	0.27	3.30	ug/kg	10/17/25 23:29	PB170142
72-43-5	Methoxychlor	3.30	U	1	0.72	3.30	ug/kg	10/17/25 23:29	PB170142
53494-70-5	Endrin ketone	3.30	U	1	0.37	3.30	ug/kg	10/17/25 23:29	PB170142
7421-93-4	Endrin aldehyde	3.30	U	1	0.72	3.30	ug/kg	10/17/25 23:29	PB170142
5103-71-9	alpha-Chlordane	3.30	U	1	0.23	3.30	ug/kg	10/17/25 23:29	PB170142
5103-74-2	gamma-Chlordane	3.30	U	1	0.29	3.30	ug/kg	10/17/25 23:29	PB170142
8001-35-2	Toxaphene	64.0	U	1	10.5	64.0	ug/kg	10/17/25 23:29	PB170142
SURROGATES									
2051-24-3	Decachlorobiphenyl	11.7			20 - 144	58%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	12.6			19 - 148	63%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3363-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090681.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 23:29
Operator : AR\AJ
Sample : Q3363-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:33:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.543	2.875	37113678	375.0E6	11.918m	12.595
28) SA Decachlor...	9.066	8.061	54539125	258.6E6	11.659	8.536 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090681.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 23:29
Operator : AR\AJ
Sample : Q3363-04
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:33:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: -0.002 min
Response: 37113678
Conc: 11.92 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 375037973
Conc: 12.59 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.001 min
Response: 54539125
Conc: 11.66 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 258558556
Conc: 8.54 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM		Date Collected:	10/15/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206		Date Received:	10/15/25
Client Sample ID:	COMP01-20251015		SDG No.:	Q3363
Lab Sample ID:	Q3363-05		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	55.4
Sample Wt/Vol:	30.05 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541B	Prep Date	10/17/25	Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	3.10	U	1	0.23	3.10	ug/kg	10/18/25 00:10	PB170142
319-85-7	beta-BHC	3.10	U	1	0.32	3.10	ug/kg	10/18/25 00:10	PB170142
319-86-8	delta-BHC	3.10	U	1	0.70	3.10	ug/kg	10/18/25 00:10	PB170142
58-89-9	gamma-BHC (Lindane)	3.10	U	1	0.25	3.10	ug/kg	10/18/25 00:10	PB170142
76-44-8	Heptachlor	3.10	U	1	0.22	3.10	ug/kg	10/18/25 00:10	PB170142
309-00-2	Aldrin	3.10	U	1	0.22	3.10	ug/kg	10/18/25 00:10	PB170142
1024-57-3	Heptachlor epoxide	3.10	U	1	0.34	3.10	ug/kg	10/18/25 00:10	PB170142
959-98-8	Endosulfan I	3.10	U	1	0.25	3.10	ug/kg	10/18/25 00:10	PB170142
60-57-1	Dieldrin	3.10	U	1	0.25	3.10	ug/kg	10/18/25 00:10	PB170142
72-55-9	4,4-DDE	3.10	U	1	0.25	3.10	ug/kg	10/18/25 00:10	PB170142
72-20-8	Endrin	3.10	U	1	0.25	3.10	ug/kg	10/18/25 00:10	PB170142
33213-65-9	Endosulfan II	3.10	U	1	0.52	3.10	ug/kg	10/18/25 00:10	PB170142
72-54-8	4,4-DDD	3.10	U	1	0.27	3.10	ug/kg	10/18/25 00:10	PB170142
1031-07-8	Endosulfan Sulfate	3.10	U	1	0.23	3.10	ug/kg	10/18/25 00:10	PB170142
50-29-3	4,4-DDT	3.10	U	1	0.25	3.10	ug/kg	10/18/25 00:10	PB170142
72-43-5	Methoxychlor	3.10	U	1	0.67	3.10	ug/kg	10/18/25 00:10	PB170142
53494-70-5	Endrin ketone	3.10	U	1	0.34	3.10	ug/kg	10/18/25 00:10	PB170142
7421-93-4	Endrin aldehyde	3.10	U	1	0.67	3.10	ug/kg	10/18/25 00:10	PB170142
5103-71-9	alpha-Chlordane	3.10	U	1	0.22	3.10	ug/kg	10/18/25 00:10	PB170142
5103-74-2	gamma-Chlordane	3.10	U	1	0.27	3.10	ug/kg	10/18/25 00:10	PB170142
8001-35-2	Toxaphene	59.5	U	1	9.70	59.5	ug/kg	10/18/25 00:10	PB170142
SURROGATES									
2051-24-3	Decachlorobiphenyl	10.1			20 - 144	51%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	15.1			19 - 148	75%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090684.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2025 00:10
 Operator : AR\AJ
 Sample : Q3363-05
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP01-20251015

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 10/21/2025
 Supervised By : mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:33:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.542	2.874	45685545	448.8E6	14.670m	15.071
28) SA Decachlor...	9.066	8.061	47371470	219.3E6	10.127	7.240 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090684.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2025 00:10
Operator : AR\AJ
Sample : Q3363-05
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
COMP01-20251015


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/21/2025
Supervised By : mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:33:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.542 min
Delta R.T.: -0.003 min
Response: 45685545
Conc: 14.67 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP01-20251015


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 448774364
Conc: 15.07 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 47371470
Conc: 10.13 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 219294910
Conc: 7.24 ng/ml



CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3363</u>
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):	<u>10/17/2025</u> <u>10/17/2025</u>
		Calibration Times:	<u>11:34</u> <u>12:28</u>

 GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PD090650.D</u>	RT 075 = <u>PD090651.D</u>
RT 050 = <u>PD090652.D</u>	RT 025 = <u>PD090653.D</u>	RT 005 = <u>PD090654.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	6.70	6.70	6.70	6.70	6.70	6.70	6.60	6.80
4,4'-DDE	6.19	6.19	6.19	6.19	6.19	6.19	6.09	6.29
4,4'-DDT	7.02	7.02	7.02	7.02	7.02	7.02	6.92	7.12
Aldrin	5.26	5.27	5.26	5.26	5.26	5.26	5.16	5.36
alpha-BHC	3.99	3.99	3.99	3.99	3.99	3.99	3.89	4.09
alpha-Chlordane	6.02	6.02	6.02	6.02	6.02	6.02	5.92	6.12
beta-BHC	4.51	4.51	4.51	4.51	4.51	4.51	4.41	4.61
Decachlorobiphenyl	9.07	9.07	9.06	9.07	9.07	9.07	8.97	9.17
delta-BHC	4.76	4.76	4.76	4.76	4.76	4.76	4.66	4.86
Dieldrin	6.34	6.34	6.34	6.34	6.34	6.34	6.24	6.44
Endosulfan I	6.07	6.07	6.07	6.07	6.07	6.07	5.97	6.17
Endosulfan II	6.78	6.78	6.78	6.78	6.78	6.78	6.68	6.88
Endosulfan sulfate	7.15	7.15	7.15	7.14	7.14	7.14	7.04	7.24
Endrin	6.57	6.57	6.57	6.57	6.57	6.57	6.47	6.67
Endrin aldehyde	6.91	6.91	6.91	6.91	6.91	6.91	6.81	7.01
Endrin ketone	7.63	7.63	7.63	7.63	7.62	7.62	7.52	7.72
gamma-BHC (Lindane)	4.32	4.33	4.33	4.32	4.32	4.32	4.22	4.42
gamma-Chlordane	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Heptachlor	4.92	4.92	4.92	4.92	4.92	4.92	4.82	5.02
Heptachlor epoxide	5.69	5.69	5.68	5.68	5.68	5.68	5.58	5.78
Methoxychlor	7.49	7.49	7.49	7.49	7.49	7.49	7.39	7.59
Tetrachloro-m-xylene	3.54	3.55	3.55	3.54	3.54	3.54	3.44	3.64

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3363</u>
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):	<u>10/17/2025</u> <u>10/17/2025</u>
		Calibration Times:	<u>11:34</u> <u>12:28</u>

 GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PD090650.D</u>	RT 075 = <u>PD090651.D</u>
RT 050 = <u>PD090652.D</u>	RT 025 = <u>PD090653.D</u>	RT 005 = <u>PD090654.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	5.92	5.92	5.92	5.92	5.92	5.92	5.82	6.02
4,4'-DDE	5.37	5.37	5.37	5.37	5.37	5.37	5.27	5.47
4,4'-DDT	6.17	6.18	6.17	6.17	6.18	6.17	6.07	6.27
Aldrin	4.36	4.36	4.36	4.36	4.36	4.36	4.26	4.46
alpha-BHC	3.39	3.39	3.39	3.39	3.39	3.39	3.29	3.49
alpha-Chlordane	5.18	5.18	5.18	5.18	5.18	5.18	5.08	5.28
beta-BHC	4.02	4.02	4.02	4.02	4.02	4.02	3.92	4.12
Decachlorobiphenyl	8.06	8.06	8.06	8.06	8.06	8.06	7.96	8.16
delta-BHC	4.26	4.26	4.26	4.26	4.26	4.26	4.16	4.36
Dieldrin	5.50	5.50	5.50	5.50	5.50	5.50	5.40	5.60
Endosulfan I	5.24	5.24	5.24	5.24	5.24	5.24	5.14	5.34
Endosulfan II	6.07	6.07	6.07	6.07	6.07	6.07	5.97	6.17
Endosulfan sulfate	6.47	6.47	6.47	6.47	6.47	6.47	6.37	6.57
Endrin	5.78	5.78	5.78	5.78	5.78	5.78	5.68	5.88
Endrin aldehyde	6.25	6.25	6.25	6.25	6.25	6.25	6.15	6.35
Endrin ketone	6.98	6.98	6.98	6.98	6.98	6.98	6.88	7.08
gamma-BHC (Lindane)	3.72	3.72	3.72	3.72	3.72	3.72	3.62	3.82
gamma-Chlordane	5.12	5.12	5.12	5.12	5.12	5.12	5.02	5.22
Heptachlor	4.08	4.08	4.08	4.07	4.08	4.07	3.97	4.17
Heptachlor epoxide	4.86	4.87	4.86	4.86	4.86	4.86	4.76	4.96
Methoxychlor	6.75	6.75	6.75	6.74	6.75	6.74	6.64	6.84
Tetrachloro-m-xylene	2.87	2.88	2.88	2.87	2.87	2.87	2.77	2.97

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: AECO02
SDG NO.: Q3363
Calibration Date(s): 10/17/2025 10/17/2025
Calibration Times: 11:34 12:28

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4150980000	4075270000	4055470000	3840760000	4148570000	4054210000	3
4,4'-DDE	5435610000	5370750000	5317610000	5053060000	5456160000	5326640000	3
4,4'-DDT	4256810000	4152600000	4117930000	3927880000	4097130000	4110470000	3
Aldrin	6694230000	6562230000	6516940000	6238140000	6724820000	6547270000	3
alpha-BHC	7752530000	7402380000	7252210000	6743070000	6735520000	7177140000	6
alpha-Chlordane	5966070000	5905470000	5938650000	6170460000	6920630000	6180250000	7
beta-BHC	2515210000	2489310000	2540370000	2560620000	2992650000	2619630000	8
Decachlorobiphenyl	4361180000	4315270000	4506230000	4707060000	5500070000	4677960000	10
delta-BHC	7169260000	6905720000	6771510000	6342450000	6619680000	6761720000	5
Dieldrin	5929890000	5841940000	5818180000	5550330000	6016400000	5831350000	3
Endosulfan I	5452270000	5404490000	5449050000	5348660000	6028770000	5536650000	5
Endosulfan II	4873230000	4835950000	4885080000	4789750000	5615800000	4999960000	7
Endosulfan sulfate	4557060000	4504850000	4556170000	4479960000	5092920000	4638190000	6
Endrin	5035940000	4962910000	4938250000	4711180000	5120260000	4953710000	3
Endrin aldehyde	3505880000	3507960000	3571680000	3579710000	4148280000	3662700000	7
Endrin ketone	4819770000	4736510000	4782320000	4695250000	5260400000	4858850000	5
gamma-BHC (Lindane)	7185880000	6905860000	6799650000	6417980000	6684750000	6798820000	4
gamma-Chlordane	5972450000	5894040000	5855640000	5592670000	6341090000	5931180000	5
Heptachlor	5757910000	5597510000	5545540000	5288860000	5757970000	5589560000	3
Heptachlor epoxide	5795570000	5708170000	5718680000	5584980000	6137920000	5789060000	4
Methoxychlor	2050060000	2015340000	2078640000	2094570000	2389310000	2125590000	7
Tetrachloro-m-xylene	3010860000	2951180000	3030100000	3124300000	3454390000	3114160000	6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: AECO02
SDG NO.: Q3363

Calibration Date(s): 10/17/2025 10/17/2025
Calibration Times: 11:34 12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	33708900000	34309400000	35560900000	36237500000	42373300000	36438000000	10
4,4'-DDE	40500800000	41191000000	42548600000	43577200000	51373700000	43838300000	10
4,4'-DDT	33888000000	33982600000	34868200000	34959100000	37667600000	35073100000	4
Aldrin	43083800000	43574800000	45033800000	46172400000	53729700000	46318900000	9
alpha-BHC	49118000000	48392000000	49683700000	50500300000	57784900000	51095800000	7
alpha-Chlordane	40004000000	40550200000	41893000000	42708900000	50532500000	43137700000	10
beta-BHC	18296700000	18365500000	19080800000	19804100000	23600800000	19829500000	11
Decachlorobiphenyl	28856800000	28247200000	29308000000	30057200000	34987600000	30291400000	9
delta-BHC	45324900000	45220200000	46466400000	47155800000	54416000000	47716700000	8
Dieldrin	40601100000	41425500000	43096200000	44203200000	51593900000	44184000000	10
Endosulfan I	35728600000	36586400000	38185700000	39478600000	46972400000	39390300000	11
Endosulfan II	34269700000	34968100000	36432200000	37520200000	44587200000	37555500000	11
Endosulfan sulfate	32963400000	33538900000	34960800000	35861000000	41900900000	35845000000	10
Endrin	36902000000	37827400000	39400900000	40693000000	47646500000	40494000000	11
Endrin aldehyde	25110900000	25739400000	26893100000	27767400000	32953200000	27692800000	11
Endrin ketone	34460000000	35020500000	36550700000	37816700000	44740900000	37717800000	11
gamma-BHC (Lindane)	44964800000	44604900000	45972800000	46808300000	53893500000	47248900000	8
gamma-Chlordane	41920800000	42422300000	43748100000	44346600000	52306700000	44948900000	9
Heptachlor	41272300000	41810600000	43239300000	44502900000	52189000000	44602800000	10
Heptachlor epoxide	38203400000	38891500000	40672900000	42374400000	53052500000	42639000000	14
Methoxychlor	16116900000	16316400000	17198900000	17735100000	19771700000	17427800000	8
Tetrachloro-m-xylene	27903400000	27689100000	28617400000	29726500000	34949400000	29777100000	10



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3363

Instrument ID: ECD_D Date(s) Analyzed: 10/17/2025 10/17/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	32564400
		2	6.44	6.34	6.54	48621800
		3	7.14	7.04	7.24	85710600
		4	7.56	7.46	7.66	95737400
		5	7.92	7.82	8.02	60480500



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3363

Instrument ID: ECD_D Date(s) Analyzed: 10/17/2025 10/17/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.47	5.37	5.57	263702000
		2	5.64	5.54	5.74	176560000
		3	6.75	6.65	6.85	658883000
		4	7.19	7.09	7.29	495104000
		5	7.32	7.22	7.42	246033000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090650.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:34
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:33:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.874	301.1E6	2790.3E6	99.682	98.737
28) SA Decachlor...	9.066	8.061	436.1E6	2885.7E6	98.364	99.224
Target Compounds						
2) A alpha-BHC	3.994	3.386	775.3E6	4911.8E6	103.334	99.427
3) MA gamma-BHC...	4.324	3.722	718.6E6	4496.5E6	102.762	98.892
4) MA Heptachlor	4.922	4.075	575.8E6	4127.2E6	101.879	97.673
5) MB Aldrin	5.264	4.360	669.4E6	4308.4E6	101.342	97.787
6) B beta-BHC	4.511	4.019	251.5E6	1829.7E6	99.502	97.902
7) B delta-BHC	4.760	4.255	716.9E6	4532.5E6	102.853	98.756
8) B Heptachlo...	5.685	4.864	579.6E6	3820.3E6	100.668	96.869
9) A Endosulfan I	6.068	5.238	545.2E6	3572.9E6	100.030	96.676
10) B gamma-Chl...	5.940	5.117	597.2E6	4192.1E6	100.988	97.867
11) B alpha-Chl...	6.020	5.181	596.6E6	4000.4E6	100.230	97.693
12) B 4,4'-DDE	6.190	5.366	543.6E6	4050.1E6	101.097	97.534
13) MA Dieldrin	6.341	5.504	593.0E6	4060.1E6	100.951	97.019
14) MA Endrin	6.568	5.780	503.6E6	3690.2E6	100.979	96.725
15) B Endosulfa...	6.781	6.072	487.3E6	3427.0E6	99.879	96.941
16) A 4,4'-DDD	6.700	5.920	415.1E6	3370.9E6	101.164	97.326
17) MA 4,4'-DDT	7.016	6.174	425.7E6	3388.8E6	101.658	98.574
18) B Endrin al...	6.910	6.250	350.6E6	2511.1E6	99.070	96.573
19) B Endosulfa...	7.145	6.474	455.7E6	3296.3E6	100.010	97.059
20) A Methoxychlor	7.489	6.745	205.0E6	1611.7E6	99.308	96.752
21) B Endrin ke...	7.625	6.982	482.0E6	3446.0E6	100.390	97.056
22) Mirex	8.107	7.175	295.5E6	2331.6E6	97.723	97.238



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090650.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:34
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:33:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090650.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 301086012
Conc: 99.68 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100


Time Response_ Signal: PD090650.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 2790335064
Conc: 98.74 ng/ml


Time Response_ Signal: PD090650.D\ECD1A.ch



#2 alpha-BHC


R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 775252906
Conc: 103.33 ng/ml

Time Response_ Signal: PD090650.D\ECD2B.ch



#2 alpha-BHC


R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 4911798337
Conc: 99.43 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 718588040
Conc: 102.76 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 4496483491
Conc: 98.89 ng/ml



#4 Heptachlor


R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 575791141
Conc: 101.88 ng/ml



#4 Heptachlor

R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 4127234386
Conc: 97.67 ng/ml

Response_ Signal: PD090650.D\ECD1A.ch




#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 669423227
Conc: 101.34 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100


Time Response_ Signal: PD090650.D\ECD2B.ch



#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 4308380411
Conc: 97.79 ng/ml


Time Response_ Signal: PD090650.D\ECD1A.ch



#6 beta-BHC


R.T.: 4.511 min
Delta R.T.: 0.000 min
Response: 251520811
Conc: 99.50 ng/ml

Time Response_ Signal: PD090650.D\ECD2B.ch



#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 1829668615
Conc: 97.90 ng/ml



R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 716926396
Conc: 102.85 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Time

Response_

Signal: PD090650.D\ECD2B.ch

#7 delta-BHC




R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 4532491918
Conc: 98.76 ng/ml

Time

Response_

Signal: PD090650.D\ECD1A.ch

#8 Heptachlor epoxide




R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 579557032
Conc: 100.67 ng/ml

Time


Response_

Signal: PD090650.D\ECD2B.ch

#8 Heptachlor epoxide




R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 3820342410
Conc: 96.87 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 545227033
Conc: 100.03 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 3572859410
Conc: 96.68 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 597245354
Conc: 100.99 ng/ml



#10 gamma-Chlordane


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 4192081714
Conc: 97.87 ng/ml



#11 alpha-Chlordane


R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 596606616
Conc: 100.23 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 4000403429
Conc: 97.69 ng/ml



#12 4,4'-DDE


R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 543560575
Conc: 101.10 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 4050079316
Conc: 97.53 ng/ml

Response_ Signal: PD090650.D\ECD1A.ch




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 592988915
Conc: 100.95 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100


Time Response_ Signal: PD090650.D\ECD2B.ch



#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 4060107897
Conc: 97.02 ng/ml


Time Response_ Signal: PD090650.D\ECD1A.ch



#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 503593872
Conc: 100.98 ng/ml


Time Response_ Signal: PD090650.D\ECD2B.ch



#14 Endrin

R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 3690195471
Conc: 96.72 ng/ml

Response_ Signal: PD090650.D\ECD1A.ch




#15 Endosulfan II

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 487323424
Conc: 99.88 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100


Time Response_ Signal: PD090650.D\ECD2B.ch



#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 3426973062
Conc: 96.94 ng/ml


Time Response_ Signal: PD090650.D\ECD1A.ch



#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 415098011
Conc: 101.16 ng/ml

Time Response_ Signal: PD090650.D\ECD2B.ch




#16 4,4'-DDD

R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 3370894623
Conc: 97.33 ng/ml

Response_ Signal: PD090650.D\ECD1A.ch

#17 4,4'-DDT




R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 425681261
Conc: 101.66 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Response_ Signal: PD090650.D\ECD2B.ch


#17 4,4'-DDT



R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 3388802413
Conc: 98.57 ng/ml

Response_ Signal: PD090650.D\ECD1A.ch


#18 Endrin aldehyde



R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 350587578
Conc: 99.07 ng/ml

Response_ Signal: PD090650.D\ECD2B.ch

#18 Endrin aldehyde




R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 2511089389
Conc: 96.57 ng/ml

Response_ Signal: PD090650.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 455705527
Conc: 100.01 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC100



Time Response_ Signal: PD090650.D\ECD2B.ch

#19 Endosulfan Sulfate


R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 3296343899
Conc: 97.06 ng/ml



Time Response_ Signal: PD090650.D\ECD1A.ch

#20 Methoxychlor


R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 205006247
Conc: 99.31 ng/ml




Time Response_ Signal: PD090650.D\ECD2B.ch

#20 Methoxychlor


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 1611694442
Conc: 96.75 ng/ml






R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 481977282
Conc: 100.39 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC100



R.T.: 6.982 min
Delta R.T.: 0.000 min
Response: 3446004756
Conc: 97.06 ng/ml



R.T.: 8.107 min
Delta R.T.: 0.000 min
Response: 295471777
Conc: 97.72 ng/ml




R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 2331618859
Conc: 97.24 ng/ml

Response_

Signal: PD090650.D\ECD1A.ch

#28 Decachlorobiphenyl

ICAL Form



R.T.: 9.066 min
Delta R.T.: 0.000 min
Response: 436118427
Conc: 98.36 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Time

Response

Signal: PD090650.D\ECD2B.ch

#28 Decachlorobiphenyl



R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 2885681422
Conc: 99.22 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090651.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:48
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:35:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	3.545	2.875	221.3E6	2076.7E6	73.844	73.982
28) SA Decachlor...	9.065	8.062	323.6E6	2118.5E6	73.652	73.550
Target Compounds						
2) A alpha-BHC	3.994	3.387	555.2E6	3629.4E6	74.331	73.972
3) MA gamma-BHC...	4.325	3.723	517.9E6	3345.4E6	74.376	74.044
4) MA Heptachlor	4.923	4.076	419.8E6	3135.8E6	74.519	74.471
5) MB Aldrin	5.265	4.361	492.2E6	3268.1E6	74.671	74.449
6) B beta-BHC	4.512	4.019	186.7E6	1377.4E6	74.235	74.130
7) B delta-BHC	4.760	4.255	517.9E6	3391.5E6	74.535	74.261
8) B Heptachlo...	5.685	4.865	428.1E6	2916.9E6	74.574	74.304
9) A Endosulfan I	6.068	5.239	405.3E6	2744.0E6	74.575	74.497
10) B gamma-Chl...	5.940	5.117	442.1E6	3181.7E6	74.831	74.517
11) B alpha-Chl...	6.021	5.182	442.9E6	3041.3E6	74.605	74.512
12) B 4,4'-DDE	6.191	5.366	402.8E6	3089.3E6	74.945	74.597
13) MA Dieldrin	6.341	5.504	438.1E6	3106.9E6	74.726	74.493
14) MA Endrin	6.568	5.781	372.2E6	2837.1E6	74.757	74.574
15) B Endosulfa...	6.781	6.072	362.7E6	2622.6E6	74.556	74.457
16) A 4,4'-DDD	6.700	5.921	305.6E6	2573.2E6	74.659	74.529
17) MA 4,4'-DDT	7.016	6.175	311.4E6	2548.7E6	74.584	74.422
18) B Endrin al...	6.910	6.251	263.1E6	1930.5E6	74.563	74.493
19) B Endosulfa...	7.145	6.474	337.9E6	2515.4E6	74.430	74.374
20) A Methoxychlor	7.488	6.745	151.2E6	1223.7E6	73.804	73.968
21) B Endrin ke...	7.625	6.983	355.2E6	2626.5E6	74.325	74.314
22) Mirex	8.108	7.176	222.5E6	1780.4E6	74.051	74.497



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090651.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:48
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:35:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm




Response_ Signal: PD090651.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 221338475
Conc: 73.84 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC075



Response_ Signal: PD090651.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 2076681010
Conc: 73.98 ng/ml



Response_ Signal: PD090651.D\ECD1A.ch

#2 alpha-BHC


R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 555178575
Conc: 74.33 ng/ml



Response_ Signal: PD090651.D\ECD2B.ch

#2 alpha-BHC


R.T.: 3.387 min
Delta R.T.: 0.000 min
Response: 3629403435
Conc: 73.97 ng/ml



Response_

Signal: PD090651.D\ECD1A.ch

#3 gamma-BHC (Lindane)



R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 517939637
Conc: 74.38 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Time

Response_

Signal: PD090651.D\ECD2B.ch

#3 gamma-BHC (Lindane)




R.T.: 3.723 min
Delta R.T.: 0.000 min
Response: 3345366469
Conc: 74.04 ng/ml

Time

Response_

Signal: PD090651.D\ECD1A.ch

#4 Heptachlor




R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 419813050
Conc: 74.52 ng/ml

Time

Response_


Signal: PD090651.D\ECD2B.ch

#4 Heptachlor



R.T.: 4.076 min
Delta R.T.: 0.000 min
Response: 3135796942
Conc: 74.47 ng/ml


Time



#5 Aldrin


R.T.: 5.265 min
Delta R.T.: 0.000 min
Response: 492167044
Conc: 74.67 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075




#5 Aldrin

R.T.: 4.361 min
Delta R.T.: 0.000 min
Response: 3268108002
Conc: 74.45 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 186698186
Conc: 74.24 ng/ml



#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 1377409505
Conc: 74.13 ng/ml



R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 517928984
Conc: 74.53 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Time

Response_

Signal: PD090651.D\ECD2B.ch

#7 delta-BHC




R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 3391516826
Conc: 74.26 ng/ml

Time

Response_

Signal: PD090651.D\ECD1A.ch

#8 Heptachlor epoxide




R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 428112423
Conc: 74.57 ng/ml

Time

Response_

Signal: PD090651.D\ECD2B.ch

#8 Heptachlor epoxide



R.T.: 4.865 min
Delta R.T.: 0.000 min
Response: 2916863262
Conc: 74.30 ng/ml

Time

Response_

Signal: PD090651.D\ECD1A.ch

#9 Endosulfan I

R.T.: 6.068 min

Delta R.T.: 0.000 min

Response: 405336624


Conc: 74.58 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDICC075



Time

Response_

Signal: PD090651.D\ECD2B.ch


#9 Endosulfan I

R.T.: 5.239 min

Delta R.T.: 0.000 min

Response: 2743979164

Conc: 74.50 ng/ml



Time

Response_

Signal: PD090651.D\ECD1A.ch


#10 gamma-Chlordane

R.T.: 5.940 min

Delta R.T.: 0.000 min

Response: 442053358

Conc: 74.83 ng/ml



Time

Response_

Signal: PD090651.D\ECD2B.ch


#10 gamma-Chlordane

R.T.: 5.117 min


Delta R.T.: 0.000 min

Response: 3181670591

Conc: 74.52 ng/ml




Time




R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 442910146
Conc: 74.61 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC075



R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 3041262707
Conc: 74.51 ng/ml




R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 402806011
Conc: 74.95 ng/ml



R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 3089327488
Conc: 74.60 ng/ml

Response_ Signal: PD090651.D\ECD1A.ch




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 438145830
Conc: 74.73 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075


Time Response_ Signal: PD090651.D\ECD2B.ch



#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 3106916142
Conc: 74.49 ng/ml


Time Response_ Signal: PD090651.D\ECD1A.ch



#14 Endrin


R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 372217983
Conc: 74.76 ng/ml

Time Response_ Signal: PD090651.D\ECD2B.ch



#14 Endrin


R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 2837058459
Conc: 74.57 ng/ml



#15 Endosulfan II


R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 362696180
Conc: 74.56 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 2622605787
Conc: 74.46 ng/ml




#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 305645357
Conc: 74.66 ng/ml



#16 4,4'-DDD


R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 2573202277
Conc: 74.53 ng/ml



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 311444736
Conc: 74.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075




#17 4,4'-DDT

R.T.: 6.175 min
Delta R.T.: 0.000 min
Response: 2548691729
Conc: 74.42 ng/ml



#18 Endrin aldehyde


R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 263096696
Conc: 74.56 ng/ml



#18 Endrin aldehyde

R.T.: 6.251 min
Delta R.T.: 0.000 min
Response: 1930452420
Conc: 74.49 ng/ml

Response_ Signal: PD090651.D\ECD1A.ch




#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 337864118
Conc: 74.43 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075


Time Response_ Signal: PD090651.D\ECD2B.ch



#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 2515420876
Conc: 74.37 ng/ml


Time Response_ Signal: PD090651.D\ECD1A.ch



#20 Methoxychlor

R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 151150843
Conc: 73.80 ng/ml


Time Response_ Signal: PD090651.D\ECD2B.ch



#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 1223732210
Conc: 73.97 ng/ml

Response_ Signal: PD090651.D\ECD1A.ch




#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 355238454
Conc: 74.32 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075


Response_ Signal: PD090651.D\ECD2B.ch



#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 2626540426
Conc: 74.31 ng/ml


Response_ Signal: PD090651.D\ECD1A.ch



#22 Mirex


R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 222493333
Conc: 74.05 ng/ml

Response_ Signal: PD090651.D\ECD2B.ch



#22 Mirex


R.T.: 7.176 min
Delta R.T.: 0.000 min
Response: 1780357534
Conc: 74.50 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 323645321
Conc: 73.65 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 2118537336
Conc: 73.55 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090652.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 12:01
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:30:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.545	2.875	151.5E6	1430.9E6	50.000	50.000
28) SA Decachlor...	9.064	8.062	225.3E6	1465.4E6	50.000	50.000
Target Compounds						
2) A alpha-BHC	3.994	3.387	362.6E6	2484.2E6	50.000	50.000
3) MA gamma-BHC...	4.325	3.723	340.0E6	2298.6E6	50.000	50.000
4) MA Heptachlor	4.923	4.075	277.3E6	2162.0E6	50.000	50.000
5) MB Aldrin	5.264	4.361	325.8E6	2251.7E6	50.000	50.000
6) B beta-BHC	4.512	4.019	127.0E6	954.0E6	50.000	50.000
7) B delta-BHC	4.760	4.255	338.6E6	2323.3E6	50.000	50.000
8) B Heptachlo...	5.684	4.864	285.9E6	2033.6E6	50.000	50.000
9) A Endosulfan I	6.068	5.238	272.5E6	1909.3E6	50.000	50.000
10) B gamma-Chl...	5.940	5.117	292.8E6	2187.4E6	50.000	50.000
11) B alpha-Chl...	6.021	5.181	296.9E6	2094.7E6	50.000	50.000
12) B 4,4'-DDE	6.190	5.366	265.9E6	2127.4E6	50.000	50.000
13) MA Dieldrin	6.340	5.504	290.9E6	2154.8E6	50.000	50.000
14) MA Endrin	6.568	5.781	246.9E6	1970.0E6	50.000	50.000
15) B Endosulfa...	6.781	6.072	244.3E6	1821.6E6	50.000	50.000
16) A 4,4'-DDD	6.700	5.921	202.8E6	1778.0E6	50.000	50.000
17) MA 4,4'-DDT	7.015	6.174	205.9E6	1743.4E6	50.000	50.000
18) B Endrin al...	6.910	6.250	178.6E6	1344.7E6	50.000	50.000
19) B Endosulfa...	7.145	6.474	227.8E6	1748.0E6	50.000	50.000
20) A Methoxychlor	7.488	6.745	103.9E6	859.9E6	50.000	50.000
21) B Endrin ke...	7.625	6.983	239.1E6	1827.5E6	50.000	50.000
22) Mirex	8.107	7.175	154.6E6	1232.0E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090652.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:01
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:30:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090652.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 151504775
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Time 3.20 3.30 3.40 3.50 3.60 3.70 3.80 3.90

Response_ Signal: PD090652.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 1430869153
Conc: 50.00 ng/ml

Time 2.70 2.75 2.80 2.85 2.90 2.95 3.00 3.05

Response_ Signal: PD090652.D\ECD1A.ch

#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 362610738
Conc: 50.00 ng/ml

Time 3.70 3.80 3.90 4.00 4.10 4.20 4.30

Response_ Signal: PD090652.D\ECD2B.ch

#2 alpha-BHC


R.T.: 3.387 min
Delta R.T.: 0.000 min
Response: 2484185326
Conc: 50.00 ng/ml

Time 3.20 3.30 3.40 3.50 3.60

Response_

Signal: PD090652.D\ECD1A.ch

#3 gamma-BHC (Lindane)



R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 339982682
Conc: 50.00 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Time

Response_

Signal: PD090652.D\ECD2B.ch

#3 gamma-BHC (Lindane)




R.T.: 3.723 min
Delta R.T.: 0.000 min
Response: 2298638612
Conc: 50.00 ng/ml

Time

Response_

Signal: PD090652.D\ECD1A.ch

#4 Heptachlor




R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 277277005
Conc: 50.00 ng/ml

Time

Response_


Signal: PD090652.D\ECD2B.ch

#4 Heptachlor



R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 2161964168
Conc: 50.00 ng/ml

Time



R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 325847234
Conc: 50.00 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Time

Response_

Signal: PD090652.D\ECD2B.ch

#5 Aldrin




R.T.: 4.361 min
Delta R.T.: 0.000 min
Response: 2251688760
Conc: 50.00 ng/ml

Time

Response_

Signal: PD090652.D\ECD1A.ch

#6 beta-BHC




R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 127018373
Conc: 50.00 ng/ml

Time

Response_

Signal: PD090652.D\ECD2B.ch

#6 beta-BHC



R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 954037739
Conc: 50.00 ng/ml

Time

Response_ Signal: PD090652.D\ECD1A.ch

#7 delta-BHC

R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 338575319
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Time

Response_ Signal: PD090652.D\ECD2B.ch

#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2323317526
Conc: 50.00 ng/ml

Time

Response_ Signal: PD090652.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 285933842
Conc: 50.00 ng/ml


Time

Response_ Signal: PD090652.D\ECD2B.ch

#8 Heptachlor epoxide

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 2033647315
Conc: 50.00 ng/ml


Time



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 272452330
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1909283878
Conc: 50.00 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 292781895
Conc: 50.00 ng/ml



#10 gamma-Chlordane


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 2187405429
Conc: 50.00 ng/ml



#11 alpha-Chlordane


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 296932529
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2094650466
Conc: 50.00 ng/ml




#12 4,4'-DDE

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 265880490
Conc: 50.00 ng/ml



#12 4,4'-DDE


R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 2127431088
Conc: 50.00 ng/ml



#13 Dieldrin


R.T.: 6.340 min
Delta R.T.: 0.000 min
Response: 290909064
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 2154808119
Conc: 50.00 ng/ml



#14 Endrin


R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 246912451
Conc: 50.00 ng/ml



#14 Endrin

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 1970044096
Conc: 50.00 ng/ml

Response_ Signal: PD090652.D\ECD1A.ch




#15 Endosulfan II

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 244253888
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050


Time Response_ Signal: PD090652.D\ECD2B.ch



#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1821608734
Conc: 50.00 ng/ml


Time Response_ Signal: PD090652.D\ECD1A.ch



#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 202773551
Conc: 50.00 ng/ml


Time Response_ Signal: PD090652.D\ECD2B.ch



#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1778046885
Conc: 50.00 ng/ml

Response_ Signal: PD090652.D\ECD1A.ch




#17 4,4'-DDT

R.T.: 7.015 min
Delta R.T.: 0.000 min
Response: 205896645
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050


Time Response_ Signal: PD090652.D\ECD2B.ch



#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1743410018
Conc: 50.00 ng/ml


Time Response_ Signal: PD090652.D\ECD1A.ch



#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 178583940
Conc: 50.00 ng/ml

Time Response_ Signal: PD090652.D\ECD2B.ch



#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1344655981
Conc: 50.00 ng/ml

Response_ Signal: PD090652.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 227808376
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Time

Response_ Signal: PD090652.D\ECD2B.ch

#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1748039260
Conc: 50.00 ng/ml

Time

Response_ Signal: PD090652.D\ECD1A.ch

#20 Methoxychlor

R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 103932037
Conc: 50.00 ng/ml


Time

Response_ Signal: PD090652.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 859946427
Conc: 50.00 ng/ml


Time



#21 Endrin ketone


R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 239116023
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1827533277
Conc: 50.00 ng/ml




#22 Mirex

R.T.: 8.107 min
Delta R.T.: 0.000 min
Response: 154621594
Conc: 50.00 ng/ml



#22 Mirex


R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1232028398
Conc: 50.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.064 min
Delta R.T.: 0.000 min
Response: 225311337
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 1465399343
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:15
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:38:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	78107465	743.2E6	25.786	26.090
28)	SA Decachlor...	9.065	8.061	117.7E6	751.4E6	26.311	25.807
Target Compounds							
2)	A alpha-BHC	3.993	3.386	168.6E6	1262.5E6	23.132	25.545
3)	MA gamma-BHC...	4.324	3.722	160.4E6	1170.2E6	23.501	25.669
4)	MA Heptachlor	4.922	4.074	132.2E6	1112.6E6	23.835	26.052
5)	MB Aldrin	5.263	4.360	156.0E6	1154.3E6	23.982	25.959
6)	B beta-BHC	4.511	4.018	64015475	495.1E6	25.339	26.214
7)	B delta-BHC	4.759	4.255	158.6E6	1178.9E6	23.327	25.605
8)	B Heptachlo...	5.684	4.864	139.6E6	1059.4E6	24.488	26.460
9)	A Endosulfan I	6.067	5.238	133.7E6	987.0E6	24.700	26.323
10)	B gamma-Chl...	5.939	5.116	139.8E6	1108.7E6	23.988	25.717
11)	B alpha-Chl...	6.020	5.181	154.3E6	1067.7E6	25.731	25.860
12)	B 4,4'-DDE	6.190	5.365	126.3E6	1089.4E6	23.861	25.967
13)	MA Dieldrin	6.340	5.503	138.8E6	1105.1E6	23.985	26.105
14)	MA Endrin	6.567	5.780	117.8E6	1017.3E6	23.978	26.284
15)	B Endosulfa...	6.781	6.071	119.7E6	938.0E6	24.710	26.203
16)	A 4,4'-DDD	6.700	5.920	96019034	905.9E6	23.822	25.918
17)	MA 4,4'-DDT	7.016	6.173	98197092	874.0E6	23.870	25.388
18)	B Endrin al...	6.909	6.250	89492653	694.2E6	25.271	26.317
19)	B Endosulfa...	7.144	6.473	112.0E6	896.5E6	24.754	26.114
20)	A Methoxychlor	7.488	6.744	52364279	443.4E6	25.424	26.326
21)	B Endrin ke...	7.625	6.982	117.4E6	945.4E6	24.668	26.289
22)	Mirex	8.108	7.175	79715192	640.8E6	26.131	26.336



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:15
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:38:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm




Response_ Signal: PD090653.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 78107465
Conc: 25.79 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025



Response_ Signal: PD090653.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 743162855
Conc: 26.09 ng/ml



Response_ Signal: PD090653.D\ECD1A.ch

#2 alpha-BHC


R.T.: 3.993 min
Delta R.T.: 0.000 min
Response: 168576733
Conc: 23.13 ng/ml




Response_ Signal: PD090653.D\ECD2B.ch

#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 1262506307
Conc: 25.54 ng/ml






#3 gamma-BHC (Lindane)


R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 160449432
Conc: 23.50 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 1170208039
Conc: 25.67 ng/ml




#4 Heptachlor

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 132221469
Conc: 23.83 ng/ml



#4 Heptachlor

R.T.: 4.074 min
Delta R.T.: 0.000 min
Response: 1112572417
Conc: 26.05 ng/ml



R.T.: 5.263 min
Delta R.T.: 0.000 min
Response: 155953472
Conc: 23.98 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Time

Response_

Signal: PD090653.D\ECD2B.ch

#5 Aldrin




R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 1154309416
Conc: 25.96 ng/ml

Time

Response_

Signal: PD090653.D\ECD1A.ch

#6 beta-BHC




R.T.: 4.511 min
Delta R.T.: 0.000 min
Response: 64015475
Conc: 25.34 ng/ml

Time

Response_


Signal: PD090653.D\ECD2B.ch

#6 beta-BHC



R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 495101315
Conc: 26.21 ng/ml


Time



#7 delta-BHC


R.T.: 4.759 min
Delta R.T.: 0.000 min
Response: 158561347
Conc: 23.33 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 1178894488
Conc: 25.60 ng/ml



#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 139624515
Conc: 24.49 ng/ml




#8 Heptachlor epoxide

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 1059359842
Conc: 26.46 ng/ml

Response_

Signal: PD090653.D\ECD1A.ch

#9 Endosulfan I



R.T.: 6.067 min
Delta R.T.: 0.000 min
Response: 133716398
Conc: 24.70 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Time

Response

Signal: PD090653.D\ECD2B.ch

#9 Endosulfan I




R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 986966005
Conc: 26.32 ng/ml

Time

Response_

Signal: PD090653.D\ECD1A.ch

#10 gamma-Chlordane




R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 139816717
Conc: 23.99 ng/ml

Time

Response


Signal: PD090653.D\ECD2B.ch

#10 gamma-Chlordane



R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 1108664102
Conc: 25.72 ng/ml


Time



#11 alpha-Chlordane


R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 154261488
Conc: 25.73 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 1067721561
Conc: 25.86 ng/ml



#12 4,4'-DDE


R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 126326499
Conc: 23.86 ng/ml



#12 4,4'-DDE

R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 1089429298
Conc: 25.97 ng/ml

Response_ Signal: PD090653.D\ECD1A.ch




#13 Dieldrin

R.T.: 6.340 min
Delta R.T.: 0.000 min
Response: 138758150
Conc: 23.99 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025


Time Response_ Signal: PD090653.D\ECD2B.ch



#13 Dieldrin

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 1105079839
Conc: 26.11 ng/ml


Time Response_ Signal: PD090653.D\ECD1A.ch



#14 Endrin

R.T.: 6.567 min
Delta R.T.: 0.000 min
Response: 117779573
Conc: 23.98 ng/ml

Time Response_ Signal: PD090653.D\ECD2B.ch



#14 Endrin


R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 1017324913
Conc: 26.28 ng/ml

Response_

Signal: PD090653.D\ECD1A.ch

#15 Endosulfan II

ICAL Form



R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 119743666
Conc: 24.71 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Time

Response_

Signal: PD090653.D\ECD2B.ch

#15 Endosulfan II




R.T.: 6.071 min
Delta R.T.: 0.000 min
Response: 938003929
Conc: 26.20 ng/ml

Time

Response_

Signal: PD090653.D\ECD1A.ch

#16 4,4'-DDD




R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 96019034
Conc: 23.82 ng/ml

Time

Response_


Signal: PD090653.D\ECD2B.ch

#16 4,4'-DDD



R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 905938376
Conc: 25.92 ng/ml


Time



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 98197092
Conc: 23.87 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025




#17 4,4'-DDT

R.T.: 6.173 min
Delta R.T.: 0.000 min
Response: 873976959
Conc: 25.39 ng/ml




#18 Endrin aldehyde

R.T.: 6.909 min
Delta R.T.: 0.000 min
Response: 89492653
Conc: 25.27 ng/ml



#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 694183960
Conc: 26.32 ng/ml



R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 111999004
Conc: 24.75 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Time

Response_

Signal: PD090653.D\ECD2B.ch

#19 Endosulfan Sulfate




R.T.: 6.473 min
Delta R.T.: 0.000 min
Response: 896525775
Conc: 26.11 ng/ml

Time

Response_

Signal: PD090653.D\ECD1A.ch

#20 Methoxychlor




R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 52364279
Conc: 25.42 ng/ml

Time

Response_

Signal: PD090653.D\ECD2B.ch


#20 Methoxychlor



R.T.: 6.744 min
Delta R.T.: 0.000 min
Response: 443376884
Conc: 26.33 ng/ml

Time

Response_ Signal: PD090653.D\ECD1A.ch




#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 117381168
Conc: 24.67 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025


Time Response_ Signal: PD090653.D\ECD2B.ch



#21 Endrin ketone

R.T.: 6.982 min
Delta R.T.: 0.000 min
Response: 945417829
Conc: 26.29 ng/ml


Time Response_ Signal: PD090653.D\ECD1A.ch



#22 Mirex


R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 79715192
Conc: 26.13 ng/ml

Time Response_ Signal: PD090653.D\ECD2B.ch



#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 640790883
Conc: 26.34 ng/ml



#28 Decachlorobiphenyl


R.T.: 9.065 min

Delta R.T.: 0.000 min

Response: 117676418

Conc: 26.31 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025



#28 Decachlorobiphenyl

R.T.: 8.061 min

Delta R.T.: 0.000 min

Response: 751430638

Conc: 25.81 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090654.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 12:28
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC005

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:41:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.543	2.874	17271955	174.7E6	5.550m	5.868
28) SA Decachlor...	9.066	8.062	27500355	174.9E6	5.879	5.775
Target Compounds						
2) A alpha-BHC	3.993	3.386	33677596	288.9E6	4.692	5.655
3) MA gamma-BHC...	4.324	3.723	33423739	269.5E6	4.916	5.703
4) MA Heptachlor	4.922	4.075	28789856	260.9E6	5.151	5.850
5) MB Aldrin	5.264	4.360	33624118	268.6E6	5.136	5.800
6) B beta-BHC	4.512	4.019	14963256	118.0E6	5.712	5.951
7) B delta-BHC	4.760	4.255	33098385	272.1E6	4.895	5.702
8) B Heptachlo...	5.684	4.864	30689599	265.3E6	5.301	6.280
9) A Endosulfan I	6.067	5.238	30143871	234.9E6	5.444	5.962
10) B gamma-Chl...	5.940	5.117	31705442	261.5E6	5.346	5.818
11) B alpha-Chl...	6.019	5.181	34603140	252.7E6	5.576m	5.857
12) B 4,4'-DDE	6.190	5.366	27280775	256.9E6	5.122	5.859
13) MA Dieldrin	6.341	5.504	30082015	258.0E6	5.159	5.839
14) MA Endrin	6.568	5.781	25601278	238.2E6	5.168	5.883
15) B Endosulfa...	6.781	6.072	28078989	222.9E6	5.616	5.936
16) A 4,4'-DDD	6.700	5.921	20742842	211.9E6	5.116	5.814
17) MA 4,4'-DDT	7.016	6.175	20485670	188.3E6	4.984	5.370
18) B Endrin al...	6.910	6.250	20741409	164.8E6	5.663	5.950
19) B Endosulfa...	7.144	6.474	25464611	209.5E6	5.490	5.845
20) A Methoxychlor	7.489	6.745	11946535	98858490	5.620	5.672
21) B Endrin ke...	7.624	6.983	26302012	223.7E6	5.414m	5.931
22) Mirex	8.108	7.175	19009608	155.6E6	5.939	6.058

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090654.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:28
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC005

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 10/19/2025


Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:41:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090654.D\ECD1A.ch



#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: 0.000 min
Response: 17271955
Conc: 5.55 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDICC005


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025


Time Response_ Signal: PD090654.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 174747032
Conc: 5.87 ng/ml


Response_ Signal: PD090654.D\ECD1A.ch



#2 alpha-BHC


R.T.: 3.993 min
Delta R.T.: 0.000 min
Response: 33677596
Conc: 4.69 ng/ml

Time Response_ Signal: PD090654.D\ECD2B.ch



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 288924308
Conc: 5.65 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 33423739
Conc: 4.92 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#3 gamma-BHC (Lindane)

R.T.: 3.723 min
Delta R.T.: 0.000 min
Response: 269467396
Conc: 5.70 ng/ml




#4 Heptachlor

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 28789856
Conc: 5.15 ng/ml



#4 Heptachlor

R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 260945196
Conc: 5.85 ng/ml




#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 33624118
Conc: 5.14 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 268648325
Conc: 5.80 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 14963256
Conc: 5.71 ng/ml



#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 118003884
Conc: 5.95 ng/ml




R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 33098385
Conc: 4.89 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC005

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 272079912
Conc: 5.70 ng/ml



R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 30689599
Conc: 5.30 ng/ml



R.T.: 4.864 min
Delta R.T.: 0.001 min
Response: 265262691
Conc: 6.28 ng/ml




#9 Endosulfan I

R.T.: 6.067 min
Delta R.T.: 0.000 min
Response: 30143871
Conc: 5.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 234861846
Conc: 5.96 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 31705442
Conc: 5.35 ng/ml



#10 gamma-Chlordane

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 261533283
Conc: 5.82 ng/ml




#11 alpha-Chlordane

R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 34603140
Conc: 5.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 252662629
Conc: 5.86 ng/ml




#12 4,4'-DDE

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 27280775
Conc: 5.12 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 256868291
Conc: 5.86 ng/ml




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 30082015
Conc: 5.16 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 257969409
Conc: 5.84 ng/ml




#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 25601278
Conc: 5.17 ng/ml



#14 Endrin

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 238232371
Conc: 5.88 ng/ml




#15 Endosulfan II

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 28078989
Conc: 5.62 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 222935892
Conc: 5.94 ng/ml




#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 20742842
Conc: 5.12 ng/ml



#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 211866590
Conc: 5.81 ng/ml




#17 4,4'-DDT

R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 20485670
Conc: 4.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#17 4,4'-DDT

R.T.: 6.175 min
Delta R.T.: 0.000 min
Response: 188338004
Conc: 5.37 ng/ml




#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 20741409
Conc: 5.66 ng/ml



#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 164766045
Conc: 5.95 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 25464611
Conc: 5.49 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 209504586
Conc: 5.84 ng/ml




#20 Methoxychlor

R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 11946535
Conc: 5.62 ng/ml



#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 98858490
Conc: 5.67 ng/ml




#21 Endrin ketone

R.T.: 7.624 min
Delta R.T.: 0.000 min
Response: 26302012
Conc: 5.41 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 223704577
Conc: 5.93 ng/ml




#22 Mirex

R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 19009608
Conc: 5.94 ng/ml



#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 155640388
Conc: 6.06 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.000 min
Response: 27500355
Conc: 5.88 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 174937832
Conc: 5.78 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090657.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 13:09
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:49:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:48:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.545	2.874	108.8E6	1215.0E6	50.000	50.000
28) SA Decachlor...	9.066	8.062	167.8E6	1105.6E6	50.000	50.000
Target Compounds						
23) Chlordane-1	4.708	3.898	92608674	605.2E6	500.000	500.000
24) Chlordane-2	5.235	4.479	92865500	631.9E6	500.000	500.000
25) Chlordane-3	5.941	5.117	379.1E6	1885.9E6	500.000	500.000
26) Chlordane-4	6.026	5.181	458.8E6	1625.9E6	500.000	500.000
27) Chlordane-5	6.865	6.081	74238996	730.3E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090657.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 13:09
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:49:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:48:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 108758097
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1214996007
Conc: 50.00 ng/ml




#23 Chlordane-1

R.T.: 4.708 min
Delta R.T.: 0.000 min
Response: 92608674
Conc: 500.00 ng/ml



#23 Chlordane-1


R.T.: 3.898 min
Delta R.T.: 0.000 min
Response: 605185458
Conc: 500.00 ng/ml



#24 Chlordane-2


R.T.: 5.235 min
Delta R.T.: 0.000 min
Response: 92865500
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500




#24 Chlordane-2

R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 631892083
Conc: 500.00 ng/ml



#25 Chlordane-3

R.T.: 5.941 min
Delta R.T.: 0.000 min
Response: 379095261
Conc: 500.00 ng/ml



#25 Chlordane-3

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 1885915535
Conc: 500.00 ng/ml

Response_ Signal: PD090657.D\ECD1A.ch

#26 Chlordane-4

R.T.: 6.026 min
Delta R.T.: 0.000 min
Response: 458791566
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Time

Response_ Signal: PD090657.D\ECD2B.ch

#26 Chlordane-4

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 1625937453
Conc: 500.00 ng/ml

Time

Response_ Signal: PD090657.D\ECD1A.ch

#27 Chlordane-5

R.T.: 6.865 min
Delta R.T.: 0.000 min
Response: 74238996
Conc: 500.00 ng/ml


Time

Response_ Signal: PD090657.D\ECD2B.ch

#27 Chlordane-5

R.T.: 6.081 min
Delta R.T.: 0.000 min
Response: 730253582
Conc: 500.00 ng/ml


Time



#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.000 min
Response: 167755700
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 1105630741
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090662.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 14:17
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:36:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:35:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.545	2.875	149.0E6	1451.5E6	50.000	50.000
7) SA Decachlor...	9.066	8.062	225.6E6	1489.1E6	50.000	50.000
Target Compounds						
2) Toxaphene-1	6.236	5.466	16282219	131.9E6	500.000	500.000
3) Toxaphene-2	6.436	5.638	24310877	88280081	500.000	500.000
4) Toxaphene-3	7.143	6.745	42855293	329.4E6	500.000	500.000
5) Toxaphene-4	7.558	7.189	47868684	247.6E6	500.000	500.000
6) Toxaphene-5	7.923	7.320	30240231	123.0E6	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090662.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 14:17
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:36:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:35:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 149025126
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXICC500




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 1451452961
Conc: 50.00 ng/ml




#2 Toxaphene-1

R.T.: 6.236 min
Delta R.T.: 0.000 min
Response: 16282219
Conc: 500.00 ng/ml



#2 Toxaphene-1


R.T.: 5.466 min
Delta R.T.: 0.000 min
Response: 131850816
Conc: 500.00 ng/ml



#3 Toxaphene-2


R.T.: 6.436 min
Delta R.T.: 0.000 min
Response: 24310877
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXICC500




#3 Toxaphene-2

R.T.: 5.638 min
Delta R.T.: 0.000 min
Response: 88280081
Conc: 500.00 ng/ml



#4 Toxaphene-3


R.T.: 7.143 min
Delta R.T.: 0.000 min
Response: 42855293
Conc: 500.00 ng/ml



#4 Toxaphene-3

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 329441315
Conc: 500.00 ng/ml

Response_ Signal: PD090662.D\ECD1A.ch




#5 Toxaphene-4

R.T.: 7.558 min
Delta R.T.: 0.000 min
Response: 47868684
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXICC500


Time Response_ Signal: PD090662.D\ECD2B.ch



#5 Toxaphene-4

R.T.: 7.189 min
Delta R.T.: 0.000 min
Response: 247552187
Conc: 500.00 ng/ml


Time Response_ Signal: PD090662.D\ECD1A.ch



#6 Toxaphene-5

R.T.: 7.923 min
Delta R.T.: 0.000 min
Response: 30240231
Conc: 500.00 ng/ml

Time Response_ Signal: PD090662.D\ECD2B.ch



#6 Toxaphene-5


R.T.: 7.320 min
Delta R.T.: 0.000 min
Response: 123016477
Conc: 500.00 ng/ml

Response_

Signal: PD090662.D\ECD1A.ch

#7 Decachlorobiphenyl

ICAL Form



R.T.: 9.066 min
Delta R.T.: 0.000 min
Response: 225576698
Conc: 50.00 ng/ml


Instrument :
ECD_D
ClientSampleId :
PTOXICC500

Time

Response_

Signal: PD090662.D\ECD2B.ch

#7 Decachlorobiphenyl



R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 1489074983
Conc: 50.00 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090665.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 16:46
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD101725

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 17:27:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:02:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	153.7E6	1464.9E6	49.347	49.195
28)	SA Decachlor...	9.065	8.062	232.3E6	1556.2E6	49.651	51.373
Target Compounds							
2)	A alpha-BHC	3.993	3.386	368.9E6	2547.7E6	51.395	49.860
3)	MA gamma-BHC...	4.325	3.722	346.1E6	2351.4E6	50.913	49.766
4)	MA Heptachlor	4.923	4.074	269.7E6	2210.7E6	48.243	49.564
5)	MB Aldrin	5.264	4.360	332.4E6	2302.8E6	50.770	49.716
6)	B beta-BHC	4.512	4.018	128.8E6	974.3E6	49.166	49.134
7)	B delta-BHC	4.760	4.255	343.7E6	2372.2E6	50.832	49.714
8)	B Heptachlo...	5.684	4.864	292.1E6	2073.7E6	50.455	48.633
9)	A Endosulfan I	6.068	5.238	276.7E6	1960.5E6	49.982	49.772
10)	B gamma-Chl...	5.940	5.117	297.4E6	2236.0E6	50.146	49.745
11)	B alpha-Chl...	6.020	5.181	299.6E6	2144.4E6	48.473	49.711
12)	B 4,4'-DDE	6.190	5.366	269.1E6	2179.5E6	50.526	49.716
13)	MA Dieldrin	6.341	5.503	295.7E6	2209.3E6	50.705	50.003
14)	MA Endrin	6.568	5.780	251.8E6	2041.8E6	50.824	50.421
15)	B Endosulfa...	6.781	6.072	251.4E6	1876.1E6	50.275	49.956
16)	A 4,4'-DDD	6.701	5.920	206.7E6	1821.0E6	50.990	49.977
17)	MA 4,4'-DDT	7.016	6.174	206.8E6	1798.7E6	50.320	51.285
18)	B Endrin al...	6.910	6.250	184.2E6	1401.3E6	50.283	50.600
19)	B Endosulfa...	7.144	6.474	231.2E6	1809.0E6	49.837	50.467
20)	A Methoxychlor	7.489	6.745	104.2E6	890.1E6	48.999	51.076
21)	B Endrin ke...	7.625	6.983	242.9E6	1911.1E6	49.982	50.668
22)	Mirex	8.108	7.176	158.1E6	1294.5E6	49.394	50.389



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090665.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 16:46
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD101725

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 17:27:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 153674114
Conc: 49.35 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1464875316
Conc: 49.19 ng/ml




#2 alpha-BHC

R.T.: 3.993 min
Delta R.T.: 0.000 min
Response: 368866522
Conc: 51.39 ng/ml



#2 alpha-BHC


R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2547652740
Conc: 49.86 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 346148511
Conc: 50.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 2351401356
Conc: 49.77 ng/ml




#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 269655012
Conc: 48.24 ng/ml



#4 Heptachlor

R.T.: 4.074 min
Delta R.T.: 0.000 min
Response: 2210699675
Conc: 49.56 ng/ml



R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 332404308
Conc: 50.77 ng/ml


Instrument :
ECD_D
ClientSampleId :
ICVPD101725

Time

Response

Signal: PD090665.D\ECD2B.ch

#5 Aldrin




R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2302805919
Conc: 49.72 ng/ml

Time

Response

Signal: PD090665.D\ECD1A.ch

#6 beta-BHC




R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 128796459
Conc: 49.17 ng/ml

Time

Response

Signal: PD090665.D\ECD2B.ch

#6 beta-BHC




R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 974305644
Conc: 49.13 ng/ml

Time

Response_ Signal: PD090665.D\ECD1A.ch

#7 delta-BHC




R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 343712775
Conc: 50.83 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725

Time Response_ Signal: PD090665.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2372178334
Conc: 49.71 ng/ml

Response_ Signal: PD090665.D\ECD1A.ch


#8 Heptachlor epoxide




R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 292086915
Conc: 50.45 ng/ml

Time Response_ Signal: PD090665.D\ECD2B.ch

#8 Heptachlor epoxide




R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 2073663307
Conc: 48.63 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 276731064
Conc: 49.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1960516174
Conc: 49.77 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 297423108
Conc: 50.15 ng/ml



#10 gamma-Chlordane


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 2235964406
Conc: 49.74 ng/ml



#11 alpha-Chlordane


R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 299573543
Conc: 48.47 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2144431564
Conc: 49.71 ng/ml




#12 4,4'-DDE

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 269131461
Conc: 50.53 ng/ml



#12 4,4'-DDE


R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 2179474893
Conc: 49.72 ng/ml



#13 Dieldrin


R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 295678653
Conc: 50.71 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#13 Dieldrin

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 2209322407
Conc: 50.00 ng/ml




#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 251765100
Conc: 50.82 ng/ml



#14 Endrin


R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 2041761986
Conc: 50.42 ng/ml



#15 Endosulfan II


R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 251374054
Conc: 50.28 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1876116999
Conc: 49.96 ng/ml




#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: 0.000 min
Response: 206725943
Conc: 50.99 ng/ml



#16 4,4'-DDD


R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 1821049286
Conc: 49.98 ng/ml



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 206839113
Conc: 50.32 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1798730996
Conc: 51.29 ng/ml



#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 184171641
Conc: 50.28 ng/ml



#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1401255725
Conc: 50.60 ng/ml

Response_ Signal: PD090665.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 231154352
Conc: 49.84 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725

Time

Response_ Signal: PD090665.D\ECD2B.ch

#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1808994693
Conc: 50.47 ng/ml

Time

Response_ Signal: PD090665.D\ECD1A.ch

#20 Methoxychlor

R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 104150715
Conc: 49.00 ng/ml


Time

Response_ Signal: PD090665.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 890144630
Conc: 51.08 ng/ml


Time



#21 Endrin ketone


R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 242853020
Conc: 49.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1911076638
Conc: 50.67 ng/ml




#22 Mirex

R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 158102045
Conc: 49.39 ng/ml



#22 Mirex


R.T.: 7.176 min
Delta R.T.: 0.000 min
Response: 1294533618
Conc: 50.39 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 232264787
Conc: 49.65 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 1556154030
Conc: 51.37 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Continuing Calib Date: 10/17/2025

Initial Calibration Date(s): 10/17/2025 10/17/2025

Continuing Calib Time: 20:18

Initial Calibration Time(s): 11:34 12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.06	8.96	9.16	-0.01
Tetrachloro-m-xylene	3.54	3.55	3.45	3.65	0.01
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.33	4.23	4.43	0.01
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.69	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.15	7.15	7.05	7.25	0.01
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.63	7.53	7.73	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Continuing Calib Date: 10/17/2025

Initial Calibration Date(s): 10/17/2025 10/17/2025

Continuing Calib Time: 20:18

Initial Calibration Time(s): 11:34 12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.26	4.26	4.16	4.36	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.08	4.08	3.98	4.18	0.01
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.24	5.24	5.14	5.34	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.37	5.37	5.27	5.47	0.01
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.75	6.75	6.65	6.85	0.01
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.12	5.12	5.02	5.22	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025
Client Sample No.: CCAL01 **Date Analyzed:** 10/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090670.D **Time Analyzed:** 20:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	53.400	50.000	6.8
4,4'-DDE	6.191	6.090	6.290	53.210	50.000	6.4
4,4'-DDT	7.016	6.915	7.115	52.590	50.000	5.2
Aldrin	5.264	5.164	5.364	53.680	50.000	7.4
alpha-BHC	3.994	3.894	4.094	54.470	50.000	8.9
alpha-Chlordane	6.021	5.921	6.121	51.170	50.000	2.3
beta-BHC	4.512	4.412	4.612	51.910	50.000	3.8
Decachlorobiphenyl	9.066	8.964	9.164	51.560	50.000	3.1
delta-BHC	4.760	4.660	4.860	53.950	50.000	7.9
Dieldrin	6.341	6.240	6.440	53.290	50.000	6.6
Endosulfan I	6.068	5.968	6.168	52.600	50.000	5.2
Endosulfan II	6.781	6.681	6.881	53.170	50.000	6.3
Endosulfan sulfate	7.145	7.045	7.245	52.300	50.000	4.6
Endrin	6.568	6.468	6.668	53.340	50.000	6.7
Endrin aldehyde	6.910	6.810	7.010	52.750	50.000	5.5
Endrin ketone	7.625	7.525	7.725	52.370	50.000	4.7
gamma-BHC (Lindane)	4.324	4.225	4.425	53.930	50.000	7.9
gamma-Chlordane	5.940	5.840	6.040	52.910	50.000	5.8
Heptachlor	4.923	4.823	5.023	51.150	50.000	2.3
Heptachlor epoxide	5.685	5.584	5.784	53.260	50.000	6.5
Methoxychlor	7.489	7.388	7.588	50.870	50.000	1.7
Tetrachloro-m-xylene	3.544	3.445	3.645	51.790	50.000	3.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090670.D **Time Analyzed:** 20:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.921	5.821	6.021	52.390	50.000	4.8
4,4'-DDE	5.365	5.266	5.466	52.220	50.000	4.4
4,4'-DDT	6.174	6.074	6.274	54.440	50.000	8.9
Aldrin	4.360	4.261	4.461	52.220	50.000	4.4
alpha-BHC	3.386	3.287	3.487	52.250	50.000	4.5
alpha-Chlordane	5.181	5.081	5.281	52.420	50.000	4.8
beta-BHC	4.018	3.919	4.119	51.770	50.000	3.5
Decachlorobiphenyl	8.062	7.962	8.162	53.450	50.000	6.9
delta-BHC	4.255	4.155	4.355	52.490	50.000	5.0
Dieldrin	5.503	5.404	5.604	52.550	50.000	5.1
Endosulfan I	5.238	5.138	5.338	52.240	50.000	4.5
Endosulfan II	6.072	5.972	6.172	52.390	50.000	4.8
Endosulfan sulfate	6.473	6.374	6.574	52.790	50.000	5.6
Endrin	5.780	5.681	5.881	52.830	50.000	5.7
Endrin aldehyde	6.250	6.150	6.350	53.080	50.000	6.2
Endrin ketone	6.983	6.883	7.083	52.900	50.000	5.8
gamma-BHC (Lindane)	3.722	3.623	3.823	52.320	50.000	4.6
gamma-Chlordane	5.116	5.017	5.217	52.450	50.000	4.9
Heptachlor	4.075	3.975	4.175	52.480	50.000	5.0
Heptachlor epoxide	4.863	4.764	4.964	51.240	50.000	2.5
Methoxychlor	6.745	6.645	6.845	53.840	50.000	7.7
Tetrachloro-m-xylene	2.874	2.775	2.975	51.420	50.000	2.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090670.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 20:18
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:30:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	161.3E6	1531.1E6	51.789	51.420
28)	SA Decachlor...	9.066	8.062	241.2E6	1619.0E6	51.558	53.449
Target Compounds							
2)	A alpha-BHC	3.994	3.386	390.9E6	2669.8E6	54.471	52.251
3)	MA gamma-BHC...	4.324	3.722	366.7E6	2472.0E6	53.935	52.319
4)	MA Heptachlor	4.923	4.075	285.9E6	2340.7E6	51.152	52.478
5)	MB Aldrin	5.264	4.360	351.5E6	2418.9E6	53.680	52.222
6)	B beta-BHC	4.512	4.018	136.0E6	1026.7E6	51.907	51.775
7)	B delta-BHC	4.760	4.255	364.8E6	2504.8E6	53.955	52.494
8)	B Heptachlo...	5.685	4.863	308.3E6	2184.8E6	53.257	51.238
9)	A Endosulfan I	6.068	5.238	291.2E6	2057.7E6	52.601	52.240
10)	B gamma-Chl...	5.940	5.116	313.8E6	2357.5E6	52.906	52.448
11)	B alpha-Chl...	6.021	5.181	316.2E6	2261.5E6	51.166	52.424
12)	B 4,4'-DDE	6.191	5.365	283.4E6	2289.0E6	53.208	52.215
13)	MA Dieldrin	6.341	5.503	310.7E6	2321.9E6	53.285	52.551
14)	MA Endrin	6.568	5.780	264.2E6	2139.5E6	53.343	52.835
15)	B Endosulfa...	6.781	6.072	265.8E6	1967.6E6	53.168	52.391
16)	A 4,4'-DDD	6.700	5.921	216.5E6	1909.0E6	53.401	52.391
17)	MA 4,4'-DDT	7.016	6.174	216.2E6	1909.3E6	52.590	54.438
18)	B Endrin al...	6.910	6.250	193.2E6	1469.9E6	52.749	53.078
19)	B Endosulfa...	7.145	6.473	242.6E6	1892.3E6	52.302	52.790
20)	A Methoxychlor	7.489	6.745	108.1E6	938.3E6	50.867	53.839
21)	B Endrin ke...	7.625	6.983	254.4E6	1995.3E6	52.366	52.900
22)	Mirex	8.107	7.175	165.2E6	1361.0E6	51.616	52.975



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090670.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 20:18
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:30:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 161277964
Conc: 51.79 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1531136723
Conc: 51.42 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 390943449
Conc: 54.47 ng/ml



#2 alpha-BHC


R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2669827138
Conc: 52.25 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 366694320
Conc: 53.93 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 2472030526
Conc: 52.32 ng/ml




#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 285916460
Conc: 51.15 ng/ml



#4 Heptachlor


R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 2340651061
Conc: 52.48 ng/ml



#5 Aldrin


R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 351457831
Conc: 53.68 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2418877727
Conc: 52.22 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 135977995
Conc: 51.91 ng/ml



#6 beta-BHC


R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 1026671754
Conc: 51.77 ng/ml



#7 delta-BHC


R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 364828655
Conc: 53.95 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2504839647
Conc: 52.49 ng/ml




#8 Heptachlor epoxide

R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 308308743
Conc: 53.26 ng/ml



#8 Heptachlor epoxide


R.T.: 4.863 min
Delta R.T.: 0.000 min
Response: 2184755762
Conc: 51.24 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 291231404
Conc: 52.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 2057740620
Conc: 52.24 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 313795143
Conc: 52.91 ng/ml



#10 gamma-Chlordane


R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 2357471424
Conc: 52.45 ng/ml



#11 alpha-Chlordane


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 316216529
Conc: 51.17 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2261459309
Conc: 52.42 ng/ml



#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 283417707
Conc: 53.21 ng/ml




#12 4,4'-DDE

R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 2289015740
Conc: 52.22 ng/ml

Response_ Signal: PD090670.D\ECD1A.ch

#13 Dieldrin




R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 310725394
Conc: 53.29 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time Response_ Signal: PD090670.D\ECD2B.ch


#13 Dieldrin



R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 2321920761
Conc: 52.55 ng/ml

Time Response_ Signal: PD090670.D\ECD1A.ch


#14 Endrin



R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 264243558
Conc: 53.34 ng/ml

Time Response_ Signal: PD090670.D\ECD2B.ch


#14 Endrin



R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 2139494067
Conc: 52.83 ng/ml

Response_ Signal: PD090670.D\ECD1A.ch

#15 Endosulfan II




R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 265839721
Conc: 53.17 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time Response_ Signal: PD090670.D\ECD2B.ch


#15 Endosulfan II



R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1967561685
Conc: 52.39 ng/ml

Time Response_ Signal: PD090670.D\ECD1A.ch


#16 4,4'-DDD



R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 216500159
Conc: 53.40 ng/ml

Time Response_ Signal: PD090670.D\ECD2B.ch


#16 4,4'-DDD



R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1909038447
Conc: 52.39 ng/ml

Response_ Signal: PD090670.D\ECD1A.ch

#17 4,4'-DDT




R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 216168926
Conc: 52.59 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time Response_ Signal: PD090670.D\ECD2B.ch


#17 4,4'-DDT



R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1909322523
Conc: 54.44 ng/ml

Time Response_ Signal: PD090670.D\ECD1A.ch


#18 Endrin aldehyde




R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 193202358
Conc: 52.75 ng/ml

Time Response_ Signal: PD090670.D\ECD2B.ch

#18 Endrin aldehyde




R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1469875105
Conc: 53.08 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 242587440
Conc: 52.30 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.473 min
Delta R.T.: 0.000 min
Response: 1892253261
Conc: 52.79 ng/ml




#20 Methoxychlor

R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 108123063
Conc: 50.87 ng/ml



#20 Methoxychlor


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 938289155
Conc: 53.84 ng/ml



#21 Endrin ketone


R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 254440173
Conc: 52.37 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1995280178
Conc: 52.90 ng/ml




#22 Mirex

R.T.: 8.107 min
Delta R.T.: 0.000 min
Response: 165216282
Conc: 51.62 ng/ml



#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1360972275
Conc: 52.97 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.066 min

Delta R.T.: 0.001 min

Response: 241187526


Conc: 51.56 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 8.062 min

Delta R.T.: 0.000 min

Response: 1619030948

Conc: 53.45 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Continuing Calib Date: 10/18/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 02:13

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.55	3.45	3.65	0.01
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.33	4.23	4.43	0.01
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.14	7.15	7.05	7.25	0.01
4,4'-DDT	7.02	7.02	6.92	7.12	0.01
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.62	7.63	7.53	7.73	0.01
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Continuing Calib Date: 10/18/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 02:13

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.26	4.16	4.36	0.01
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.08	3.98	4.18	0.01
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.24	5.24	5.14	5.34	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.37	5.37	5.27	5.47	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.75	6.75	6.65	6.85	0.01
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.12	5.12	5.02	5.22	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090688.D **Time Analyzed:** 02:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.699	6.600	6.800	53.090	50.000	6.2
4,4'-DDE	6.189	6.090	6.290	52.010	50.000	4.0
4,4'-DDT	7.015	6.915	7.115	54.040	50.000	8.1
Aldrin	5.264	5.164	5.364	53.520	50.000	7.0
alpha-BHC	3.993	3.894	4.094	54.690	50.000	9.4
alpha-Chlordane	6.020	5.921	6.121	51.580	50.000	3.2
beta-BHC	4.511	4.412	4.612	51.940	50.000	3.9
Decachlorobiphenyl	9.064	8.964	9.164	47.910	50.000	-4.2
delta-BHC	4.759	4.660	4.860	54.270	50.000	8.5
Dieldrin	6.340	6.240	6.440	52.930	50.000	5.9
Endosulfan I	6.067	5.968	6.168	52.190	50.000	4.4
Endosulfan II	6.781	6.681	6.881	52.290	50.000	4.6
Endosulfan sulfate	7.144	7.045	7.245	49.960	50.000	-0.1
Endrin	6.568	6.468	6.668	52.810	50.000	5.6
Endrin aldehyde	6.909	6.810	7.010	52.080	50.000	4.2
Endrin ketone	7.624	7.525	7.725	49.990	50.000	0.0
gamma-BHC (Lindane)	4.324	4.225	4.425	54.150	50.000	8.3
gamma-Chlordane	5.939	5.840	6.040	52.490	50.000	5.0
Heptachlor	4.922	4.823	5.023	53.990	50.000	8.0
Heptachlor epoxide	5.684	5.584	5.784	53.100	50.000	6.2
Methoxychlor	7.488	7.388	7.588	51.500	50.000	3.0
Tetrachloro-m-xylene	3.544	3.445	3.645	52.290	50.000	4.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3363

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No.: CCAL02 Date Analyzed: 10/18/2025

Lab Sample No.: PSTDCCC050 Data File : PD090688.D Time Analyzed: 02:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.821	6.021	50.750	50.000	1.5
4,4'-DDE	5.366	5.266	5.466	51.070	50.000	2.1
4,4'-DDT	6.174	6.074	6.274	54.490	50.000	9.0
Aldrin	4.360	4.261	4.461	51.420	50.000	2.8
alpha-BHC	3.386	3.287	3.487	51.750	50.000	3.5
alpha-Chlordane	5.181	5.081	5.281	50.910	50.000	1.8
beta-BHC	4.018	3.919	4.119	50.520	50.000	1.0
Decachlorobiphenyl	8.061	7.962	8.162	44.850	50.000	-10.3
delta-BHC	4.254	4.155	4.355	51.790	50.000	3.6
Dieldrin	5.504	5.404	5.604	51.270	50.000	2.5
Endosulfan I	5.238	5.138	5.338	49.630	50.000	-0.7
Endosulfan II	6.071	5.972	6.172	50.880	50.000	1.8
Endosulfan sulfate	6.473	6.374	6.574	49.300	50.000	-1.4
Endrin	5.780	5.681	5.881	52.570	50.000	5.1
Endrin aldehyde	6.250	6.150	6.350	51.130	50.000	2.3
Endrin ketone	6.982	6.883	7.083	48.210	50.000	-3.6
gamma-BHC (Lindane)	3.722	3.623	3.823	51.710	50.000	3.4
gamma-Chlordane	5.116	5.017	5.217	51.230	50.000	2.5
Heptachlor	4.074	3.975	4.175	51.800	50.000	3.6
Heptachlor epoxide	4.863	4.764	4.964	50.240	50.000	0.5
Methoxychlor	6.745	6.645	6.845	52.120	50.000	4.2
Tetrachloro-m-xylene	2.874	2.775	2.975	51.130	50.000	2.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090688.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2025 02:13
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:35:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	162.8E6	1522.4E6	52.293	51.128
28)	SA Decachlor...	9.064	8.061	224.1E6	1358.5E6	47.912	44.848
Target Compounds							
2)	A alpha-BHC	3.993	3.386	392.5E6	2644.4E6	54.686	51.754
3)	MA gamma-BHC...	4.324	3.722	368.2E6	2443.3E6	54.150	51.711
4)	MA Heptachlor	4.922	4.074	301.8E6	2310.4E6	53.995	51.800
5)	MB Aldrin	5.264	4.360	350.4E6	2381.8E6	53.518	51.423
6)	B beta-BHC	4.511	4.018	136.1E6	1001.7E6	51.942	50.516
7)	B delta-BHC	4.759	4.254	366.9E6	2471.2E6	54.266	51.788
8)	B Heptachlo...	5.684	4.863	307.4E6	2142.3E6	53.101	50.243
9)	A Endosulfan I	6.067	5.238	289.0E6	1954.8E6	52.190	49.628
10)	B gamma-Chl...	5.939	5.116	311.3E6	2302.8E6	52.493	51.233
11)	B alpha-Chl...	6.020	5.181	318.8E6	2196.2E6	51.583	50.912
12)	B 4,4'-DDE	6.189	5.366	277.0E6	2239.0E6	52.012	51.075
13)	MA Dieldrin	6.340	5.504	308.7E6	2265.4E6	52.931	51.271
14)	MA Endrin	6.568	5.780	261.6E6	2128.7E6	52.807	52.569
15)	B Endosulfa...	6.781	6.071	261.5E6	1910.9E6	52.293	50.883
16)	A 4,4'-DDD	6.699	5.920	215.2E6	1849.3E6	53.091	50.752
17)	MA 4,4'-DDT	7.015	6.174	222.1E6	1911.2E6	54.036	54.491
18)	B Endrin al...	6.909	6.250	190.7E6	1415.9E6	52.077	51.129
19)	B Endosulfa...	7.144	6.473	231.7E6	1767.2E6	49.957	49.302
20)	A Methoxychlor	7.488	6.745	109.5E6	908.3E6	51.500	52.118
21)	B Endrin ke...	7.624	6.982	242.9E6	1818.5E6	49.991	48.213
22)	Mirex	8.108	7.175	153.9E6	1192.2E6	48.074	46.405



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090688.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2025 02:13
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:35:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 162847718
Conc: 52.29 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1522440144
Conc: 51.13 ng/ml



#2 alpha-BHC

R.T.: 3.993 min
Delta R.T.: 0.000 min
Response: 392491238
Conc: 54.69 ng/ml




#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2644430290
Conc: 51.75 ng/ml

Response_ Signal: PD090688.D\ECD1A.ch

#3 gamma-BHC (Lindane)




R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 368155481
Conc: 54.15 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time Response_ Signal: PD090688.D\ECD2B.ch


#3 gamma-BHC (Lindane)



R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 2443301154
Conc: 51.71 ng/ml

Time Response_ Signal: PD090688.D\ECD1A.ch


#4 Heptachlor



R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 301805703
Conc: 53.99 ng/ml

Time Response_ Signal: PD090688.D\ECD2B.ch

#4 Heptachlor



R.T.: 4.074 min
Delta R.T.: 0.000 min
Response: 2310429151
Conc: 51.80 ng/ml

Response_ Signal: PD090688.D\ECD1A.ch

#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 350399128
Conc: 53.52 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time

Response_ Signal: PD090688.D\ECD2B.ch

#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2381845649
Conc: 51.42 ng/ml

Time

Response_ Signal: PD090688.D\ECD1A.ch

#6 beta-BHC

R.T.: 4.511 min
Delta R.T.: -0.001 min
Response: 136069610
Conc: 51.94 ng/ml


Time

Response_ Signal: PD090688.D\ECD2B.ch

#6 beta-BHC

R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 1001709776
Conc: 50.52 ng/ml


Time



#7 delta-BHC


R.T.: 4.759 min
Delta R.T.: -0.001 min
Response: 366932996
Conc: 54.27 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.254 min
Delta R.T.: 0.000 min
Response: 2471159561
Conc: 51.79 ng/ml




#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 307406481
Conc: 53.10 ng/ml



#8 Heptachlor epoxide


R.T.: 4.863 min
Delta R.T.: 0.000 min
Response: 2142316710
Conc: 50.24 ng/ml



#9 Endosulfan I


R.T.: 6.067 min
Delta R.T.: 0.000 min
Response: 288956165
Conc: 52.19 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1954846515
Conc: 49.63 ng/ml




#10 gamma-Chlordane

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 311347987
Conc: 52.49 ng/ml



#10 gamma-Chlordane


R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 2302844676
Conc: 51.23 ng/ml



#11 alpha-Chlordane


R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 318798736
Conc: 51.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2196243414
Conc: 50.91 ng/ml




#12 4,4'-DDE

R.T.: 6.189 min
Delta R.T.: 0.000 min
Response: 277049157
Conc: 52.01 ng/ml



#12 4,4'-DDE


R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 2239036388
Conc: 51.07 ng/ml



#13 Dieldrin


R.T.: 6.340 min
Delta R.T.: 0.000 min
Response: 308661900
Conc: 52.93 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 2265378080
Conc: 51.27 ng/ml




#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 261592017
Conc: 52.81 ng/ml



#14 Endrin


R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 2128726920
Conc: 52.57 ng/ml



#15 Endosulfan II


R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 261463471
Conc: 52.29 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.071 min
Delta R.T.: 0.000 min
Response: 1910945675
Conc: 50.88 ng/ml



#16 4,4'-DDD

R.T.: 6.699 min
Delta R.T.: 0.000 min
Response: 215242196
Conc: 53.09 ng/ml



#16 4,4'-DDD


R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 1849297090
Conc: 50.75 ng/ml

Response_ Signal: PD090688.D\ECD1A.ch

#17 4,4'-DDT

R.T.: 7.015 min
Delta R.T.: 0.000 min
Response: 222115125
Conc: 54.04 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDCCC050



Time Response_ Signal: PD090688.D\ECD2B.ch

#17 4,4'-DDT


R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1911167341
Conc: 54.49 ng/ml



Time Response_ Signal: PD090688.D\ECD1A.ch

#18 Endrin aldehyde


R.T.: 6.909 min
Delta R.T.: 0.000 min
Response: 190743549
Conc: 52.08 ng/ml



Time Response_ Signal: PD090688.D\ECD2B.ch

#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1415896707
Conc: 51.13 ng/ml




Response_ Signal: PD090688.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 231711199
Conc: 49.96 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDCCC050



Time Response_ Signal: PD090688.D\ECD2B.ch

#19 Endosulfan Sulfate


R.T.: 6.473 min
Delta R.T.: 0.000 min
Response: 1767241268
Conc: 49.30 ng/ml



Time Response_ Signal: PD090688.D\ECD1A.ch

#20 Methoxychlor


R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 109466607
Conc: 51.50 ng/ml




Time Response_ Signal: PD090688.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 908306281
Conc: 52.12 ng/ml






#21 Endrin ketone


R.T.: 7.624 min
Delta R.T.: 0.000 min
Response: 242900927
Conc: 49.99 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.982 min
Delta R.T.: 0.000 min
Response: 1818499620
Conc: 48.21 ng/ml



#22 Mirex

R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 153877408
Conc: 48.07 ng/ml




#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1192189722
Conc: 46.41 ng/ml

Signal: PD090688.D\ECD1A.ch

#28 Decachlorobiphenyl




R.T.: 9.064 min
Delta R.T.: 0.000 min
Response: 224128569
Conc: 47.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Signal: PD090688.D\ECD2B.ch

#28 Decachlorobiphenyl



R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 1358508029
Conc: 44.85 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.066	8.970	9.170	22.860	20.000	14.3
Tetrachloro-m-xylene	3.544	3.490	3.590	21.710	20.000	8.6
alpha-BHC	3.994	3.940	4.040	9.050	10.000	-9.5
beta-BHC	4.512	4.460	4.560	10.070	10.000	0.7
gamma-BHC (Lindane)	4.324	4.270	4.370	9.330	10.000	-6.7
Endrin	6.568	6.500	6.640	49.200	50.000	-1.6
4,4'-DDT	7.016	6.950	7.090	100.270	100.000	0.3
Methoxychlor	7.489	7.420	7.560	236.810	250.000	-5.3

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.062	7.960	8.160	21.400	20.000	7.0
Tetrachloro-m-xylene	2.875	2.820	2.930	21.220	20.000	6.1
alpha-BHC	3.386	3.340	3.440	10.300	10.000	3.0
beta-BHC	4.019	3.970	4.070	10.460	10.000	4.6
gamma-BHC (Lindane)	3.723	3.670	3.770	10.340	10.000	3.4
Endrin	5.781	5.710	5.850	47.940	50.000	-4.1
4,4'-DDT	6.174	6.100	6.240	95.670	100.000	-4.3
Methoxychlor	6.745	6.670	6.820	201.830	250.000	-19.3

PEM
Data File: PD090648.D Date Acquired 10/17/2025 11:07
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	243728295.7	249054181.9	5325886.17	2.14
Endrin aldehyde	6.91	1372426.584			
Endrin ketone	7.63	3953459.588			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.78	1941282940	1991060515	49777575.7	2.50
Endrin aldehyde #2	6.25	19940698.04			
Endrin ketone #2	6.98	29836877.65			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	412144601.3	415215115.8	3070514.5	0.74
4,4'-DDE	0.00	0			
4,4'-DDD	6.70	3070514.503			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.17	3355351424	3394188860	38837436.1	1.14
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.92	38837436.09			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090648.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 11:07
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 10/19/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:04:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:02:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.875	67593250	632.0E6	21.705	21.224
28)	SA Decachlor...	9.066	8.062	106.9E6	648.1E6	22.859	21.395
Target Compounds							
2)	A alpha-BHC	3.994	3.386	64971308	526.5E6	9.053	10.304
3)	MA gamma-BHC...	4.324	3.723	63399907	488.4E6	9.325	10.337
6)	B beta-BHC	4.512	4.019	26386693	207.4E6	10.073	10.459
14)	MA Endrin	6.568	5.781	243.7E6	1941.3E6	49.201	47.940
16)	A 4,4'-DDD	6.701	5.923	3070515	38837436	0.757	1.066 #
17)	MA 4,4'-DDT	7.016	6.174	412.1E6	3355.4E6	100.267	95.667
18)	B Endrin al...	6.913	6.248	1372427	19940698	0.375m	0.720 #
20)	A Methoxychlor	7.489	6.745	503.3E6	3517.5E6	236.805	201.834
21)	B Endrin ke...	7.625	6.981	3953460	29836878	0.814	0.791m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090648.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:07
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:04:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 67593250
Conc: 21.71 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 631992202
Conc: 21.22 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 64971308
Conc: 9.05 ng/ml



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 526499291
Conc: 10.30 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 63399907
Conc: 9.33 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#3 gamma-BHC (Lindane)

R.T.: 3.723 min
Delta R.T.: 0.000 min
Response: 488402734
Conc: 10.34 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 26386693
Conc: 10.07 ng/ml



#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 207399327
Conc: 10.46 ng/ml




#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 243728296
Conc: 49.20 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#14 Endrin

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 1941282940
Conc: 47.94 ng/ml




#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: 0.000 min
Response: 3070515
Conc: 0.76 ng/ml



#16 4,4'-DDD

R.T.: 5.923 min
Delta R.T.: 0.003 min
Response: 38837436
Conc: 1.07 ng/ml



#17 4,4'-DDT

R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 412144601
Conc: 100.27 ng/ml

Instrument :

ECD_D

ClientSampleId :


PEM

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 3355351424
Conc: 95.67 ng/ml




#18 Endrin aldehyde

R.T.: 6.913 min
Delta R.T.: 0.003 min
Response: 1372427
Conc: 0.37 ng/ml m



#18 Endrin aldehyde

R.T.: 6.248 min
Delta R.T.: -0.002 min
Response: 19940698
Conc: 0.72 ng/ml



#20 Methoxychlor

R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 503349948
Conc: 236.81 ng/ml

Instrument :

ECD_D

ClientSampleId :


PEM

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025




#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 3517519034
Conc: 201.83 ng/ml




#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 3953460
Conc: 0.81 ng/ml



#21 Endrin ketone

R.T.: 6.981 min
Delta R.T.: -0.002 min
Response: 29836878
Conc: 0.79 ng/ml m




#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 106933782
Conc: 22.86 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 648085268
Conc: 21.40 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090669.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.065	8.960	9.170	23.950	20.000	19.8
Tetrachloro-m-xylene	3.544	3.490	3.590	22.660	20.000	13.3
alpha-BHC	3.993	3.940	4.040	9.630	10.000	-3.7
beta-BHC	4.512	4.460	4.560	10.740	10.000	7.4
gamma-BHC (Lindane)	4.324	4.270	4.370	9.940	10.000	-0.6
Endrin	6.568	6.500	6.640	52.910	50.000	5.8
4,4'-DDT	7.016	6.950	7.090	103.900	100.000	3.9
Methoxychlor	7.489	7.420	7.560	242.540	250.000	-3.0

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090669.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.061	7.960	8.160	24.170	20.000	20.9
Tetrachloro-m-xylene	2.874	2.820	2.920	22.460	20.000	12.3
alpha-BHC	3.386	3.340	3.440	10.980	10.000	9.8
beta-BHC	4.019	3.970	4.070	11.300	10.000	13.0
gamma-BHC (Lindane)	3.723	3.670	3.770	11.120	10.000	11.2
Endrin	5.781	5.710	5.850	53.090	50.000	6.2
4,4'-DDT	6.175	6.100	6.250	107.030	100.000	7.0
Methoxychlor	6.746	6.680	6.820	223.020	250.000	-10.8

Data File: PEM
 PD090669.D Date Acquired 10/17/2025 20:04
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	262109893.9	266219540.1	4109646.21	Down 1.54
Endrin aldehyde	6.91	958907.135			
Endrin ketone	7.62	3150739.072			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.78	2149779887	2200728231	50948344.2	2.32
Endrin aldehyde #2	6.25	20792419.64			
Endrin ketone #2	6.98	30155924.53			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	427080097.1	429084732.8	2004635.72	0.47
4,4'-DDE	0.00	0			
4,4'-DDD	6.70	2004635.721			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.17	3754007771	3787186925	33179154.1	0.88
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.93	33179154.12			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090669.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 20:04
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 10/21/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:29:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	70552955	668.8E6	22.655	22.461
28)	SA Decachlor...	9.065	8.061	112.0E6	732.0E6	23.948	24.167
Target Compounds							
2)	A alpha-BHC	3.993	3.386	69124953	561.1E6	9.631	10.980
3)	MA gamma-BHC...	4.324	3.723	67550443	525.3E6	9.936	11.117
6)	B beta-BHC	4.512	4.019	28133405	224.0E6	10.739	11.299
14)	MA Endrin	6.568	5.781	262.1E6	2149.8E6	52.912	53.089
16)	A 4,4'-DDD	6.702	5.926	2004636	33179154	0.494m	0.911m#
17)	MA 4,4'-DDT	7.016	6.175	427.1E6	3754.0E6	103.901	107.034
18)	B Endrin al...	6.910	6.248	958907	20792420	0.262m	0.751 #
20)	A Methoxychlor	7.489	6.746	515.5E6	3886.8E6	242.543	223.020
21)	B Endrin ke...	7.625	6.983	3150739	30155925	0.648	0.800

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090669.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 20:04
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/21/2025
Supervised By : mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:29:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 70552955
Conc: 22.66 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 668820442
Conc: 22.46 ng/ml




#2 alpha-BHC

R.T.: 3.993 min
Delta R.T.: 0.000 min
Response: 69124953
Conc: 9.63 ng/ml



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 561054467
Conc: 10.98 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 67550443
Conc: 9.94 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#3 gamma-BHC (Lindane)

R.T.: 3.723 min
Delta R.T.: 0.000 min
Response: 525287565
Conc: 11.12 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 28133405
Conc: 10.74 ng/ml



#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 224046983
Conc: 11.30 ng/ml




#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 262109894
Conc: 52.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#14 Endrin

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 2149779887
Conc: 53.09 ng/ml




#16 4,4'-DDD

R.T.: 6.702 min
Delta R.T.: 0.002 min
Response: 2004636
Conc: 0.49 ng/ml m



#16 4,4'-DDD

R.T.: 5.926 min
Delta R.T.: 0.006 min
Response: 33179154
Conc: 0.91 ng/ml m




#17 4,4'-DDT

R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 427080097
Conc: 103.90 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#17 4,4'-DDT

R.T.: 6.175 min
Delta R.T.: 0.000 min
Response: 3754007771
Conc: 107.03 ng/ml




#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 958907
Conc: 0.26 ng/ml m



#18 Endrin aldehyde

R.T.: 6.248 min
Delta R.T.: -0.002 min
Response: 20792420
Conc: 0.75 ng/ml




#20 Methoxychlor

R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 515545040
Conc: 242.54 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#20 Methoxychlor

R.T.: 6.746 min
Delta R.T.: 0.000 min
Response: 3886751103
Conc: 223.02 ng/ml




#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 3150739
Conc: 0.65 ng/ml



#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 30155925
Conc: 0.80 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.001 min
Response: 112029810
Conc: 23.95 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 732040565
Conc: 24.17 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090649.D
Acq On : 17 Oct 2025 11:20
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Title : GC Extractables
Last Update : Fri Oct 17 17:43:28 2025
Integrator: ChemStation

RT#1	RT#2	Resolution
3.544	5.940	100.00%
5.940	6.068	100.00%
6.068	6.190	100.00%
6.190	6.341	100.00%
6.341	7.144	100.00%
7.144	7.488	100.00%
7.488	7.625	100.00%
7.625	9.066	100.00%

Signal #2

2.874	5.116	100.00%
5.116	5.238	100.00%
5.238	5.366	100.00%
5.366	5.503	100.00%
5.503	6.474	100.00%
6.474	6.745	100.00%
6.745	6.983	100.00%
6.983	8.062	100.00%

PD101725.M Sat Oct 25 06:10:53 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090649.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 11:20
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESCHK

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 10/19/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:04:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:02:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.874	55125186	545.1E6	17.701m	18.307
28)	SA Decachlor...	9.066	8.062	88858238	547.9E6	18.995	18.088
Target Compounds							
9)	A Endosulfan I	6.068	5.238	43926319	330.7E6	7.934	8.397
10)	B gamma-Chl...	5.940	5.116	47696248	387.3E6	8.042	8.616
12)	B 4,4'-DDE	6.190	5.366	86477351	754.1E6	16.235	17.202
13)	MA Dieldrin	6.341	5.503	92552800	748.0E6	15.872	16.930
19)	B Endosulfa...	7.144	6.474	76754913	612.1E6	16.548	17.078
20)	A Methoxychlor	7.488	6.745	169.7E6	1331.6E6	79.844	76.409
21)	B Endrin ke...	7.625	6.983	80182457	638.1E6	16.502	16.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090649.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:20
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

RESCHK

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:04:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: -0.002 min
Response: 55125186
Conc: 17.70 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 545126372
Conc: 18.31 ng/ml




#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 43926319
Conc: 7.93 ng/ml



#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 330745194
Conc: 8.40 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 47696248
Conc: 8.04 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#10 gamma-Chlordane

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 387274886
Conc: 8.62 ng/ml




#12 4,4'-DDE

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 86477351
Conc: 16.23 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 754114700
Conc: 17.20 ng/ml




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 92552800
Conc: 15.87 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#13 Dieldrin

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 748035470
Conc: 16.93 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 76754913
Conc: 16.55 ng/ml



#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 612147653
Conc: 17.08 ng/ml



#20 Methoxychlor

R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 169715831
Conc: 79.84 ng/ml


Instrument :

ECD_D

ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 1331648566
Conc: 76.41 ng/ml




#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 80182457
Conc: 16.50 ng/ml



#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 638061752
Conc: 16.92 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.001 min
Response: 88858238
Conc: 19.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 547910430
Conc: 18.09 ng/ml

Analytical Sequence

Client: AECOM	SDG No.: Q3363
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 6069.	Instrument ID: ECD_D
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 10/17/2025 10/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
LBLK	LBLK	10/17/2025	10:53	PD090647.D	9.07	3.55
PEM	PEM	10/17/2025	11:07	PD090648.D	9.07	3.54
RESCHK	RESCHK	10/17/2025	11:20	PD090649.D	9.07	3.54
PSTDICC100	PSTDICC100	10/17/2025	11:34	PD090650.D	9.07	3.54
PSTDICC075	PSTDICC075	10/17/2025	11:48	PD090651.D	9.07	3.55
PSTDICC050	PSTDICC050	10/17/2025	12:01	PD090652.D	9.06	3.55
PSTDICC025	PSTDICC025	10/17/2025	12:15	PD090653.D	9.07	3.54
PSTDICC005	PSTDICC005	10/17/2025	12:28	PD090654.D	9.07	3.54
PCHLORICC500	PCHLORICC500	10/17/2025	13:09	PD090657.D	9.07	3.55
PTOXICC500	PTOXICC500	10/17/2025	14:17	PD090662.D	9.07	3.55
LBLK	LBLK	10/17/2025	19:50	PD090668.D	9.07	3.54
PEM	PEM	10/17/2025	20:04	PD090669.D	9.07	3.54
PSTDCCC050	PSTDCCC050	10/17/2025	20:18	PD090670.D	9.07	3.54
PB170142BL	PB170142BL	10/17/2025	20:31	PD090671.D	9.07	3.54
PB170142BS	PB170142BS	10/17/2025	20:45	PD090672.D	9.07	3.54
PR132-SO1-085011-20251015	Q3363-01	10/17/2025	22:48	PD090678.D	9.07	3.54
PR132-SO1-039085-20251015	Q3363-02	10/17/2025	23:02	PD090679.D	9.07	3.54
PR132-SO3-028010-20251015	Q3363-03	10/17/2025	23:15	PD090680.D	9.07	3.54
PR132-SO3-010013-20251015	Q3363-04	10/17/2025	23:29	PD090681.D	9.07	3.54
PR132-SO3-010013-20251015MS	Q3363-04MS	10/17/2025	23:43	PD090682.D	9.07	3.54
PR132-SO3-010013-20251015MSD	Q3363-04MSD	10/17/2025	23:56	PD090683.D	9.07	3.54
COMP01-20251015	Q3363-05	10/18/2025	00:10	PD090684.D	9.07	3.54
LBLK	LBLK	10/18/2025	00:51	PD090687.D	9.07	3.54
PSTDCCC050	PSTDCCC050	10/18/2025	02:13	PD090688.D	9.06	3.54

Analytical Sequence

Client: AECOM	SDG No.: Q3363
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 6069.	Instrument ID: ECD_D
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 10/17/2025 10/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/17/2025	10:53	PD090647.D	8.06	2.88
PEM	PEM	10/17/2025	11:07	PD090648.D	8.06	2.88
RESCHK	RESCHK	10/17/2025	11:20	PD090649.D	8.06	2.87
PSTDICC100	PSTDICC100	10/17/2025	11:34	PD090650.D	8.06	2.87
PSTDICC075	PSTDICC075	10/17/2025	11:48	PD090651.D	8.06	2.88
PSTDICC050	PSTDICC050	10/17/2025	12:01	PD090652.D	8.06	2.88
PSTDICC025	PSTDICC025	10/17/2025	12:15	PD090653.D	8.06	2.87
PSTDICC005	PSTDICC005	10/17/2025	12:28	PD090654.D	8.06	2.87
PCHLORICC500	PCHLORICC500	10/17/2025	13:09	PD090657.D	8.06	2.87
PTOXICC500	PTOXICC500	10/17/2025	14:17	PD090662.D	8.06	2.88
IBLK	IBLK	10/17/2025	19:50	PD090668.D	8.06	2.87
PEM	PEM	10/17/2025	20:04	PD090669.D	8.06	2.87
PSTDCCC050	PSTDCCC050	10/17/2025	20:18	PD090670.D	8.06	2.87
PB170142BL	PB170142BL	10/17/2025	20:31	PD090671.D	8.06	2.87
PB170142BS	PB170142BS	10/17/2025	20:45	PD090672.D	8.06	2.87
PR132-SO1-085011-20251015	Q3363-01	10/17/2025	22:48	PD090678.D	8.06	2.87
PR132-SO1-039085-20251015	Q3363-02	10/17/2025	23:02	PD090679.D	8.06	2.87
PR132-SO3-028010-20251015	Q3363-03	10/17/2025	23:15	PD090680.D	8.06	2.87
PR132-SO3-010013-20251015	Q3363-04	10/17/2025	23:29	PD090681.D	8.06	2.88
PR132-SO3-010013-20251015MS	Q3363-04MS	10/17/2025	23:43	PD090682.D	8.06	2.88
PR132-SO3-010013-20251015MSD	Q3363-04MSD	10/17/2025	23:56	PD090683.D	8.06	2.88
COMP01-20251015	Q3363-05	10/18/2025	00:10	PD090684.D	8.06	2.87
IBLK	IBLK	10/18/2025	00:51	PD090687.D	8.06	2.87
PSTDCCC050	PSTDCCC050	10/18/2025	02:13	PD090688.D	8.06	2.87

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170142BS

Lab Name: Alliance

Contract: AEEO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: PB170142BS

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	15.6	1.3
	2	5.92	5.87	5.97	15.8	
4,4'-DDE	1	6.19	6.14	6.24	15.5	1.3
	2	5.37	5.32	5.42	15.7	
4,4'-DDT	1	7.02	6.97	7.07	15.4	5.1
	2	6.17	6.12	6.22	16.2	
alpha-BHC	1	3.99	3.94	4.04	15.4	0.6
	2	3.39	3.34	3.44	15.5	
Aldrin	1	5.26	5.21	5.31	15.6	0.6
	2	4.36	4.31	4.41	15.7	
alpha-Chlordane	1	6.02	5.97	6.07	16.0	1.9
	2	5.18	5.13	5.23	15.7	
Endosulfan II	1	6.78	6.73	6.83	15.6	1.3
	2	6.07	6.02	6.12	15.8	
Endosulfan sulfate	1	7.15	7.10	7.20	15.4	3.2
	2	6.47	6.42	6.52	15.9	
beta-BHC	1	4.51	4.46	4.56	15.0	3.3
	2	4.02	3.97	4.07	15.5	
delta-BHC	1	4.76	4.71	4.81	15.4	1.3
	2	4.26	4.21	4.31	15.6	
Endosulfan I	1	6.07	6.02	6.12	15.5	1.9
	2	5.24	5.19	5.29	15.8	
Dieldrin	1	6.34	6.29	6.39	15.6	1.3
	2	5.50	5.45	5.55	15.8	
Endrin aldehyde	1	6.91	6.86	6.96	15.4	2.6
	2	6.25	6.20	6.30	15.8	
Methoxychlor	1	7.49	7.44	7.54	15.0	5.8
	2	6.75	6.70	6.80	15.9	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170142BS

Lab Name: Alliance

Contract: AEEO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: PB170142BS

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin ketone	1	7.63	7.58	7.68	15.5	2.5
	2	6.98	6.93	7.03	15.9	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	15.4	1.3
	2	3.72	3.67	3.77	15.6	
Heptachlor	1	4.92	4.87	4.97	14.7	6.6
	2	4.07	4.02	4.12	15.7	
Heptachlor epoxide	1	5.68	5.63	5.73	15.6	0
	2	4.86	4.81	4.91	15.6	
gamma-Chlordane	1	5.94	5.89	5.99	15.4	1.9
	2	5.12	5.07	5.17	15.7	
Endrin	1	6.57	6.52	6.62	15.7	1.9
	2	5.78	5.73	5.83	16.0	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PR132-SO1-039085-20251015

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3363-02

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	5.70	37.5
	2	5.92	5.87	5.97	3.90	
Endosulfan sulfate	1	7.13	7.08	7.18	0.38	47.9
	2	6.47	6.42	6.52	0.62	
alpha-Chlordane	1	6.02	5.97	6.07	0.49	3.5
	2	5.18	5.13	5.23	0.51	
4,4'-DDE	1	6.19	6.14	6.24	9.70	45.6
	2	5.37	5.32	5.42	6.10	
Endrin	1	6.57	6.52	6.62	1.40	120.5
	2	5.78	5.73	5.83	0.35	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PR132-SO3-010013-20251015MS

Lab Name: Alliance

Contract: AEEO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3363-04MS

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.78	6.73	6.83	26.8	10.6
	2	6.07	6.02	6.12	24.1	
4,4'-DDD	1	6.70	6.65	6.75	27.0	20
	2	5.92	5.87	5.97	22.1	
4,4'-DDT	1	7.02	6.97	7.07	27.1	12.5
	2	6.17	6.12	6.22	23.9	
Endrin aldehyde	1	6.91	6.86	6.96	27.0	13.4
	2	6.25	6.20	6.30	23.6	
Endosulfan sulfate	1	7.14	7.09	7.19	26.3	11.7
	2	6.47	6.42	6.52	23.4	
Methoxychlor	1	7.49	7.44	7.54	26.9	13.1
	2	6.75	6.70	6.80	23.6	
Endrin ketone	1	7.63	7.58	7.68	26.8	21
	2	6.98	6.93	7.03	21.7	
alpha-BHC	1	3.99	3.94	4.04	29.0	12.1
	2	3.39	3.34	3.44	25.7	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	28.5	11.1
	2	3.72	3.67	3.77	25.5	
Heptachlor	1	4.92	4.87	4.97	27.7	5.2
	2	4.07	4.02	4.12	26.3	
Aldrin	1	5.26	5.21	5.31	28.9	13.7
	2	4.36	4.31	4.41	25.2	
beta-BHC	1	4.51	4.46	4.56	24.9	0.8
	2	4.02	3.97	4.07	24.7	
delta-BHC	1	4.76	4.71	4.81	29.0	12.5
	2	4.26	4.21	4.31	25.6	
Heptachlor epoxide	1	5.68	5.63	5.73	28.0	8.2
	2	4.86	4.81	4.91	25.8	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PR132-SO3-010013-20251015MS

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3363-04MS

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.07	6.02	6.12	27.5	20.4
	2	5.24	5.19	5.29	22.4	
gamma-Chlordane	1	5.94	5.89	5.99	27.1	4.1
	2	5.12	5.07	5.17	26.0	
alpha-Chlordane	1	6.02	5.97	6.07	26.6	5.4
	2	5.18	5.13	5.23	25.2	
4,4'-DDE	1	6.19	6.14	6.24	25.3	9.1
	2	5.37	5.32	5.42	23.1	
Dieldrin	1	6.34	6.29	6.39	27.1	6.9
	2	5.50	5.45	5.55	25.3	
Endrin	1	6.57	6.52	6.62	27.8	12.6
	2	5.78	5.73	5.83	24.5	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PR132-SO3-010013-20251015MSI

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3363-04MSD

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.78	6.73	6.83	28.0	13.7
	2	6.07	6.02	6.12	24.4	
4,4'-DDD	1	6.70	6.65	6.75	28.4	23.6
	2	5.92	5.87	5.97	22.4	
4,4'-DDT	1	7.02	6.97	7.07	28.2	14.9
	2	6.17	6.12	6.22	24.3	
Endrin aldehyde	1	6.91	6.86	6.96	28.2	16.1
	2	6.25	6.20	6.30	24.0	
Endosulfan sulfate	1	7.15	7.10	7.20	27.5	15.7
	2	6.47	6.42	6.52	23.5	
Methoxychlor	1	7.49	7.44	7.54	27.7	13.5
	2	6.75	6.70	6.80	24.2	
Endrin ketone	1	7.63	7.58	7.68	28.4	26.7
	2	6.98	6.93	7.03	21.7	
alpha-BHC	1	3.99	3.94	4.04	30.2	14.9
	2	3.39	3.34	3.44	26.0	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	29.6	14.1
	2	3.72	3.67	3.77	25.7	
Heptachlor	1	4.92	4.87	4.97	28.7	7.2
	2	4.08	4.03	4.13	26.7	
Aldrin	1	5.26	5.21	5.31	29.8	14.8
	2	4.36	4.31	4.41	25.7	
beta-BHC	1	4.51	4.46	4.56	25.8	3.1
	2	4.02	3.97	4.07	25.0	
delta-BHC	1	4.76	4.71	4.81	30.0	15.1
	2	4.26	4.21	4.31	25.8	
Heptachlor epoxide	1	5.68	5.63	5.73	29.0	12.5
	2	4.86	4.81	4.91	25.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PR132-SO3-010013-20251015MSI

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3363-04MSD

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.07	6.02	6.12	28.8	23.3
	2	5.24	5.19	5.29	22.8	
gamma-Chlordane	1	5.94	5.89	5.99	28.3	6.9
	2	5.12	5.07	5.17	26.4	
alpha-Chlordane	1	6.02	5.97	6.07	27.6	7.9
	2	5.18	5.13	5.23	25.5	
4,4'-DDE	1	6.19	6.14	6.24	26.6	12.4
	2	5.37	5.32	5.42	23.5	
Dieldrin	1	6.34	6.29	6.39	28.3	8.9
	2	5.50	5.45	5.55	25.9	
Endrin	1	6.57	6.52	6.62	29.1	15.6
	2	5.78	5.73	5.83	24.9	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PR132-SO3-028010-20251015

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3363-03

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	6.80	10.9
	2	5.92	5.87	5.97	6.10	
gamma-Chlordane	1	5.94	5.89	5.99	0.38	104.5
	2	5.12	5.07	5.17	1.20	
alpha-Chlordane	1	6.02	5.97	6.07	0.59	8.6
	2	5.18	5.13	5.23	0.54	
4,4'-DDE	1	6.19	6.14	6.24	4.40	28.6
	2	5.36	5.31	5.41	3.30	
Dieldrin	1	6.34	6.29	6.39	0.36	89.4
	2	5.50	5.45	5.55	0.95	
Endrin	1	6.57	6.52	6.62	0.47	45.7
	2	5.78	5.73	5.83	0.30	



QC SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM		Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206		Date Received:	
Client Sample ID:	PB170142BL		SDG No.:	Q3363
Lab Sample ID:	PB170142BL		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	100
Sample Wt/Vol:	30.01 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541B	Prep Date	10/17/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	1.70	U	1	0.13	1.70	ug/kg	10/17/25 20:31	PB170142
319-85-7	beta-BHC	1.70	U	1	0.18	1.70	ug/kg	10/17/25 20:31	PB170142
319-86-8	delta-BHC	1.70	U	1	0.39	1.70	ug/kg	10/17/25 20:31	PB170142
58-89-9	gamma-BHC (Lindane)	1.70	U	1	0.14	1.70	ug/kg	10/17/25 20:31	PB170142
76-44-8	Heptachlor	1.70	U	1	0.12	1.70	ug/kg	10/17/25 20:31	PB170142
309-00-2	Aldrin	1.70	U	1	0.12	1.70	ug/kg	10/17/25 20:31	PB170142
1024-57-3	Heptachlor epoxide	1.70	U	1	0.19	1.70	ug/kg	10/17/25 20:31	PB170142
959-98-8	Endosulfan I	1.70	U	1	0.14	1.70	ug/kg	10/17/25 20:31	PB170142
60-57-1	Dieldrin	1.70	U	1	0.14	1.70	ug/kg	10/17/25 20:31	PB170142
72-55-9	4,4-DDE	1.70	U	1	0.14	1.70	ug/kg	10/17/25 20:31	PB170142
72-20-8	Endrin	1.70	U	1	0.14	1.70	ug/kg	10/17/25 20:31	PB170142
33213-65-9	Endosulfan II	1.70	U	1	0.29	1.70	ug/kg	10/17/25 20:31	PB170142
72-54-8	4,4-DDD	1.70	U	1	0.15	1.70	ug/kg	10/17/25 20:31	PB170142
1031-07-8	Endosulfan Sulfate	1.70	U	1	0.13	1.70	ug/kg	10/17/25 20:31	PB170142
50-29-3	4,4-DDT	1.70	U	1	0.14	1.70	ug/kg	10/17/25 20:31	PB170142
72-43-5	Methoxychlor	1.70	U	1	0.37	1.70	ug/kg	10/17/25 20:31	PB170142
53494-70-5	Endrin ketone	1.70	U	1	0.19	1.70	ug/kg	10/17/25 20:31	PB170142
7421-93-4	Endrin aldehyde	1.70	U	1	0.37	1.70	ug/kg	10/17/25 20:31	PB170142
5103-71-9	alpha-Chlordane	1.70	U	1	0.12	1.70	ug/kg	10/17/25 20:31	PB170142
5103-74-2	gamma-Chlordane	1.70	U	1	0.15	1.70	ug/kg	10/17/25 20:31	PB170142
8001-35-2	Toxaphene	33.0	U	1	5.40	33.0	ug/kg	10/17/25 20:31	PB170142
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.5			20 - 144	103%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.0			19 - 148	95%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090671.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 20:31
Operator : AR\AJ
Sample : PB170142BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170142BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:30:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	58078094	565.6E6	18.650	18.995
28)	SA Decachlor...	9.065	8.062	96076687	621.6E6	20.538	20.521

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090671.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 20:31
Operator : AR\AJ
Sample : PB170142BL
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PB170142BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:30:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090671.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 58078094
Conc: 18.65 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170142BL


Response_ Signal: PD090671.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 565608356
Conc: 18.99 ng/ml


Response_ Signal: PD090671.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 96076687
Conc: 20.54 ng/ml

Response_ Signal: PD090671.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 621623695
Conc: 20.52 ng/ml

Report of Analysis

Client:	AECOM		Date Collected:	10/17/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206		Date Received:	10/17/25
Client Sample ID:	PIBLK-PD090647.D		SDG No.:	Q3363
Lab Sample ID:	I.BLK-PD090647.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.050	U	1	0.0039	0.050	ug/L	10/17/25	pd101725
319-85-7	beta-BHC	0.050	U	1	0.0049	0.050	ug/L	10/17/25	pd101725
319-86-8	delta-BHC	0.050	U	1	0.011	0.050	ug/L	10/17/25	pd101725
58-89-9	gamma-BHC (Lindane)	0.050	U	1	0.0037	0.050	ug/L	10/17/25	pd101725
76-44-8	Heptachlor	0.050	U	1	0.0027	0.050	ug/L	10/17/25	pd101725
309-00-2	Aldrin	0.050	U	1	0.0036	0.050	ug/L	10/17/25	pd101725
1024-57-3	Heptachlor epoxide	0.050	U	1	0.0096	0.050	ug/L	10/17/25	pd101725
959-98-8	Endosulfan I	0.050	U	1	0.0031	0.050	ug/L	10/17/25	pd101725
60-57-1	Dieldrin	0.050	U	1	0.0036	0.050	ug/L	10/17/25	pd101725
72-55-9	4,4-DDE	0.050	U	1	0.0037	0.050	ug/L	10/17/25	pd101725
72-20-8	Endrin	0.050	U	1	0.0032	0.050	ug/L	10/17/25	pd101725
33213-65-9	Endosulfan II	0.050	U	1	0.0079	0.050	ug/L	10/17/25	pd101725
72-54-8	4,4-DDD	0.050	U	1	0.0071	0.050	ug/L	10/17/25	pd101725
1031-07-8	Endosulfan Sulfate	0.050	U	1	0.0037	0.050	ug/L	10/17/25	pd101725
50-29-3	4,4-DDT	0.050	U	1	0.0035	0.050	ug/L	10/17/25	pd101725
72-43-5	Methoxychlor	0.050	U	1	0.011	0.050	ug/L	10/17/25	pd101725
53494-70-5	Endrin ketone	0.050	U	1	0.0093	0.050	ug/L	10/17/25	pd101725
7421-93-4	Endrin aldehyde	0.050	U	1	0.011	0.050	ug/L	10/17/25	pd101725
5103-71-9	alpha-Chlordane	0.050	U	1	0.0035	0.050	ug/L	10/17/25	pd101725
5103-74-2	gamma-Chlordane	0.050	U	1	0.0039	0.050	ug/L	10/17/25	pd101725
8001-35-2	Toxaphene	1.00	U	1	0.17	1.00	ug/L	10/17/25	pd101725
SURROGATES									
2051-24-3	Decachlorobiphenyl	21.3			57 - 171	106%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.0			61 - 148	100%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3363-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 10:53
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:03:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	3.545	2.875	60526543	594.5E6	19.436	19.964
28) SA Decachlor...	9.066	8.062	99514776	616.2E6	21.273	20.342

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 10:53
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:03:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 60526543
Conc: 19.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 594475365
Conc: 19.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 99514776
Conc: 21.27 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 616188932
Conc: 20.34 ng/ml

Report of Analysis

Client: AECOM	Date Collected: 10/17/25	
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received: 10/17/25	
Client Sample ID: PIBLK-PD090668.D	SDG No.: Q3363	
Lab Sample ID: I.BLK-PD090668.D	Matrix: WATER	
Analytical Method: 8081B	% Solid: 0	
Sample Wt/Vol: 1000 mL	Final Vol: 10000 uL	Test: Pesticide-TCL
Prep Method: 3510C	Prep Date:	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.050	U	1	0.0039	0.050	ug/L	10/17/25	pd101725
319-85-7	beta-BHC	0.050	U	1	0.0049	0.050	ug/L	10/17/25	pd101725
319-86-8	delta-BHC	0.050	U	1	0.011	0.050	ug/L	10/17/25	pd101725
58-89-9	gamma-BHC (Lindane)	0.050	U	1	0.0037	0.050	ug/L	10/17/25	pd101725
76-44-8	Heptachlor	0.050	U	1	0.0027	0.050	ug/L	10/17/25	pd101725
309-00-2	Aldrin	0.050	U	1	0.0036	0.050	ug/L	10/17/25	pd101725
1024-57-3	Heptachlor epoxide	0.050	U	1	0.0096	0.050	ug/L	10/17/25	pd101725
959-98-8	Endosulfan I	0.050	U	1	0.0031	0.050	ug/L	10/17/25	pd101725
60-57-1	Dieldrin	0.050	U	1	0.0036	0.050	ug/L	10/17/25	pd101725
72-55-9	4,4-DDE	0.050	U	1	0.0037	0.050	ug/L	10/17/25	pd101725
72-20-8	Endrin	0.050	U	1	0.0032	0.050	ug/L	10/17/25	pd101725
33213-65-9	Endosulfan II	0.050	U	1	0.0079	0.050	ug/L	10/17/25	pd101725
72-54-8	4,4-DDD	0.050	U	1	0.0071	0.050	ug/L	10/17/25	pd101725
1031-07-8	Endosulfan Sulfate	0.050	U	1	0.0037	0.050	ug/L	10/17/25	pd101725
50-29-3	4,4-DDT	0.050	U	1	0.0035	0.050	ug/L	10/17/25	pd101725
72-43-5	Methoxychlor	0.050	U	1	0.011	0.050	ug/L	10/17/25	pd101725
53494-70-5	Endrin ketone	0.050	U	1	0.0093	0.050	ug/L	10/17/25	pd101725
7421-93-4	Endrin aldehyde	0.050	U	1	0.011	0.050	ug/L	10/17/25	pd101725
5103-71-9	alpha-Chlordane	0.050	U	1	0.0035	0.050	ug/L	10/17/25	pd101725
5103-74-2	gamma-Chlordane	0.050	U	1	0.0039	0.050	ug/L	10/17/25	pd101725
8001-35-2	Toxaphene	1.00	U	1	0.17	1.00	ug/L	10/17/25	pd101725
SURROGATES									
2051-24-3	Decachlorobiphenyl	23.1			57 - 171	116%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.1			61 - 148	106%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3363-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090668.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 19:50
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:29:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.874	63598406	628.4E6	20.422	21.105
28) SA Decachlor...	9.065	8.061	106.8E6	700.9E6	22.833	23.139

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090668.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 19:50
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:29:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 63598406
Conc: 20.42 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: -0.001 min
Response: 628442425
Conc: 21.10 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 106809690
Conc: 22.83 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: -0.001 min
Response: 700921484
Conc: 23.14 ng/ml

Report of Analysis

Client: AECOM	Date Collected: 10/18/25	
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received: 10/18/25	
Client Sample ID: PIBLK-PD090687.D	SDG No.: Q3363	
Lab Sample ID: I.BLK-PD090687.D	Matrix: WATER	
Analytical Method: 8081B	% Solid: 0	
Sample Wt/Vol: 1000 mL	Final Vol: 10000 uL	Test: Pesticide-TCL
Prep Method: 3510C	Prep Date:	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.050	U	1	0.0039	0.050	ug/L	10/18/25	pd101725
319-85-7	beta-BHC	0.050	U	1	0.0049	0.050	ug/L	10/18/25	pd101725
319-86-8	delta-BHC	0.050	U	1	0.011	0.050	ug/L	10/18/25	pd101725
58-89-9	gamma-BHC (Lindane)	0.050	U	1	0.0037	0.050	ug/L	10/18/25	pd101725
76-44-8	Heptachlor	0.050	U	1	0.0027	0.050	ug/L	10/18/25	pd101725
309-00-2	Aldrin	0.050	U	1	0.0036	0.050	ug/L	10/18/25	pd101725
1024-57-3	Heptachlor epoxide	0.050	U	1	0.0096	0.050	ug/L	10/18/25	pd101725
959-98-8	Endosulfan I	0.050	U	1	0.0031	0.050	ug/L	10/18/25	pd101725
60-57-1	Dieldrin	0.050	U	1	0.0036	0.050	ug/L	10/18/25	pd101725
72-55-9	4,4-DDE	0.050	U	1	0.0037	0.050	ug/L	10/18/25	pd101725
72-20-8	Endrin	0.050	U	1	0.0032	0.050	ug/L	10/18/25	pd101725
33213-65-9	Endosulfan II	0.050	U	1	0.0079	0.050	ug/L	10/18/25	pd101725
72-54-8	4,4-DDD	0.050	U	1	0.0071	0.050	ug/L	10/18/25	pd101725
1031-07-8	Endosulfan Sulfate	0.050	U	1	0.0037	0.050	ug/L	10/18/25	pd101725
50-29-3	4,4-DDT	0.050	U	1	0.0035	0.050	ug/L	10/18/25	pd101725
72-43-5	Methoxychlor	0.050	U	1	0.011	0.050	ug/L	10/18/25	pd101725
53494-70-5	Endrin ketone	0.050	U	1	0.0093	0.050	ug/L	10/18/25	pd101725
7421-93-4	Endrin aldehyde	0.050	U	1	0.011	0.050	ug/L	10/18/25	pd101725
5103-71-9	alpha-Chlordane	0.050	U	1	0.0035	0.050	ug/L	10/18/25	pd101725
5103-74-2	gamma-Chlordane	0.050	U	1	0.0039	0.050	ug/L	10/18/25	pd101725
8001-35-2	Toxaphene	1.00	U	1	0.17	1.00	ug/L	10/18/25	pd101725
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.7			57 - 171	103%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.0			61 - 148	105%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3363-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090687.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2025 00:51
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:34:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	65293304	624.9E6	20.967	20.985
28)	SA Decachlor...	9.065	8.061	96652967	540.4E6	20.661	17.842

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090687.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2025 00:51
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:34:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 65293304
Conc: 20.97 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 624877076
Conc: 20.99 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.001 min
Response: 96652967
Conc: 20.66 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 540444429
Conc: 17.84 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM		Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206		Date Received:	
Client Sample ID:	PB170142BS		SDG No.:	Q3363
Lab Sample ID:	PB170142BS		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	100
Sample Wt/Vol:	30.03 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541B	Prep Date	10/17/25	Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	15.5		1	0.13	1.70	ug/kg	10/17/25 20:45	PB170142
319-85-7	beta-BHC	15.5		1	0.18	1.70	ug/kg	10/17/25 20:45	PB170142
319-86-8	delta-BHC	15.6		1	0.39	1.70	ug/kg	10/17/25 20:45	PB170142
58-89-9	gamma-BHC (Lindane)	15.6		1	0.14	1.70	ug/kg	10/17/25 20:45	PB170142
76-44-8	Heptachlor	15.7		1	0.12	1.70	ug/kg	10/17/25 20:45	PB170142
309-00-2	Aldrin	15.7		1	0.12	1.70	ug/kg	10/17/25 20:45	PB170142
1024-57-3	Heptachlor epoxide	15.6		1	0.19	1.70	ug/kg	10/17/25 20:45	PB170142
959-98-8	Endosulfan I	15.8		1	0.14	1.70	ug/kg	10/17/25 20:45	PB170142
60-57-1	Dieldrin	15.8		1	0.14	1.70	ug/kg	10/17/25 20:45	PB170142
72-55-9	4,4-DDE	15.7		1	0.14	1.70	ug/kg	10/17/25 20:45	PB170142
72-20-8	Endrin	16.0		1	0.14	1.70	ug/kg	10/17/25 20:45	PB170142
33213-65-9	Endosulfan II	15.8		1	0.29	1.70	ug/kg	10/17/25 20:45	PB170142
72-54-8	4,4-DDD	15.8		1	0.15	1.70	ug/kg	10/17/25 20:45	PB170142
1031-07-8	Endosulfan Sulfate	15.9		1	0.13	1.70	ug/kg	10/17/25 20:45	PB170142
50-29-3	4,4-DDT	16.2		1	0.14	1.70	ug/kg	10/17/25 20:45	PB170142
72-43-5	Methoxychlor	15.9		1	0.37	1.70	ug/kg	10/17/25 20:45	PB170142
53494-70-5	Endrin ketone	15.9		1	0.19	1.70	ug/kg	10/17/25 20:45	PB170142
7421-93-4	Endrin aldehyde	15.8		1	0.37	1.70	ug/kg	10/17/25 20:45	PB170142
5103-71-9	alpha-Chlordane	16.0		1	0.12	1.70	ug/kg	10/17/25 20:45	PB170142
5103-74-2	gamma-Chlordane	15.7		1	0.15	1.70	ug/kg	10/17/25 20:45	PB170142
8001-35-2	Toxaphene	33.0	U	1	5.40	33.0	ug/kg	10/17/25 20:45	PB170142
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.6			20 - 144	103%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.0			19 - 148	95%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090672.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 20:45
 Operator : AR\AJ
 Sample : PB170142BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB170142BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:30:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	56803309	566.9E6	18.240	19.038
28)	SA Decachlor...	9.066	8.062	93682539	623.5E6	20.026	20.583
Target Compounds							
2)	A alpha-BHC	3.994	3.386	332.9E6	2374.3E6	46.379	46.467
3)	MA gamma-BHC...	4.325	3.723	314.6E6	2213.0E6	46.270	46.838
4)	MA Heptachlor	4.923	4.074	247.3E6	2098.1E6	44.239	47.039
5)	MB Aldrin	5.264	4.360	306.6E6	2180.9E6	46.828	47.084
6)	B beta-BHC	4.512	4.019	117.6E6	925.0E6	44.908	46.649
7)	B delta-BHC	4.760	4.255	313.6E6	2236.0E6	46.380	46.860
8)	B Heptachlo...	5.684	4.864	270.9E6	1998.6E6	46.796	46.872
9)	A Endosulfan I	6.068	5.238	257.2E6	1867.8E6	46.458	47.417
10)	B gamma-Chl...	5.940	5.116	273.9E6	2122.5E6	46.176	47.221
11)	B alpha-Chl...	6.021	5.181	296.4E6	2036.6E6	47.959	47.213
12)	B 4,4'-DDE	6.190	5.365	248.2E6	2066.4E6	46.599	47.137
13)	MA Dieldrin	6.341	5.504	273.4E6	2098.4E6	46.882	47.491
14)	MA Endrin	6.568	5.780	233.2E6	1940.2E6	47.069	47.912
15)	B Endosulfa...	6.781	6.072	233.8E6	1782.1E6	46.765	47.452
16)	A 4,4'-DDD	6.700	5.921	190.3E6	1725.6E6	46.944	47.358
17)	MA 4,4'-DDT	7.016	6.174	190.1E6	1710.1E6	46.245	48.759
18)	B Endrin al...	6.910	6.250	169.8E6	1311.5E6	46.352	47.360
19)	B Endosulfa...	7.145	6.474	215.1E6	1713.7E6	46.382	47.808
20)	A Methoxychlor	7.488	6.745	95704276	834.3E6	45.025	47.872
21)	B Endrin ke...	7.625	6.983	225.8E6	1802.7E6	46.478	47.795
22)	Mirex	8.108	7.175	144.1E6	1194.8E6	45.023	46.509



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090672.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 20:45
Operator : AR\AJ
Sample : PB170142BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170142BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:30:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090672.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 56803309
Conc: 18.24 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170142BS

Time

Response_ Signal: PD090672.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 566898947
Conc: 19.04 ng/ml

Time

Response_ Signal: PD090672.D\ECD1A.ch

#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 332865321
Conc: 46.38 ng/ml

Time

Response_ Signal: PD090672.D\ECD2B.ch

#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2374282869
Conc: 46.47 ng/ml

Time

Response_

Signal: PD090672.D\ECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.325 min

Delta R.T.: 0.000 min

Response: 314583674


Conc: 46.27 ng/ml

Instrument :

ECD_D

ClientSampleId :

PB170142BS



Time

Response_

Signal: PD090672.D\ECD2B.ch


#3 gamma-BHC (Lindane)

R.T.: 3.723 min

Delta R.T.: 0.000 min

Response: 2213031313

Conc: 46.84 ng/ml



Time

Response_

Signal: PD090672.D\ECD1A.ch


#4 Heptachlor

R.T.: 4.923 min

Delta R.T.: 0.000 min

Response: 247275435

Conc: 44.24 ng/ml



Time

Response_

Signal: PD090672.D\ECD2B.ch


#4 Heptachlor

R.T.: 4.074 min


Delta R.T.: 0.000 min

Response: 2098084266

Conc: 47.04 ng/ml




Time



#5 Aldrin


R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 306597602
Conc: 46.83 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170142BS




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2180889952
Conc: 47.08 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 117642586
Conc: 44.91 ng/ml



#6 beta-BHC


R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 925036774
Conc: 46.65 ng/ml



#7 delta-BHC


R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 313608082
Conc: 46.38 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170142BS




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2235994398
Conc: 46.86 ng/ml



#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 270903307
Conc: 46.80 ng/ml




#8 Heptachlor epoxide

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 1998579813
Conc: 46.87 ng/ml

Response_

Signal: PD090672.D\ECD1A.ch

#9 Endosulfan I



R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 257220946
Conc: 46.46 ng/ml


Instrument :
ECD_D
ClientSampleId :
PB170142BS

Time

Response_

Signal: PD090672.D\ECD2B.ch

#9 Endosulfan I




R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1867785926
Conc: 47.42 ng/ml

Time

Response_

Signal: PD090672.D\ECD1A.ch

#10 gamma-Chlordane




R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 273877669
Conc: 46.18 ng/ml

Time

Response_

Signal: PD090672.D\ECD2B.ch

#10 gamma-Chlordane



R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 2122527117
Conc: 47.22 ng/ml

Time

Response_

Signal: PD090672.D\ECD1A.ch

#11 alpha-Chlordane

R.T.: 6.021 min

Delta R.T.: 0.000 min

Response: 296397895

Conc: 47.96 ng/ml

Instrument :

ECD_D

ClientSampleId :

PB170142BS

2e+07

1e+07

0

6.019

+

Time

5.80 5.85 5.90 5.95 6.00 6.05 6.10 6.15 6.20

Response_

Signal: PD090672.D\ECD2B.ch

#11 alpha-Chlordane

R.T.: 5.181 min

Delta R.T.: 0.000 min

Response: 2036649524

Conc: 47.21 ng/ml

2e+08

1.5e+08

1e+08

5e+07

0

5.180

+

Time

5.05 5.10 5.15 5.20 5.25 5.30

Response_

Signal: PD090672.D\ECD1A.ch

#12 4,4'-DDE

R.T.: 6.190 min

Delta R.T.: 0.000 min

Response: 248216951

Conc: 46.60 ng/ml

2e+07

1e+07

0

6.189

+

Time

6.00 6.10 6.20 6.30 6.40

Response_

Signal: PD090672.D\ECD2B.ch

#12 4,4'-DDE

R.T.: 5.365 min

Delta R.T.: 0.000 min

Response: 2066382846

Conc: 47.14 ng/ml

2e+08

1.5e+08

1e+08

5e+07

0

5.364

+


Time

5.20 5.30 5.40 5.50 5.60

Response_

Signal: PD090672.D\ECD1A.ch

#13 Dieldrin



R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 273385328
Conc: 46.88 ng/ml


Instrument :
ECD_D
ClientSampleId :
PB170142BS

Time

Response_

Signal: PD090672.D\ECD2B.ch

#13 Dieldrin




R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 2098360634
Conc: 47.49 ng/ml

Time

Response_

Signal: PD090672.D\ECD1A.ch

#14 Endrin




R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 233168211
Conc: 47.07 ng/ml

Time


Response_

Signal: PD090672.D\ECD2B.ch

#14 Endrin




R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 1940158737
Conc: 47.91 ng/ml



#15 Endosulfan II


R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 233822432
Conc: 46.76 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170142BS




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1782082435
Conc: 47.45 ng/ml




#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 190320354
Conc: 46.94 ng/ml



#16 4,4'-DDD


R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1725627117
Conc: 47.36 ng/ml



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 190089573
Conc: 46.25 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170142BS




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1710135391
Conc: 48.76 ng/ml



#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 169773008
Conc: 46.35 ng/ml



#18 Endrin aldehyde


R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1311540473
Conc: 47.36 ng/ml

Response_ Signal: PD090672.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 215130518
Conc: 46.38 ng/ml


Instrument :
ECD_D
ClientSampleId :
PB170142BS



Time Response_ Signal: PD090672.D\ECD2B.ch

#19 Endosulfan Sulfate


R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1713663483
Conc: 47.81 ng/ml



Time Response_ Signal: PD090672.D\ECD1A.ch

#20 Methoxychlor


R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 95704276
Conc: 45.02 ng/ml



Time Response_ Signal: PD090672.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 834298776
Conc: 47.87 ng/ml




Response_ Signal: PD090672.D\ECD1A.ch

#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 225827351
Conc: 46.48 ng/ml


Instrument :
ECD_D
ClientSampleId :
PB170142BS



Time Response_ Signal: PD090672.D\ECD2B.ch

#21 Endrin ketone


R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1802707194
Conc: 47.79 ng/ml



Time Response_ Signal: PD090672.D\ECD1A.ch

#22 Mirex


R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 144111926
Conc: 45.02 ng/ml




Time Response_ Signal: PD090672.D\ECD2B.ch

#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1194849561
Conc: 46.51 ng/ml






#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 93682539
Conc: 20.03 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170142BS



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 623493713
Conc: 20.58 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM		Date Collected:	10/15/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206		Date Received:	10/15/25
Client Sample ID:	PR132-SO3-010013-20251015MS		SDG No.:	Q3363
Lab Sample ID:	Q3363-04MS		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	51.5
Sample Wt/Vol:	30.06 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541B	Prep Date	10/17/25	Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	29.0		1	0.25	3.30	ug/kg	10/17/25 23:43	PB170142
319-85-7	beta-BHC	24.9		1	0.35	3.30	ug/kg	10/17/25 23:43	PB170142
319-86-8	delta-BHC	29.0		1	0.76	3.30	ug/kg	10/17/25 23:43	PB170142
58-89-9	gamma-BHC (Lindane)	28.5		1	0.27	3.30	ug/kg	10/17/25 23:43	PB170142
76-44-8	Heptachlor	27.7		1	0.23	3.30	ug/kg	10/17/25 23:43	PB170142
309-00-2	Aldrin	28.9		1	0.23	3.30	ug/kg	10/17/25 23:43	PB170142
1024-57-3	Heptachlor epoxide	28.0		1	0.37	3.30	ug/kg	10/17/25 23:43	PB170142
959-98-8	Endosulfan I	27.5		1	0.27	3.30	ug/kg	10/17/25 23:43	PB170142
60-57-1	Dieldrin	27.1		1	0.27	3.30	ug/kg	10/17/25 23:43	PB170142
72-55-9	4,4-DDE	25.3		1	0.27	3.30	ug/kg	10/17/25 23:43	PB170142
72-20-8	Endrin	27.8		1	0.27	3.30	ug/kg	10/17/25 23:43	PB170142
33213-65-9	Endosulfan II	26.8		1	0.56	3.30	ug/kg	10/17/25 23:43	PB170142
72-54-8	4,4-DDD	27.0		1	0.29	3.30	ug/kg	10/17/25 23:43	PB170142
1031-07-8	Endosulfan Sulfate	26.3		1	0.25	3.30	ug/kg	10/17/25 23:43	PB170142
50-29-3	4,4-DDT	27.1		1	0.27	3.30	ug/kg	10/17/25 23:43	PB170142
72-43-5	Methoxychlor	26.9		1	0.72	3.30	ug/kg	10/17/25 23:43	PB170142
53494-70-5	Endrin ketone	26.8		1	0.37	3.30	ug/kg	10/17/25 23:43	PB170142
7421-93-4	Endrin aldehyde	27.0		1	0.72	3.30	ug/kg	10/17/25 23:43	PB170142
5103-71-9	alpha-Chlordane	26.6		1	0.23	3.30	ug/kg	10/17/25 23:43	PB170142
5103-74-2	gamma-Chlordane	27.1		1	0.29	3.30	ug/kg	10/17/25 23:43	PB170142
8001-35-2	Toxaphene	63.9	U	1	10.5	63.9	ug/kg	10/17/25 23:43	PB170142
SURROGATES									
2051-24-3	Decachlorobiphenyl	12.7			20 - 144	63%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	13.7			19 - 148	69%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3363-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090682.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 23:43
 Operator : AR\AJ
 Sample : Q3363-04MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PR132-SO3-010013-20251015MS

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 10/21/2025
 Supervised By : mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:33:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.543	2.875	40355759	408.0E6	12.959m	13.701
28) SA Decachlor...	9.066	8.062	59203025	292.6E6	12.656	9.659
Target Compounds						
2) A alpha-BHC	3.994	3.386	321.9E6	2035.2E6	44.845	39.830
3) MA gamma-BHC...	4.325	3.722	300.4E6	1861.6E6	44.187	39.400
4) MA Heptachlor	4.923	4.074	239.8E6	1813.8E6	42.901	40.666
5) MB Aldrin	5.264	4.360	293.1E6	1810.2E6	44.773	39.081
6) B beta-BHC	4.512	4.017	101.0E6	759.6E6	38.573	38.306m
7) B delta-BHC	4.760	4.255	304.1E6	1891.2E6	44.972	39.634
8) B Heptachlo...	5.684	4.863	250.6E6	1701.5E6	43.287	39.904m
9) A Endosulfan I	6.068	5.239	235.9E6	1364.6E6	42.599	34.643
10) B gamma-Chl...	5.940	5.116	248.9E6	1809.3E6	41.963	40.252
11) B alpha-Chl...	6.020	5.181	254.6E6	1685.2E6	41.196	39.065
12) B 4,4'-DDE	6.190	5.366	208.8E6	1569.9E6	39.199	35.810
13) MA Dieldrin	6.341	5.503	245.1E6	1732.5E6	42.023	39.210
14) MA Endrin	6.568	5.780	213.5E6	1537.2E6	43.096	37.961
15) B Endosulfa...	6.781	6.070	207.1E6	1400.7E6	41.419	37.297m
16) A 4,4'-DDD	6.700	5.921	169.4E6	1245.2E6	41.789	34.174
17) MA 4,4'-DDT	7.015	6.174	172.3E6	1297.4E6	41.912	36.991
18) B Endrin al...	6.910	6.250	153.2E6	1009.9E6	41.827	36.468
19) B Endosulfa...	7.144	6.474	188.7E6	1296.6E6	40.692	36.174
20) A Methoxychlor	7.488	6.745	88514245	637.2E6	41.642	36.560
21) B Endrin ke...	7.625	6.983	201.9E6	1267.6E6	41.558	33.608
22) Mirex	8.106	7.175	126.0E6	842.3E6	39.380m	32.787

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090682.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 23:43
Operator : AR\AJ
Sample : Q3363-04MS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PR132-SO3-010013-20251015MS

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 10/21/2025

Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:33:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: -0.002 min
Response: 40355759
Conc: 12.96 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 407985494
Conc: 13.70 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 321856084
Conc: 44.84 ng/ml



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2035157045
Conc: 39.83 ng/ml



#3 gamma-BHC (Lindane)

R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 300418994
Conc: 44.19 ng/ml

Instrument :

ECD_D

ClientSampleId :


PR132-SO3-010013-20251015MS

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 10/21/2025

Supervised By :mohammad ahmed 10/24/2025




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 1861605360
Conc: 39.40 ng/ml



#4 Heptachlor


R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 239796224
Conc: 42.90 ng/ml



#4 Heptachlor

R.T.: 4.074 min
Delta R.T.: 0.000 min
Response: 1813827920
Conc: 40.67 ng/ml

Response_ Signal: PD090682.D\ECD1A.ch



#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 293138835
Conc: 44.77 ng/ml

Instrument :

ECD_D

ClientSampleId :

PR132-SO3-010013-20251015MS


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/21/2025

Supervised By :mohammad ahmed 10/24/2025


Response_ Signal: PD090682.D\ECD2B.ch



#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 1810176572
Conc: 39.08 ng/ml


Response_ Signal: PD090682.D\ECD1A.ch



#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 101047794
Conc: 38.57 ng/ml

Response_ Signal: PD090682.D\ECD2B.ch



#6 beta-BHC

R.T.: 4.017 min
Delta R.T.: -0.001 min
Response: 759594568
Conc: 38.31 ng/ml m

Response_ Signal: PD090682.D\ECD1A.ch

#7 delta-BHC

R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 304088622
Conc: 44.97 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Time Response_ Signal: PD090682.D\ECD2B.ch

#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 1891196954
Conc: 39.63 ng/ml

Time Response_ Signal: PD090682.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 250590453
Conc: 43.29 ng/ml

Time Response_ Signal: PD090682.D\ECD2B.ch

#8 Heptachlor epoxide

R.T.: 4.863 min
Delta R.T.: -0.001 min
Response: 1701482894
Conc: 39.90 ng/ml m

Time Response_ Signal: PD090682.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 250590453
Conc: 43.29 ng/ml

Time Response_ Signal: PD090682.D\ECD2B.ch

#8 Heptachlor epoxide

R.T.: 4.863 min
Delta R.T.: -0.001 min
Response: 1701482894
Conc: 39.90 ng/ml m

Time Response_ Signal: PD090682.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 250590453
Conc: 43.29 ng/ml

Time Response_ Signal: PD090682.D\ECD2B.ch

#8 Heptachlor epoxide

R.T.: 4.863 min
Delta R.T.: -0.001 min
Response: 1701482894
Conc: 39.90 ng/ml m

Time Response_ Signal: PD090682.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 250590453
Conc: 43.29 ng/ml

Time Response_ Signal: PD090682.D\ECD2B.ch

#8 Heptachlor epoxide

R.T.: 4.863 min
Delta R.T.: -0.001 min
Response: 1701482894
Conc: 39.90 ng/ml m

Time Response_ Signal: PD090682.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 250590453
Conc: 43.29 ng/ml

Time Response_ Signal: PD090682.D\ECD2B.ch


#8 Heptachlor epoxide

R.T.: 4.863 min
Delta R.T.: -0.001 min
Response: 1701482894
Conc: 39.90 ng/ml m

Time Response_ Signal: PD090682.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 250590453
Conc: 43.29 ng/ml




#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 235855871
Conc: 42.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#9 Endosulfan I

R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 1364583597
Conc: 34.64 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 248890004
Conc: 41.96 ng/ml



#10 gamma-Chlordane

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 1809288107
Conc: 40.25 ng/ml




#11 alpha-Chlordane

R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 254599978
Conc: 41.20 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 1685166630
Conc: 39.06 ng/ml



#12 4,4'-DDE


R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 208797473
Conc: 39.20 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 1569867061
Conc: 35.81 ng/ml

Response_ Signal: PD090682.D\ECD1A.ch



#13 Dieldrin


R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 245052298
Conc: 42.02 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025


Time Response_ Signal: PD090682.D\ECD2B.ch



#13 Dieldrin

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 1732466558
Conc: 39.21 ng/ml


Time Response_ Signal: PD090682.D\ECD1A.ch



#14 Endrin


R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 213484245
Conc: 43.10 ng/ml

Time Response_ Signal: PD090682.D\ECD2B.ch



#14 Endrin

R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 1537198731
Conc: 37.96 ng/ml




#15 Endosulfan II

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 207092477
Conc: 41.42 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#15 Endosulfan II

R.T.: 6.070 min
Delta R.T.: -0.002 min
Response: 1400714599
Conc: 37.30 ng/ml m



#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 169420152
Conc: 41.79 ng/ml



#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1245223566
Conc: 34.17 ng/ml

Response_ Signal: PD090682.D\ECD1A.ch


#17 4,4'-DDT

R.T.: 7.015 min
Delta R.T.: 0.000 min
Response: 172277888
Conc: 41.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MS

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



Response_ Signal: PD090682.D\ECD2B.ch

#17 4,4'-DDT


R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1297377661
Conc: 36.99 ng/ml



Response_ Signal: PD090682.D\ECD1A.ch

#18 Endrin aldehyde


R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 153199665
Conc: 41.83 ng/ml




Response_ Signal: PD090682.D\ECD2B.ch

#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1009891427
Conc: 36.47 ng/ml






#19 Endosulfan Sulfate

R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 188739099
Conc: 40.69 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1296640213
Conc: 36.17 ng/ml



#20 Methoxychlor


R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 88514245
Conc: 41.64 ng/ml



#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 637167640
Conc: 36.56 ng/ml

Response_ Signal: PD090682.D\ECD1A.ch



#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 201921711
Conc: 41.56 ng/ml

Instrument :

ECD_D


ClientSampleId :

PR132-SO3-010013-20251015MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025


Time Response_ Signal: PD090682.D\ECD2B.ch



#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1267609134
Conc: 33.61 ng/ml


Time Response_ Signal: PD090682.D\ECD1A.ch



#22 Mirex


R.T.: 8.106 min
Delta R.T.: -0.002 min
Response: 126048619
Conc: 39.38 ng/ml m

Time Response_ Signal: PD090682.D\ECD2B.ch



#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 842317051
Conc: 32.79 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 59203025
Conc: 12.66 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 292594483
Conc: 9.66 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM		Date Collected:	10/15/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206		Date Received:	10/15/25
Client Sample ID:	PR132-SO3-010013-20251015MSD		SDG No.:	Q3363
Lab Sample ID:	Q3363-04MSD		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	51.5
Sample Wt/Vol:	30.08 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541B	Prep Date	10/17/25	Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	30.2		1	0.25	3.30	ug/kg	10/17/25 23:56	PB170142
319-85-7	beta-BHC	25.8		1	0.35	3.30	ug/kg	10/17/25 23:56	PB170142
319-86-8	delta-BHC	30.0		1	0.76	3.30	ug/kg	10/17/25 23:56	PB170142
58-89-9	gamma-BHC (Lindane)	29.6		1	0.27	3.30	ug/kg	10/17/25 23:56	PB170142
76-44-8	Heptachlor	28.7		1	0.23	3.30	ug/kg	10/17/25 23:56	PB170142
309-00-2	Aldrin	29.8		1	0.23	3.30	ug/kg	10/17/25 23:56	PB170142
1024-57-3	Heptachlor epoxide	29.0		1	0.37	3.30	ug/kg	10/17/25 23:56	PB170142
959-98-8	Endosulfan I	28.8		1	0.27	3.30	ug/kg	10/17/25 23:56	PB170142
60-57-1	Dieldrin	28.3		1	0.27	3.30	ug/kg	10/17/25 23:56	PB170142
72-55-9	4,4-DDE	26.6		1	0.27	3.30	ug/kg	10/17/25 23:56	PB170142
72-20-8	Endrin	29.1		1	0.27	3.30	ug/kg	10/17/25 23:56	PB170142
33213-65-9	Endosulfan II	28.0		1	0.56	3.30	ug/kg	10/17/25 23:56	PB170142
72-54-8	4,4-DDD	28.4		1	0.29	3.30	ug/kg	10/17/25 23:56	PB170142
1031-07-8	Endosulfan Sulfate	27.5		1	0.25	3.30	ug/kg	10/17/25 23:56	PB170142
50-29-3	4,4-DDT	28.2		1	0.27	3.30	ug/kg	10/17/25 23:56	PB170142
72-43-5	Methoxychlor	27.7		1	0.72	3.30	ug/kg	10/17/25 23:56	PB170142
53494-70-5	Endrin ketone	28.4		1	0.37	3.30	ug/kg	10/17/25 23:56	PB170142
7421-93-4	Endrin aldehyde	28.2		1	0.72	3.30	ug/kg	10/17/25 23:56	PB170142
5103-71-9	alpha-Chlordane	27.6		1	0.23	3.30	ug/kg	10/17/25 23:56	PB170142
5103-74-2	gamma-Chlordane	28.3		1	0.29	3.30	ug/kg	10/17/25 23:56	PB170142
8001-35-2	Toxaphene	63.9	U	1	10.5	63.9	ug/kg	10/17/25 23:56	PB170142
SURROGATES									
2051-24-3	Decachlorobiphenyl	13.2			20 - 144	66%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	13.8			19 - 148	69%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3363-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090683.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 23:56
 Operator : AR\AJ
 Sample : Q3363-04MSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PR132-SO3-010013-20251015MSD

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 10/21/2025
 Supervised By : mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:33:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.875	42156550	412.0E6	13.537m	13.837
28)	SA Decachlor...	9.066	8.062	61743198	295.4E6	13.199	9.752 #
Target Compounds							
2)	A alpha-BHC	3.994	3.386	335.6E6	2058.1E6	46.753	40.280
3)	MA gamma-BHC...	4.324	3.723	311.6E6	1880.3E6	45.824	39.796
4)	MA Heptachlor	4.923	4.075	248.8E6	1843.7E6	44.514	41.336
5)	MB Aldrin	5.264	4.360	302.3E6	1844.3E6	46.170	39.818
6)	B beta-BHC	4.512	4.017	104.9E6	769.2E6	40.040	38.789m
7)	B delta-BHC	4.760	4.255	314.6E6	1910.5E6	46.519	40.038
8)	B Heptachlo...	5.684	4.862	260.3E6	1693.4E6	44.965	39.715m
9)	A Endosulfan I	6.068	5.239	246.6E6	1390.4E6	44.541	35.298
10)	B gamma-Chl...	5.940	5.116	260.4E6	1841.2E6	43.898	40.962
11)	B alpha-Chl...	6.021	5.181	264.4E6	1706.9E6	42.783	39.569
12)	B 4,4'-DDE	6.190	5.366	219.8E6	1598.4E6	41.264	36.461
13)	MA Dieldrin	6.341	5.503	256.1E6	1773.8E6	43.915	40.146
14)	MA Endrin	6.568	5.780	223.3E6	1564.9E6	45.070	38.645
15)	B Endosulfa...	6.781	6.071	217.0E6	1418.4E6	43.394	37.768m
16)	A 4,4'-DDD	6.700	5.921	178.3E6	1264.3E6	43.978	34.698
17)	MA 4,4'-DDT	7.016	6.174	179.9E6	1321.9E6	43.757	37.689
18)	B Endrin al...	6.910	6.250	159.7E6	1030.0E6	43.608	37.194
19)	B Endosulfa...	7.145	6.474	197.4E6	1307.8E6	42.549	36.484
20)	A Methoxychlor	7.488	6.745	91330010	654.2E6	42.967	37.536
21)	B Endrin ke...	7.625	6.982	213.9E6	1270.2E6	44.019	33.676
22)	Mirex	8.106	7.175	130.4E6	853.2E6	40.740m	33.211

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090683.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 23:56
Operator : AR\AJ
Sample : Q3363-04MSD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PR132-SO3-010013-20251015MSD

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 10/21/2025

Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:33:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: -0.002 min
Response: 42156550
Conc: 13.54 ng/ml

Instrument :

ECD_D

ClientSampleId :


PR132-SO3-010013-20251015MSD

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 10/21/2025

Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 412014015
Conc: 13.84 ng/ml



#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 335554713
Conc: 46.75 ng/ml



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2058147056
Conc: 40.28 ng/ml

Response_ Signal: PD090683.D\ECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 311550671
Conc: 45.82 ng/ml

Instrument :

ECD_D

ClientSampleId :

PR132-SO3-010013-20251015MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Time

Response_ Signal: PD090683.D\ECD2B.ch

#3 gamma-BHC (Lindane)

R.T.: 3.723 min
Delta R.T.: 0.000 min
Response: 1880311983
Conc: 39.80 ng/ml

Time

Response_ Signal: PD090683.D\ECD1A.ch

#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 248812749
Conc: 44.51 ng/ml


Time

Response_ Signal: PD090683.D\ECD2B.ch

#4 Heptachlor

R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 1843699703
Conc: 41.34 ng/ml

Time



#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 302286392
Conc: 46.17 ng/ml

Instrument :

ECD_D


ClientSampleId :

PR132-SO3-010013-20251015MSD

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 1844333628
Conc: 39.82 ng/ml



#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 104889029
Conc: 40.04 ng/ml



#6 beta-BHC

R.T.: 4.017 min
Delta R.T.: -0.001 min
Response: 769176059
Conc: 38.79 ng/ml m

Response_ Signal: PD090683.D\ECD1A.ch

#7 delta-BHC

R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 314551580
Conc: 46.52 ng/ml

Instrument :

ECD_D

ClientSampleId :

PR132-SO3-010013-20251015MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Time

Response_ Signal: PD090683.D\ECD2B.ch

#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 1910501544
Conc: 40.04 ng/ml

Time

Response_ Signal: PD090683.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 260306809
Conc: 44.97 ng/ml

Time

Response_ Signal: PD090683.D\ECD2B.ch

#8 Heptachlor epoxide

R.T.: 4.862 min
Delta R.T.: -0.002 min
Response: 1693402159
Conc: 39.71 ng/ml m

Time

Response_ Signal: PD090683.D\ECD1A.ch

#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 246606994
Conc: 44.54 ng/ml

Instrument :

ECD_D



ClientSampleId :

PR132-SO3-010013-20251015MSD

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#9 Endosulfan I

R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 1390399325
Conc: 35.30 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 260364990
Conc: 43.90 ng/ml



#10 gamma-Chlordane

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 1841198200
Conc: 40.96 ng/ml



#11 alpha-Chlordane

R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 264407286
Conc: 42.78 ng/ml

Instrument :

ECD_D


ClientSampleId :

PR132-SO3-010013-20251015MSD

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 1706927189
Conc: 39.57 ng/ml



#12 4,4'-DDE

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 219799396
Conc: 41.26 ng/ml




#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 1598387778
Conc: 36.46 ng/ml

Response_ Signal: PD090683.D\ECD1A.ch

#13 Dieldrin



R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 256083353
Conc: 43.91 ng/ml


Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Response_ Signal: PD090683.D\ECD2B.ch


#13 Dieldrin



R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 1773794283
Conc: 40.15 ng/ml

Response_ Signal: PD090683.D\ECD1A.ch


#14 Endrin




R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 223261788
Conc: 45.07 ng/ml

Response_ Signal: PD090683.D\ECD2B.ch

#14 Endrin



R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 1564894393
Conc: 38.65 ng/ml



#15 Endosulfan II

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 216969415
Conc: 43.39 ng/ml

Instrument :

ECD_D


ClientSampleId :

PR132-SO3-010013-20251015MSD

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#15 Endosulfan II

R.T.: 6.071 min
Delta R.T.: -0.002 min
Response: 1418381315
Conc: 37.77 ng/ml m



#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 178294481
Conc: 43.98 ng/ml




#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1264325312
Conc: 34.70 ng/ml

Response_ Signal: PD090683.D\ECD1A.ch

#17 4,4'-DDT



R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 179862617
Conc: 43.76 ng/ml

Instrument :

ECD_D

ClientSampleId :

PR132-SO3-010013-20251015MSD

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 10/21/2025

Supervised By :mohammad ahmed 10/24/2025

Time Response_ Signal: PD090683.D\ECD2B.ch


#17 4,4'-DDT



R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1321886221
Conc: 37.69 ng/ml

Time Response_ Signal: PD090683.D\ECD1A.ch


#18 Endrin aldehyde



R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 159721564
Conc: 43.61 ng/ml


Time Response_ Signal: PD090683.D\ECD2B.ch

#18 Endrin aldehyde



R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1030004149
Conc: 37.19 ng/ml

Response_ Signal: PD090683.D\ECD1A.ch



#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 197351892
Conc: 42.55 ng/ml

Instrument :

ECD_D

ClientSampleId :

PR132-SO3-010013-20251015MSD


Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/21/2025

Supervised By :mohammad ahmed 10/24/2025


Time Response_ Signal: PD090683.D\ECD2B.ch



#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1307772787
Conc: 36.48 ng/ml


Time Response_ Signal: PD090683.D\ECD1A.ch



#20 Methoxychlor

R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 91330010
Conc: 42.97 ng/ml

Time Response_ Signal: PD090683.D\ECD2B.ch



#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 654178585
Conc: 37.54 ng/ml

Response_ Signal: PD090683.D\ECD1A.ch

#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 213879941
Conc: 44.02 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025

Time

Response_ Signal: PD090683.D\ECD2B.ch

#21 Endrin ketone

R.T.: 6.982 min
Delta R.T.: 0.000 min
Response: 1270168384
Conc: 33.68 ng/ml

Time

Response_ Signal: PD090683.D\ECD1A.ch

#22 Mirex

R.T.: 8.106 min
Delta R.T.: 0.000 min
Response: 130402880
Conc: 40.74 ng/ml m

Time

Response_ Signal: PD090683.D\ECD2B.ch

#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 853222972
Conc: 33.21 ng/ml

Time

Response_ Signal: PD090683.D\ECD1A.ch

#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 853222972
Conc: 33.21 ng/ml


Time

Response_ Signal: PD090683.D\ECD2B.ch

#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 853222972
Conc: 33.21 ng/ml

Time




#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 61743198
Conc: 13.20 ng/ml

Instrument :
ECD_D
ClientSampleId :
PR132-SO3-010013-20251015MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 295409697
Conc: 9.75 ng/ml

Manual Integration Report

Sequence:	pd101725	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD090648.D	Endrin aldehyde	Abdul	10/19/2025 5:20:46 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090648.D	Endrin ketone #2	Abdul	10/19/2025 5:20:46 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
RESCHK	PD090649.D	Tetrachloro-m-xylene	Abdul	10/19/2025 5:20:37 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PSTDICC005	PD090654.D	alpha-Chlordane	Abdul	10/19/2025 5:19:40 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PSTDICC005	PD090654.D	Endrin ketone	Abdul	10/19/2025 5:19:40 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PSTDICC005	PD090654.D	Tetrachloro-m-xylene	Abdul	10/19/2025 5:19:40 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PTOXICV500	PD090667.D	Toxaphene-5 #2	Abdul	10/21/2025 8:49:16 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090669.D	4,4"-DDD	Abdul	10/21/2025 8:49:12 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090669.D	4,4"-DDD #2	Abdul	10/21/2025 8:49:12 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090669.D	Endrin aldehyde	Abdul	10/21/2025 8:49:12 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-01	PD090678.D	Tetrachloro-m-xylene	Abdul	10/21/2025 8:48:40 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-02	PD090679.D	4,4"-DDD	Abdul	10/21/2025 8:48:37 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-02	PD090679.D	alpha-Chlordane #2	Abdul	10/21/2025 8:48:37 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software

Manual Integration Report

Sequence:	pd101725	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3363-02	PD090679.D	Endrin	Abdul	10/21/2025 8:48:37 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-02	PD090679.D	Endrin #2	Abdul	10/21/2025 8:48:37 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-02	PD090679.D	Tetrachloro-m-xylene #2	Abdul	10/21/2025 8:48:37 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-03	PD090680.D	4,4"-DDD	Abdul	10/21/2025 8:48:34 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-03	PD090680.D	4,4"-DDD #2	Abdul	10/21/2025 8:48:34 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-03	PD090680.D	alpha-Chlordane	Abdul	10/21/2025 8:48:34 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-03	PD090680.D	alpha-Chlordane #2	Abdul	10/21/2025 8:48:34 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-03	PD090680.D	Dieldrin	Abdul	10/21/2025 8:48:34 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-03	PD090680.D	Endrin	Abdul	10/21/2025 8:48:34 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-03	PD090680.D	Endrin #2	Abdul	10/21/2025 8:48:34 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-03	PD090680.D	gamma-Chlordane	Abdul	10/21/2025 8:48:34 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-03	PD090680.D	Tetrachloro-m-xylene	Abdul	10/21/2025 8:48:34 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04	PD090681.D	Tetrachloro-m-xylene	Abdul	10/21/2025 8:48:31 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software

Manual Integration Report

Sequence:	pd101725	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3363-04MS	PD090682.D	beta-BHC #2	Abdul	10/21/2025 8:48:28 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04MS	PD090682.D	Endosulfan II #2	Abdul	10/21/2025 8:48:28 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04MS	PD090682.D	Heptachlor epoxide #2	Abdul	10/21/2025 8:48:28 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04MS	PD090682.D	Mirex	Abdul	10/21/2025 8:48:28 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04MS	PD090682.D	Tetrachloro-m-xylene	Abdul	10/21/2025 8:48:28 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04MSD	PD090683.D	beta-BHC #2	Abdul	10/21/2025 8:48:24 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04MSD	PD090683.D	Endosulfan II #2	Abdul	10/21/2025 8:48:24 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04MSD	PD090683.D	Heptachlor epoxide #2	Abdul	10/21/2025 8:48:24 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04MSD	PD090683.D	Mirex	Abdul	10/21/2025 8:48:24 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-04MSD	PD090683.D	Tetrachloro-m-xylene	Abdul	10/21/2025 8:48:24 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
Q3363-05	PD090684.D	Tetrachloro-m-xylene	Abdul	10/21/2025 8:48:21 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090646.D	17 Oct 2025 10:40	AR\AJ	Ok
2	I.BLK	PD090647.D	17 Oct 2025 10:53	AR\AJ	Ok
3	PEM	PD090648.D	17 Oct 2025 11:07	AR\AJ	Ok,M
4	RESCHK	PD090649.D	17 Oct 2025 11:20	AR\AJ	Ok,M
5	PSTDICC100	PD090650.D	17 Oct 2025 11:34	AR\AJ	Ok
6	PSTDICC075	PD090651.D	17 Oct 2025 11:48	AR\AJ	Ok
7	PSTDICC050	PD090652.D	17 Oct 2025 12:01	AR\AJ	Ok
8	PSTDICC025	PD090653.D	17 Oct 2025 12:15	AR\AJ	Ok
9	PSTDICC005	PD090654.D	17 Oct 2025 12:28	AR\AJ	Ok,M
10	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42	AR\AJ	Ok
11	PCHLORICC750	PD090656.D	17 Oct 2025 12:56	AR\AJ	Ok
12	PCHLORICC500	PD090657.D	17 Oct 2025 13:09	AR\AJ	Ok
13	PCHLORICC250	PD090658.D	17 Oct 2025 13:23	AR\AJ	Ok
14	PCHLORICC050	PD090659.D	17 Oct 2025 13:36	AR\AJ	Ok,M
15	PTOXICC1000	PD090660.D	17 Oct 2025 13:50	AR\AJ	Ok
16	PTOXICC750	PD090661.D	17 Oct 2025 14:03	AR\AJ	Ok
17	PTOXICC500	PD090662.D	17 Oct 2025 14:17	AR\AJ	Ok
18	PTOXICC250	PD090663.D	17 Oct 2025 14:30	AR\AJ	Ok
19	PTOXICC100	PD090664.D	17 Oct 2025 16:18	AR\AJ	Ok
20	PSTDICV050	PD090665.D	17 Oct 2025 16:46	AR\AJ	Ok
21	PCHLORICV500	PD090666.D	17 Oct 2025 17:50	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PD090667.D	17 Oct 2025 18:42	AR\AJ	Ok,M
23	I.BLK	PD090668.D	17 Oct 2025 19:50	AR\AJ	Ok
24	PEM	PD090669.D	17 Oct 2025 20:04	AR\AJ	Ok,M
25	PSTDCCC050	PD090670.D	17 Oct 2025 20:18	AR\AJ	Ok
26	PB170142BL	PD090671.D	17 Oct 2025 20:31	AR\AJ	Ok
27	PB170142BS	PD090672.D	17 Oct 2025 20:45	AR\AJ	Ok
28	Q3366-01	PD090673.D	17 Oct 2025 20:59	AR\AJ	Ok,M
29	Q3366-03	PD090674.D	17 Oct 2025 21:12	AR\AJ	Ok,M
30	Q3367-01	PD090675.D	17 Oct 2025 21:26	AR\AJ	Ok,M
31	Q3369-01	PD090676.D	17 Oct 2025 21:40	AR\AJ	Ok,M
32	Q3369-05	PD090677.D	17 Oct 2025 21:53	AR\AJ	Ok,M
33	Q3363-01	PD090678.D	17 Oct 2025 22:48	AR\AJ	Ok,M
34	Q3363-02	PD090679.D	17 Oct 2025 23:02	AR\AJ	Ok,M
35	Q3363-03	PD090680.D	17 Oct 2025 23:15	AR\AJ	Ok,M
36	Q3363-04	PD090681.D	17 Oct 2025 23:29	AR\AJ	Ok,M
37	Q3363-04MS	PD090682.D	17 Oct 2025 23:43	AR\AJ	Ok,M
38	Q3363-04MSD	PD090683.D	17 Oct 2025 23:56	AR\AJ	Ok,M
39	Q3363-05	PD090684.D	18 Oct 2025 00:10	AR\AJ	Ok,M
40	Q3343-04MS	PD090685.D	18 Oct 2025 00:24	AR\AJ	Ok,M
41	Q3343-04MSD	PD090686.D	18 Oct 2025 00:37	AR\AJ	Ok
42	I.BLK	PD090687.D	18 Oct 2025 00:51	AR\AJ	Ok
43	PSTDCCC050	PD090688.D	18 Oct 2025 02:13	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090646.D	17 Oct 2025 10:40		AR\AJ	Ok
2	I.BLK	I.BLK	PD090647.D	17 Oct 2025 10:53		AR\AJ	Ok
3	PEM	PEM	PD090648.D	17 Oct 2025 11:07		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD090649.D	17 Oct 2025 11:20		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PD090650.D	17 Oct 2025 11:34		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD090651.D	17 Oct 2025 11:48		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD090652.D	17 Oct 2025 12:01		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD090653.D	17 Oct 2025 12:15		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD090654.D	17 Oct 2025 12:28		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD090656.D	17 Oct 2025 12:56		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD090657.D	17 Oct 2025 13:09		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD090658.D	17 Oct 2025 13:23		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD090659.D	17 Oct 2025 13:36		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PD090660.D	17 Oct 2025 13:50		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD090661.D	17 Oct 2025 14:03		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD090662.D	17 Oct 2025 14:17		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD090663.D	17 Oct 2025 14:30		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD090664.D	17 Oct 2025 16:18		AR\AJ	Ok
20	PSTDICV050	ICVPD101725	PD090665.D	17 Oct 2025 16:46		AR\AJ	Ok
21	PCHLORICV500	ICVPD101725CHLOR	PD090666.D	17 Oct 2025 17:50		AR\AJ	Ok
22	PTOXICV500	ICVPD101725TOX	PD090667.D	17 Oct 2025 18:42		AR\AJ	Ok,M
23	I.BLK	I.BLK	PD090668.D	17 Oct 2025 19:50		AR\AJ	Ok
24	PEM	PEM	PD090669.D	17 Oct 2025 20:04		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD090670.D	17 Oct 2025 20:18		AR\AJ	Ok
26	PB170142BL	PB170142BL	PD090671.D	17 Oct 2025 20:31		AR\AJ	Ok
27	PB170142BS	PB170142BS	PD090672.D	17 Oct 2025 20:45		AR\AJ	Ok
28	Q3366-01	TR-06-101625	PD090673.D	17 Oct 2025 20:59		AR\AJ	Ok,M
29	Q3366-03	TR-1-101625	PD090674.D	17 Oct 2025 21:12		AR\AJ	Ok,M
30	Q3367-01	BU-03-101625	PD090675.D	17 Oct 2025 21:26		AR\AJ	Ok,M
31	Q3369-01	MH-22	PD090676.D	17 Oct 2025 21:40		AR\AJ	Ok,M
32	Q3369-05	MH-21	PD090677.D	17 Oct 2025 21:53		AR\AJ	Ok,M
33	Q3363-01	PR132-SO1-085011-20	PD090678.D	17 Oct 2025 22:48		AR\AJ	Ok,M
34	Q3363-02	PR132-SO1-039085-20	PD090679.D	17 Oct 2025 23:02		AR\AJ	Ok,M
35	Q3363-03	PR132-SO3-028010-20	PD090680.D	17 Oct 2025 23:15		AR\AJ	Ok,M
36	Q3363-04	PR132-SO3-010013-20	PD090681.D	17 Oct 2025 23:29		AR\AJ	Ok,M
37	Q3363-04MS	PR132-SO3-010013-20	PD090682.D	17 Oct 2025 23:43		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q3363-04MSD	PR132-SO3-010013-20	PD090683.D	17 Oct 2025 23:56		AR\AJ	Ok,M
39	Q3363-05	COMP01-20251015	PD090684.D	18 Oct 2025 00:10		AR\AJ	Ok,M
40	Q3343-04MS	SOIL-ROLLOFF-WC-1	PD090685.D	18 Oct 2025 00:24		AR\AJ	Ok,M
41	Q3343-04MSD	SOIL-ROLLOFF-WC-1	PD090686.D	18 Oct 2025 00:37		AR\AJ	Ok
42	I.BLK	I.BLK	PD090687.D	18 Oct 2025 00:51		AR\AJ	Ok
43	PSTDCCC050	PSTDCCC050	PD090688.D	18 Oct 2025 02:13		AR\AJ	Ok

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/17/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/16/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/17/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137562

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q3307-02	S-39(9.5-10.0)	1	1.14	10.71	11.85	10.87	90.8	
Q3307-03	S-39(14.5-15.0)	2	1.15	10.38	11.53	10.4	89.1	
Q3307-04	S-37(7.0-7.5)	3	1.13	10.10	11.23	10.06	88.4	
Q3307-05	S-37(14.5-15.0)	4	1.16	10.22	11.38	10.35	89.9	
Q3307-06	S-38(0-0.5)	5	1.19	10.65	11.84	10.77	90.0	
Q3307-07	S-38(6.0-6.5)	6	1.18	10.35	11.53	10.42	89.3	
Q3307-08	S-38(14.5-15.0)	7	1.17	10.18	11.35	10.39	90.6	
Q3307-09	S-36(0-0.5)	8	1.13	10.34	11.47	10.73	92.8	
Q3307-10	S-36(7.0-7.5)	9	1.18	10.53	11.71	10.63	89.7	
Q3307-11	S-36(14.5-15.0)	10	1.13	11.04	12.17	10.97	89.1	
Q3307-17	S-35(7.0-7.5)	11	1.15	10.83	11.98	10.77	88.8	
Q3307-18	S-35(14.5-15.0)	12	1.12	10.40	11.52	10.33	88.6	
Q3363-01	PR132-SOL-085011-20251015	13	1.13	10.82	11.95	6.55	50.1	
Q3363-02	PR132-SOL-039085-20251015	14	1.15	10.53	11.68	7.11	56.6	
Q3363-03	PR132-SOL-028010-20251015	15	1.15	10.29	11.44	6.79	54.8	
Q3363-04	PR132-SOL-010013-20251015	16	1.18	10.94	12.12	6.81	51.5	
Q3363-05	COMP01-20251015	17	1.19	10.32	11.51	6.91	55.4	
Q3366-01	TR-06-101625	18	1.14	10.33	11.47	10.95	95.0	
Q3366-02	TR-06-101625-E2	19	1.15	10.32	11.47	11.00	95.4	
Q3366-03	TR-1-101625	20	1.19	10.30	11.49	10.7	92.3	
Q3366-04	TR-01-101625-E2	21	1.19	10.43	11.62	11.12	95.2	
Q3367-01	BU-03-101625	22	1.11	10.51	11.62	10.81	92.3	
Q3369-01	MH-22	23	1.16	10.61	11.77	10.62	89.2	
Q3369-02	MH-22-EPH	24	1.15	10.38	11.53	10.4	89.1	
Q3369-03	MH-22-VOC	25	1.16	10.36	11.52	10.4	89.2	
Q3369-05	MH-21	26	1.19	10.76	11.95	10.81	89.4	
Q3369-06	MH-21-EPH	27	1.12	11.10	12.22	11.12	90.1	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/17/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/16/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/17/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137562

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q3369-07	MH-21-VOC	28	1.16	10.45	11.61	10.66	90.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

VB 137562

WorkList Name : %1-101625

WorkList ID : 192491

Department : Wet-Chemistry

Date : 10-16-2025 07:46:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3307-02	S-39(9.5-10.0)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-03	S-39(14.5-15.0)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-04	S-37(7.0-7.5)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-05	S-37(14.5-15.0)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-06	S-38(0-0.5)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-07	S-38(6.0-6.5)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-08	S-38(14.5-15.0)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-09	S-36(0-0.5)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-10	S-36(7.0-7.5)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-11	S-36(14.5-15.0)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-17	S-35(7.0-7.5)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3307-18	S-35(14.5-15.0)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/07/2025	Chemtech -SO
Q3363-01	PR132-SOL-085011-20251015	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/15/2025	Chemtech -SO
Q3363-02	PR132-SOL-039085-20251015	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/15/2025	Chemtech -SO
Q3363-03	PR132-SOL-028010-20251015	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/15/2025	Chemtech -SO
Q3363-04	PR132-SOL-010013-20251015	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/15/2025	Chemtech -SO
Q3363-05	COMP01-20251015	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/15/2025	Chemtech -SO
Q3366-01	TR-06-101625	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/16/2025	Chemtech -SO
Q3366-02	TR-06-101625-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/16/2025	Chemtech -SO
Q3366-03	TR-1-101625	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/16/2025	Chemtech -SO
Q3366-04	TR-01-101625-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/16/2025	Chemtech -SO

Date/Time 10-16-25 13:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 10-16-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

137562

WorkList Name : %1-101625 WorkList ID : 192491 Department : Wet-Chemistry Date : 10-16-2025 07:46:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3367-01	BU-03-101625	Solid	Percent Solids	Cool 4 deg C	PSEG05	D11	10/16/2025	Chemtech -SO
Q3369-01	MH-22	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/16/2025	Chemtech -SO
Q3369-02	MH-22-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/16/2025	Chemtech -SO
Q3369-03	MH-22-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/16/2025	Chemtech -SO
Q3369-05	MH-21	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/16/2025	Chemtech -SO
Q3369-06	MH-21-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/16/2025	Chemtech -SO
Q3369-07	MH-21-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/16/2025	Chemtech -SO

Date/Time 10-16-25 15:20 Date/Time 10-16-25 17:30
Raw Sample Received by: [Signature] Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature] Raw Sample Relinquished by: [Signature]

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Florisil

Matrix : Solid

Welgh By: RJ **Extraction By:** RJ

Balance check: RJ **Filter By:** RJ

Balance ID: EX-SC-2 **pH Meter ID:** N/A

pH Strip Lot#: N/A **Hood ID:** 3,7

Extraction Method: ☐ Seperatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Extraction Start Date : 10/17/2025

Extraction Start Time : 08:51

Extraction End Date : 10/17/2025

Extraction End Time : 12:10

Concentration By: RS

Supervisor By : RUPESH

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24766
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2643
Baked Na2SO4	N/A	EP2652
Sand	N/A	E3951
Hexane	N/A	E3978
Florisil	N/A	E3976
9:1 Hexane:Acetone Mixture	N/A	EP2644
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: N/A **Envap ID:** NE VAP-02


KD Bath Temperature: N/A **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/17/25	RS (Ext Lab)	RS (Ext Lab)
12:15	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 10/17/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170142BL	PBLK142	Pesticide-TCL	30.01	N/A	RUPESH	ritesh	10			U1-1
PB170142BS	PLCS142	Pesticide-TCL	30.03	N/A	RUPESH	ritesh	10			2
Q3363-01	PR132-SOL-085011-2025 1015	Pesticide-TCL	30.06	N/A	RUPESH	ritesh	10	E		3
Q3363-02	PR132-SOL-039085-2025 1015	Pesticide-TCL	30.04	N/A	RUPESH	ritesh	10	E		4
Q3363-03	PR132-SOL-028010-2025 1015	Pesticide-TCL	30.08	N/A	RUPESH	ritesh	10	E		5
Q3363-04	PR132-SOL-010013-2025 1015	Pesticide-TCL	30.03	N/A	RUPESH	ritesh	10	E		6
Q3363-04MS	PR132-SOL-010013-2025 1015MS	Pesticide-TCL	30.06	N/A	RUPESH	ritesh	10	E		U2-1
Q3363-04MS D	PR132-SOL-010013-2025 1015MSD	Pesticide-TCL	30.08	N/A	RUPESH	ritesh	10	E		2
Q3363-05	COMP01-20251015	Pesticide-TCL	30.05	N/A	RUPESH	ritesh	10	E		3
Q3366-01	TR-06-101625	Pesticide-TCL	30.07	N/A	RUPESH	ritesh	10	E		4
Q3366-03	TR-1-101625	Pesticide-TCL	30.01	N/A	RUPESH	ritesh	10	E		5
Q3367-01	BU-03-101625	Pesticide-TCL	30.06	N/A	RUPESH	ritesh	10	E		6
Q3369-01	MH-22	Pesticide-TCL	30.09	N/A	RUPESH	ritesh	10	E		U3-1
Q3369-05	MH-21	Pesticide-TCL	30.04	N/A	RUPESH	ritesh	10	E		2




* Extracts relinquished on the same date as received.

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3363P WorkList ID : 192515 Department : Extraction Date : 10-17-2025 08:29:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3363-01	PR132-SOL-085011-20251015	Solid	Pesticide-TCL	Cool 4 deg C	AECO02	D31	10/15/2025	8081B
Q3363-02	PR132-SOL-039085-20251015	Solid	Pesticide-TCL	Cool 4 deg C	AECO02	D31	10/15/2025	8081B
Q3363-03	PR132-SOL-028010-20251015	Solid	Pesticide-TCL	Cool 4 deg C	AECO02	D31	10/15/2025	8081B
Q3363-04	PR132-SOL-010013-20251015	Solid	Pesticide-TCL	Cool 4 deg C	AECO02	D31	10/15/2025	8081B
Q3363-05	COMP01-20251015	Solid	Pesticide-TCL	Cool 4 deg C	AECO02	D31	10/15/2025	8081B
Q3366-01	TR-06-101625	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D31	10/16/2025	8081B
Q3366-03	TR-1-101625	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D31	10/16/2025	8081B
Q3367-01	BU-03-101625	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D11	10/16/2025	8081B
Q3369-01	MH-22	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D31	10/16/2025	8081B
Q3369-05	MH-21	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D31	10/16/2025	8081B

Date/Time 10/17/25 8:46
 Raw Sample Received by: R3 [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10/17/25 9:05
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: R3 [Signature]

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

Prep Standard - Chemical Standard Summary

Order ID : Q3363
Test : Pesticide-TCL
Prepbatch ID : PB170142,
Sequence ID/Qc Batch ID: pd101725,

Standard ID :

EP2643,EP2652,PP24651,PP24663,PP24738,PP24739,PP24740,PP24741,PP24742,PP24744,PP24745,PP24746,PP24747,PP24748,PP24749,PP24750,PP24751,PP24752,PP24753,PP24754,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764,PP24766,PP24942,

Chemical ID :

E3875,E3941,E3944,E3951,E3955,E3956,E3971,E3972,E3976,E3978,P12604,P12610,P13038,P13041,P13196,P13246,P13774,P13780,P13786,P13788,P13862,P9053,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2643	09/16/2025	03/16/2026	Evelyn Huang	None	None	Riteshkumar Patel
09/16/2025								

FROM 4000.00000ml of E3971 + 4000.00000ml of E3972 = Final Quantity: 8000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2652	10/10/2025	01/28/2026	Evelyn Huang	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel
10/10/2025								

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4027	Pesticide resolution Check Mixture 8081	PP24651	06/16/2025	12/11/2025	Abdul Mirza	None	None	Yogesh Patel
								07/22/2025

FROM 1.00000ml of P13246 + 99.00000ml of E3941 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel
								07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3629	20 PPM PEST stock Solution 1st source(RESTEK)	PP24738	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 1.00000ml of P13038 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1472	20 PPM Pest Stock Solution 2nd Source	PP24739	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 1.00000ml of P13041 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1273	20 PPM Mirex Stock (Primary Source)	PP24740	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P9053 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3663	20 PPM MIREX Stock STD (Secondary source)	PP24741	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13196 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24742	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13788 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3630	100/100 PPB PEST Working std.1st Source(RESTEK)	PP24744	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 98.50000ml of E3956 + 0.50000ml of PP24738 + 0.50000ml of PP24740 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
80	100/100 PPB Pesticide Working Solution 2nd Source	PP24745	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 98.50000ml of E3956 + 0.50000ml of PP24739 + 0.50000ml of PP24741 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
386	1000/100 PPB Chlordane STD (Restek)	PP24746	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12604 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3746	1000/100 ppb Chlordane STD-RESTEK 2ND SOURCE	PP24747	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12610 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
383	1000/100 PPB Toxaphene STD (Restek)	PP24748	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P13774 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3669	1000/100 PPB TOXAPHENE STD 2nd source (RESTEK)	PP24749	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.10000ml of P13862 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3631	75 PPB ICAL PEST STD(RESTEK)	PP24750	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.25000ml of E3956 + 0.75000ml of PP24744 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3632	50 PPB ICAL PEST STD(RESTEK)	PP24751	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24744 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3633	25 PPB ICAL PEST STD(RESTEK)	PP24752	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.75000ml of E3956 + 0.25000ml of PP24744 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3634	5 PPB ICAL PEST STD(RESTEK)	PP24753	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.90000ml of E3956 + 0.10000ml of PP24751 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3988	50 PPB PEST ICV STD(RESTEK)	PP24754	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24745 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
528	CHLOR 750 PPB STD	PP24755	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.25000ml of E3956 + 0.75000ml of PP24746 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
529	CHLOR 500 PPB STD	PP24756	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24746 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
530	CHLOR 250 PPB STD	PP24757	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.75000ml of E3956 + 0.25000ml of PP24746 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3408	CHLOR 50 PPB STD	PP24758	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.90000ml of E3956 + 0.10000ml of PP24756 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
532	CHLOR 500 PPB ICV STD	PP24759	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24747 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
533	TOX 750 PPB STD	PP24760	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.25000ml of E3956 + 0.75000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
534	TOX 500 PPB STD	PP24761	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
535	TOX 250 PPB STD	PP24762	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.75000ml of E3956 + 0.25000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2217	TOX 100 PPB STD	PP24763	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.90000ml of E3956 + 0.10000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3670	TOX 500 PPB ICV std (RESTEK)	PP24764	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of PP24749 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
79	500 PPB Pesticide Spike Solution	PP24766	07/28/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								08/19/2025

FROM 95.00000ml of E3955 + 2.50000ml of PP24739 + 2.50000ml of PP24741 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
758	PEM Mix w/Surr	PP24942	09/24/2025	03/24/2026	Abdul Mirza	None	None	Yogesh Patel
								10/10/2025

FROM 1.00000ml of P13780 + 99.00000ml of E3971 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	12/11/2025	06/11/2025 / Rajesh	06/04/2025 / Rajesh	E3941

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	25A2756718	12/31/2028	07/09/2025 / RUPESH	04/28/2020 / RUPESH	E3951

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	04/18/2027	07/16/2025 / RUPESH	07/16/2025 / RUPESH	E3955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/16/2025 / RUPESH	07/16/2025 / RUPESH	E3956

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	09/15/2025 / Evelyn	09/04/2025 / Riteshkumar	E3971

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agela Technologies Inc.	FS0006 / Cleanert Florisil cartridge	M07456	04/03/2026	10/03/2025 / Evelyn	10/03/2025 / Evelyn	E3976

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/13/2026	10/13/2025 / RUPESH	10/08/2025 / Riteshkumar	E3978

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12604

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12610

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0200423	01/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13038

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0199099	07/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13041

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	042022	01/21/2026	07/21/2025 / Abdul	01/17/2024 / Abdul	P13196

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	19161 / 8081 pesticide resolution check mixture	013124	12/17/2025	06/17/2025 / Abdul	02/09/2024 / Abdul	P13246

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203038	01/21/2026	07/21/2025 / Abdul	05/03/2024 / Ankita	P13774

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32074 / Pesticide Performance Evaluation Mix w/Surrogate	A0213999	03/24/2026	09/24/2025 / Abdul	11/19/2024 / Ankita	P13780

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	01/21/2026	07/21/2025 / Abdul	11/19/2024 / Ankita	P13788

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0210240	01/21/2026	07/21/2025 / Abdul	12/09/2024 / Abdul	P13862

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	112018	01/21/2026	07/21/2025 / Abdul	11/01/2019 / Stephen	P9053



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Harout Sahagian

Recd. by RI on 6/11/25

E3941

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Material	BDH9274-2.5KG
Material Description	BDH SAND STDD OTTAWA W+I 2.5KG
Grade	NOT APPLICABLE
Batch	25A2756718
Reassay Date	12/31/2028
CAS Number	14808-60-7
Molecular Formula	SiO ₂
Molecular Mass	60.09
Date of Manufacture	12/05/2024
Storage	Room Temperature

Characteristics	Specifications	Measured Values
Appearance	Beige granules.	Beige granules.
Moisture	<= 0.1 %	0.1 %
Particle Size 30-40 mesh	>= 80 %	99 %
CUSTOMER PART # BDH9274-2.5KG		

Received on 7/9/25.

E3951

Internal ID #: 793

Signature	Additional Information
We certify that this batch conforms to the specifications listed above. This document has been electronically produced and is valid without a signature. Leona Edwardson, Quality Control Sr. Manager - Solon VWR Chemicals, LLC. 28600 Fountain Parkway, Solon OH 44139 USA	Analysis may have been rounded to significant digits in specification limits Product meets analytical specifications of the grades listed.

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/16/25

E 3955

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/16/25

E3956

Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3971

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700



Cleanert Florisil

1g/6ml 30/pkg

LOT#:M07456

MFG#:G02147

CAT#FS0006



固相萃取产品

Made in China

Agela Technologies

E 3976



n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors**TM



Material No.: 9262-03

Batch No.: 25C0362006

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	4
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5\%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95\%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3978



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0199099

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

P130397 5
↓
P13043 1
✓
12-26-2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.0 µg/mL	+/- 8.9732
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	200.1 µg/mL	+/- 8.9762
3	beta-BHC	319-85-7	BCCC6425	99%	200.3 µg/mL	+/- 8.9844
4	delta-BHC	319-86-8	14450800	98%	200.0 µg/mL	+/- 8.9740
5	Heptachlor	76-44-8	813251	99%	200.1 µg/mL	+/- 8.9754
6	Aldrin	309-00-2	14389400	98%	200.0 µg/mL	+/- 8.9718
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.1 µg/mL	+/- 8.9754
8	trans-Chlordane	5103-74-2	32943	98%	199.9 µg/mL	+/- 8.9696
9	cis-Chlordane	5103-71-9	31766	98%	200.1 µg/mL	+/- 8.9762
10	Endosulfan I	959-98-8	BCCF4060	99%	200.1 µg/mL	+/- 8.9754
11	4,4'-DDE	72-55-9	GHYQG	99%	200.1 µg/mL	+/- 8.9777
12	Dieldrin	60-57-1	11129900	98%	200.0 µg/mL	+/- 8.9718
13	Endrin	72-20-8	14123200	98%	199.9 µg/mL	+/- 8.9696
14	4,4'-DDD	72-54-8	HAN02	99%	200.1 µg/mL	+/- 8.9777
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	200.0 µg/mL	+/- 8.9718

17	Endrin aldehyde	7421-93-4	30720	98%	200.1 µg/mL	+/- 8.9784
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.0 µg/mL	+/- 8.9732
19	Methoxychlor	72-43-5	13668200	99%	200.1 µg/mL	+/- 8.9777
20	Endrin ketone	53494-70-5	1-ABS-16-7	98%	200.0 µg/mL	+/- 8.9740

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13039
↓
P13043
1
JAW
12/26/23

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Josh McCloskey
Josh McCloskey - Operations Technician I

Date Mixed: 19-Jun-2023 **Balance Serial #** 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 23-Jun-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number: **79136**
Lot Number: **042022**
Description: **Mirex**Solvent(s):
Acetone Lot#
81025

		042022
Formulated By:	Prashant Chauhan	DATE
		042022
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 042027

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 1000

NIST Test ID#: 6UTB

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 50.0 0.006 Flask Uncertainty

Expanded
SDS Information
(Solvent Safety Info. On Attached pg.)

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Mirex	437	9492400	1000	99.4	0.5	0.05034	0.05040	1001.1	10.3	2385-85-5	N/A	ori-rat 306mg/kg

1. Mirex

437

9492400

1000

99.4

0.5

0.05034

0.05040

1001.1




10.3

2385-85-5

N/A

ori-rat 306mg/kg

Method GC7MSD-1.M: Column: SPB-608 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 150°C (4min.), Temp 2 = 290°C (13.5 min.), Rate = 8°C/min., Injector B = 200°C, Detector B = 290°C. Split Ratio = 100:1, Scan Rate = 2. Analysis performed by Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5 % of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

01/17/2024

✓
True

D
C/D
D ← D
A C/D S
S

(5)



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Ship: Ambient

P 12603
↓
P 12605
3
Kauf
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

Solvent: Hexane
CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician


Date Mixed: 11-May-2023

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 12603 } (3)
↓
P 12605

7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *chromatographic plus*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Ship: Ambient

P 12606
P 12610
5 Five
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. { hold 10 min.}

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 11-May-2023

Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0200423

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 10°C or colder

Ship: Ambient

P 13034
↓
P 13038
12.26.2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.5 µg/mL	+/- 8.9956
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	199.9 µg/mL	+/- 8.9696
3	beta-BHC	319-85-7	BCCC6425	99%	200.0 µg/mL	+/- 8.9732
4	delta-BHC	319-86-8	14450800	98%	199.9 µg/mL	+/- 8.9696
5	Heptachlor	76-44-8	813251	99%	202.0 µg/mL	+/- 9.0629
6	Aldrin	309-00-2	14389400	98%	200.9 µg/mL	+/- 9.0136
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.0 µg/mL	+/- 8.9732
8	trans-Chlordane	5103-74-2	34616	99%	200.5 µg/mL	+/- 8.9956
9	cis-Chlordane	5103-71-9	31766	98%	201.4 µg/mL	+/- 9.0356
10	Endosulfan I	959-98-8	BCCF4060	99%	200.0 µg/mL	+/- 8.9732
11	4,4'-DDE	72-55-9	GHYQG	99%	201.5 µg/mL	+/- 9.0405
12	Dieldrin	60-57-1	14515000	98%	199.9 µg/mL	+/- 8.9696
13	Endrin	72-20-8	14485300	98%	200.4 µg/mL	+/- 8.9916
14	4,4'-DDD	72-54-8	HAN02	99%	200.5 µg/mL	+/- 8.9956
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	201.9 µg/mL	+/- 9.0575

17	Endrin aldehyde	7421-93-4	30720	98%	201.4 µg/mL	+/- 9.0356
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.5 µg/mL	+/- 8.9956
19	Methoxychlor	72-43-5	14563200	98%	200.9 µg/mL	+/- 9.0136
20	Endrin ketone	53494-70-5	14537700	98%	199.9 µg/mL	+/- 8.9696

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13034
P13038
1
5
12/26/2023

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 31-Jul-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 03-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number: **19161**
 Lot Number: **013124**
 Description: **CLP Pesticides & PCB's Resolution Check Standard**
 9 components
 Expiration Date: **013129**
 Recommended Storage: **Refrigerate (4 °C)**
 Nominal Concentration (µg/mL): **Varied**
 NIST Test ID#: **6UTB**
 Volume(s) shown below were combined and diluted to (mL): **100.0**

Solvent(s):
 Hexane
 Toluene

Lot#
 273615 (50%)
 28508 (50%)

5E-05 Balance Uncertainty
 0.021 Flask Uncertainty

		013124
Formulated By:	Lawrence Barry	DATE
		013124
Reviewed By:	Pedro L. Rentas	DATE

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(ug/mL)	Final Conc.(ug/mL)	Expanded Uncertainty (+/-) µg/mL	SDS Information (Solvent Safety Info. On Attached pg.)		
									CAS#	OSHA PEL (TWA)	LD50
1. trans-Chlordane	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	5103-74-2	0.5mg/m3 (skin)	ori-rat 500mg/kg
2. Endosulfan I	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	959-98-8	0.1mg/m3 (skin)	ori-rat 18mg/kg
3. 4,4'-DDE	19361	013124	0.010	1.00	0.004	201.6	2.0	0.03	72-55-9	N/A	ori-rat 880mg/kg
4. Dieldrin	19361	013124	0.010	1.00	0.004	202.8	2.0	0.03	60-57-1	0.25mg/m3 (skin)	ori-rat 38300ug/kg
5. Endosulfan sulfate	19361	013124	0.010	1.00	0.004	204.2	2.0	0.03	1031-07-8	N/A	ori-rat 18mg/kg
6. Endrin ketone	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	53494-70-5	N/A	N/A
7. 4,4'-Methoxychlor	19361	013124	0.010	1.00	0.004	1000.7	10.0	0.09	72-43-5	10mg/m3	ori-rat 6000mg/kg
8. 2,4,5,6-Tetrachloro-m-xylene	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	877-09-8	N/A	N/A
9. Decachlorobiphenyl (209)	19361	013124	0.010	1.00	0.004	202.0	2.0	0.03	2051-24-3	N/A	N/A

• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 • Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
 • Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 • All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 • Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

P13243 } (5)
 ↓
 P13247 }
 ↓
 02/9/2024



110 Benner Circle
Belleville, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	—%	1,009.0 µg/mL	+/- 55.9920

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13773
↓
P13779
AI
11/19/24

05-06-2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 10-Oct-2023

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

Handwritten notes:

- P13388
- P13389
- P13773
- P13779
- AJ
- 11/19/24
- 05-06-2024



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32074 **Lot No.:** A0213999

Description : Pesticide Performance Eval Mix w/Surrogate

Performance Evaluation Std. 3/90 SOW w/surrogates 1-25µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2028 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13780
↓
P13784
AJ
11/19/20

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	2.0 µg/mL	+/- 0.1506
2	alpha-BHC	319-84-6	15279000	99%	1.0 µg/mL	+/- 0.0749
3	gamma-BHC (Lindane)	58-89-9	14931300	97%	1.0 µg/mL	+/- 0.0756
4	beta-BHC	319-85-7	BCCC6425	99%	1.0 µg/mL	+/- 0.0749
5	Endrin	72-20-8	15076100	98%	5.0 µg/mL	+/- 0.3769
6	4,4'-DDT	50-29-3	231103JLM	97%	10.0 µg/mL	+/- 0.7512
7	Methoxychlor	72-43-5	14979500	98%	24.8 µg/mL	+/- 1.8543
8	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	2.0 µg/mL	+/- 0.1513

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

150°C to 300°C
@ 4°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 50:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan
Malina Homan - Operations Technician I

Date Mixed: 17-Jul-2024

Balance Serial # 1128353505

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Jul-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
↓
P19789 AJ
11/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

heliu-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A.O.E.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024

Balance Serial # B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
↓
P19789 AJ
11/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A.O.E.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024

Balance Serial # B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0210240
Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2028 **Storage:** 10°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.3 µg/mL	+/- 56.0105

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13861
P13862

12/9/2024

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Amanda Miller - Operations Tech III - ARM QC


Date Mixed: 11-Apr-2024 **Balance Serial #** B442140311


Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 26-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13861 } ②
P13862 }
↑

12/9/2024



CERTIFIED WEIGHT REPORT

Part Number: **72072**
Lot Number: **112018**
Description: **n-Tetracosane-d50**

Solvent(s):
Methylene chloride
Lot#: 102669

Expiration Date: 112028
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 1000
NIST Test ID#: 2684186

Weight(s) shown below were combined and diluted to (mL):



200.0 0.058 Balance Uncertainty
0.058 Flask Uncertainty

Received by
SG on 11/1/19
P9044 - P9053

		112018
Formulated By:	Prashant Chauhan	DATE
		112018
Reviewed By:	Pedro Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. n-Tetracosane-d50	2072	PR-17753/09216TC1	1000	98	0.2	0.20411	0.20415	1000.2	4.2	16416-32-3	N/A	N/A

Method GC8MSD-3.M: Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 250°C, Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Run 40, "P72072 L112018 [1000µg/mL in MeCl2]"

Run Length: 35.00 min, 20999 points at 10 points/second.

Created: Thu, Nov 22, 2018 at 7:23:18 AM.

Sampled: Sequence "112018-GC4M1", Method "GC4-M1".

Analyzed using Method "GC4-M1".

Comments

GC4-M1 Analysis by Melissa Stonier


Column ID SPB5 L#60062-01A : 30 meter x 0.53mm x 1.5µm Film Thickness

Flow rates: Total Flow = 300 mL/min, Helium (carrier) = 6.5 mL, Helium (make-up) = 25 mL, Hydrogen (detector) = 30 mL, Air (detector) = 360 mL

Oven Temp 1 = 50°C (1 min), Rate = 10°C/min, Oven Temp 2 = 300°C (9 min), Total Run Time = 35 Minutes.

Injector Temp = 200°C, FID Temp = 300°C, FID Signal = eDag Channel 1.

Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 µL, Range = 3





SHIPPING DOCUMENTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: AECOM
ADDRESS: 605 Third Avenue
CITY: New York STATE: NY ZIP: 10158
ATTENTION: Robert Forstner
PHONE: 212-377-8721 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Port Richmond 132
PROJECT NO.: 60693795 LOCATION: Staten Island
PROJECT MANAGER: Rob Forstner
e-mail: robert.forstner@aecom.com
PHONE: 212-377-8721 FAX:

CLIENT BILLING INFORMATION

BILL TO: AECOM PO#:
ADDRESS: 605 Third Avenue
CITY: New York STATE: NY ZIP: 10158
ATTENTION: Rob Forstner PHONE: 212-377-8721

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

- ☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☒ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1. SVOCs
2. Pesticides-TCLs
3. Mercury-Metals
4. PCBs / Hexachloro
5. Dioxin
6. Grain Size
7. TOC
8. VOCs
9.

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES										
			COMP	GRAB	DATE	TIME		E	E	E	E	E	E	E	F			
								1	2	3	4	5	6	7	8	9		
1.	PR132-S01-085011-20251015	Soil		X	10/15	1157	11	X	X	X	X	X	X	X	X			
2.	PR132-S01-039085-20251015	Soil		X	10/15	1210	11	X	X	X	X	X	X	X	X			
3.	PR132-S03-028010-20251015	Soil		X	10/15	1315	11	X	X	X	X	X	X	X	X			
4.	PR132-S03-010013-20251015	Soil		X	10/15	1330	11	X	X	X	X	X	X	X	X			
5.	COMP01-20251015	Soil	X		10/15	1353	11	X	X	X	X	X	X	X	X			
6.	TB-01	Water			10/15		1	X	X	X	X	X	X	X	X	X		Temp Blank
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. Jim Alderson DATE/TIME: 1700 10/15/25 RECEIVED BY: [Signature] 1700 10-15-25
RELINQUISHED BY SAMPLER: 2. [Signature] DATE/TIME: [Signature] RECEIVED BY: [Signature]
RELINQUISHED BY SAMPLER: 3. [Signature] DATE/TIME: 1750 10-15-25 RECEIVED BY: [Signature]

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.1 °C
Comments:

Page ____ of CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3363	AECO02	Order Date : 10/16/2025 10:56:42 AM	Project Mgr :
Client Name : AECOM		Project Name : AB1-CTY 3.2.Z-PR-DOCK	Report Type : NYS ASP <i>Ed</i>
Client Contact : Rob Forstner		Receive DateTime : 10/16/2025 5:50:00 PM	EDD Type : EQUIS
Invoice Name : AECOM		Purchase Order : 10/15/2025	Hard Copy Date :
Invoice Contact : Rob Forstner			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3363-01	PR132-SO 1 ₄ -085011-20251015	Solid	10/15/2025	11:57					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3363-02	PR132-SO 1 ₄ -039085-20251015	Solid	10/15/2025	12:10					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3363-03	PR132-SO 1 ₃ -028010-20251015	Solid	10/15/2025	13:15					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3363-04	PR132-SO 1 ₃ -010013-20251015	Solid	10/15/2025	13:30					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3363-05	COMP01-20251015	Solid	10/15/2025	13:53					
					VOC-TCLVOA-10		8260D	10 Bus. Days	

Relinquished By :

Date / Time : 10/16/25 140

Received By :

Date / Time : 10/16/25

Storage Area : VOA Refridgerator Room

12:18