

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3363

Client ID : SSTDCCC040

Date Analyzed: 10/17/2025

Lab File ID: BG064531.D

Time Analyzed: 11:39

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	15109	7.82	59855	10.61	45410	14.45
UPPER LIMIT	30218	8.32	119710	11.105	90820	14.947
LOWER LIMIT	7554.5	7.32	29927.5	10.105	22705	13.947
EPA SAMPLE NO.						
01 PB170144BL	15169	7.82	63951	10.60	47990	14.44
02 PB170144BS	13548	7.82	60553	10.60	47398	14.45
03 PR132-S01-085011-20251015	15692	7.82	62273	10.60	44210	14.44
04 COMP01-20251015	13378	7.82	53157	10.61	39158	14.44
05 COMP01-20251015MS	14953	7.82	59781	10.60	44289	14.44
06 COMP01-20251015MSD	12958	7.82	53760	10.61	42324	14.45
07 PR132-S01-039085-20251015	14350	7.82	59289	10.60	41893	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.