

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143874.D
Instrument ID: BNA_F

Contract: AECO02
SDG NO.: Q3363
DFTPP Injection Date: 10/06/2025
DFTPP Injection Time: 09:30

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	4.3
441	Present, but less than mass 443	78
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF143875.D	10/06/2025	09:59
SSTDICC005	SSTDICC005	BF143876.D	10/06/2025	10:28
SSTDICC010	SSTDICC010	BF143877.D	10/06/2025	10:57
SSTDICC020	SSTDICC020	BF143878.D	10/06/2025	11:56
SSTDICCC040	SSTDICCC040	BF143879.D	10/06/2025	12:54
SSTDICC050	SSTDICC050	BF143880.D	10/06/2025	13:23
SSTDICC060	SSTDICC060	BF143881.D	10/06/2025	13:53
SSTDICC080	SSTDICC080	BF143882.D	10/06/2025	14:22