

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG No.: Q3363
 Instrument ID: MSVOA_W Calibration Date/Time: 10/20/2025 10:20
 Lab File ID: VW032378.D Init. Calib. Date(s): 10/01/2025 10/01/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:12 15:20
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	54.56	-9.124001	20