

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100125\
 Data File : VW032314.D
 Acq On : 01 Oct 2025 13:34
 Operator : SY/MD
 Sample : VSTDICC010
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICC010

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025
 Supervised By :Semsettin Yesilyurt 10/03/2025

Quant Time: Oct 02 21:13:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:10:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	280850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	525265	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	471729	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	242027	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	42595	10.398	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery =	20.800%	#	
35) Dibromofluoromethane	7.898	113	35230	10.357	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery =	20.720%	#	
50) Toluene-d8	10.325	98	126526	10.078	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery =	20.160%	#	
62) 4-Bromofluorobenzene	12.617	95	47150	9.936	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery =	19.880%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.039	85	19283	10.941	ug/l	98
3) Chloromethane	2.259	50	29583	10.879	ug/l	97
4) Vinyl Chloride	2.405	62	35737	11.008	ug/l	98
5) Bromomethane	2.826	94	26494	11.105	ug/l	90
6) Chloroethane	2.972	64	22994	10.728	ug/l	98
7) Trichlorofluoromethane	3.308	101	25472	10.479	ug/l	99
8) Diethyl Ether	3.722	74	22779	10.814	ug/l	86
9) 1,1,2-Trichlorotrifluo...	4.100	101	31968	10.965	ug/l	94
10) Methyl Iodide	4.307	142	43646	10.492	ug/l	98
11) Tert butyl alcohol	5.216	59	17352	55.962	ug/l #	80
12) 1,1-Dichloroethene	4.082	96	34142	10.786	ug/l	97
13) Acrolein	3.929	56	27390	54.429	ug/l	97
14) Allyl chloride	4.710	41	55724	10.423	ug/l	89
15) Acrylonitrile	5.405	53	56775	51.811	ug/l	99
16) Acetone	4.161	43	64630	56.980	ug/l	98
17) Carbon Disulfide	4.423	76	88918	10.296	ug/l	98
18) Methyl Acetate	4.716	43	33520	10.316	ug/l	98
19) Methyl tert-butyl Ether	5.466	73	65224	10.462	ug/l	97
20) Methylene Chloride	4.960	84	59087	11.634	ug/l	94
21) trans-1,2-Dichloroethene	5.460	96	35963	10.262	ug/l	95
22) Diisopropyl ether	6.344	45	117148	10.536	ug/l #	94
23) Vinyl Acetate	6.283	43	372258	50.126	ug/l	98
24) 1,1-Dichloroethane	6.246	63	70852	10.489	ug/l	98
25) 2-Butanone	7.197	43	82281	53.263	ug/l	97
26) 2,2-Dichloropropane	7.191	77	36712	10.625	ug/l	99
27) cis-1,2-Dichloroethene	7.191	96	42849	10.178	ug/l	93
28) Bromochloromethane	7.532	49	33043	10.397	ug/l	87
29) Tetrahydrofuran	7.557	42	49488	50.941	ug/l	96
30) Chloroform	7.691	83	71607	10.453	ug/l	98
31) Cyclohexane	7.971	56	62233	10.792	ug/l #	93
32) 1,1,1-Trichloroethane	7.886	97	51138	10.534	ug/l	97
36) 1,1-Dichloropropene	8.099	75	49414	10.293	ug/l	98
37) Ethyl Acetate	7.276	43	33130	10.368	ug/l	99
38) Carbon Tetrachloride	8.087	117	46310	10.251	ug/l	94
39) Methylcyclohexane	9.343	83	56991	9.842	ug/l	92
40) Benzene	8.337	78	156957	10.431	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100125\
 Data File : VW032314.D
 Acq On : 01 Oct 2025 13:34
 Operator : SY/MD
 Sample : VSTDICC010
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICC010

Quant Time: Oct 02 21:13:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:10:58 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025
 Supervised By :Semsettin Yesilyurt 10/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.502	41	16618	9.650	ug/l	98
42) 1,2-Dichloroethane	8.416	62	52201	10.602	ug/l	99
43) Isopropyl Acetate	8.441	43	56692	10.001	ug/l	98
44) Trichloroethene	9.105	130	36599	10.387	ug/l	87
45) 1,2-Dichloropropane	9.374	63	38953	10.371	ug/l	100
46) Dibromomethane	9.465	93	24036	10.407	ug/l	94
47) Bromodichloromethane	9.654	83	54338	10.286	ug/l	99
48) Methyl methacrylate	9.447	41	27343	9.921	ug/l	94
49) 1,4-Dioxane	9.465	88	6100	193.232	ug/l #	64
51) 4-Methyl-2-Pentanone	10.215	43	170455	51.916	ug/l	98
52) Toluene	10.392	92	95076	10.146	ug/l	99
53) t-1,3-Dichloropropene	10.611	75	50624	9.878	ug/l	99
54) cis-1,3-Dichloropropene	10.081	75	59970	10.158	ug/l	97
55) 1,1,2-Trichloroethane	10.794	97	32248	10.447	ug/l	95
56) Ethyl methacrylate	10.648	69	42126	9.750	ug/l	93
57) 1,3-Dichloropropane	10.934	76	57249	10.501	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.928	63	104587	45.484	ug/l	95
59) 2-Hexanone	10.971	43	118088	50.996	ug/l	98
60) Dibromochloromethane	11.129	129	34371	9.806	ug/l	97
61) 1,2-Dibromoethane	11.233	107	31025	10.308	ug/l	97
64) Tetrachloroethene	10.867	164	30629	10.851	ug/l	95
65) Chlorobenzene	11.660	112	107554	10.362	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.733	131	33104	10.240	ug/l	97
67) Ethyl Benzene	11.733	91	175582	10.149	ug/l	95
68) m/p-Xylenes	11.837	106	135895	20.370	ug/l	98
69) o-Xylene	12.166	106	62715	10.052	ug/l	98
70) Styrene	12.178	104	107824	9.899	ug/l	99
71) Bromoform	12.349	173	19572	10.562	ug/l #	92
73) Isopropylbenzene	12.464	105	152894	9.234	ug/l	99
74) N-amyl acetate	12.269	43	49718	9.412	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.714	83	41991	10.213	ug/l	100
76) 1,2,3-Trichloropropane	12.763	75	31741m	9.939	ug/l	
77) Bromobenzene	12.745	156	38105	9.547	ug/l	84
78) n-propylbenzene	12.800	91	194851	9.421	ug/l	97
79) 2-Chlorotoluene	12.891	91	124638	9.855	ug/l	98
80) 1,3,5-Trimethylbenzene	12.940	105	135870	9.490	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.513	75	11399	9.067	ug/l	91
82) 4-Chlorotoluene	12.989	91	132770	9.874	ug/l	99
83) tert-Butylbenzene	13.202	119	113595	9.669	ug/l	99
84) 1,2,4-Trimethylbenzene	13.245	105	139474	9.580	ug/l	98
85) sec-Butylbenzene	13.379	105	171153	9.533	ug/l	99
86) p-Isopropyltoluene	13.495	119	139064	9.335	ug/l	99
87) 1,3-Dichlorobenzene	13.495	146	79860	9.881	ug/l	94
88) 1,4-Dichlorobenzene	13.574	146	80973	10.019	ug/l	98
89) n-Butylbenzene	13.818	91	136393	9.378	ug/l	99
90) Hexachloroethane	14.086	117	25898	9.699	ug/l	98
91) 1,2-Dichlorobenzene	13.867	146	74980	10.262	ug/l	94
92) 1,2-Dibromo-3-Chloropr...	14.482	75	6955	9.662	ug/l	87
93) 1,2,4-Trichlorobenzene	15.129	180	40772	9.306	ug/l	95
94) Hexachlorobutadiene	15.226	225	20605	10.287	ug/l	96
95) Naphthalene	15.360	128	92896	8.829	ug/l	99
96) 1,2,3-Trichlorobenzene	15.555	180	39577	9.584	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100125\
Data File : VW032314.D
Acq On : 01 Oct 2025 13:34
Operator : SY/MD
Sample : VSTDICC010
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTDICC010

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025
Supervised By :Semsettin Yesilyurt 10/03/2025

Quant Time: Oct 02 21:13:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:10:58 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100125\
Data File : VW032314.D
Acq On : 01 Oct 2025 13:34
Operator : SY/MD
Sample : VSTDIC010
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTDIC010

Quant Time: Oct 02 21:13:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:10:58 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/03/2025
Supervised By :Semsettin Yesilyurt 10/03/2025

