

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3363</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>AECOM</u>	Datafile :	<u>VY023484.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1016SBSD01	Dichlorodifluoromethane	20	21.5	ug/Kg	108	0		64	136	20
	Chloromethane	20	20.3	ug/Kg	102	2		52	151	20
	Vinyl chloride	20	21.2	ug/Kg	106	5		56	148	20
	Bromomethane	20	21.6	ug/Kg	108	5		58	141	20
	Chloroethane	20	21.5	ug/Kg	108	2		69	130	20
	Trichlorofluoromethane	20	20.4	ug/Kg	102	4		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	21.5	ug/Kg	108	3		81	123	20
	1,1-Dichloroethene	20	20.8	ug/Kg	104	1		79	121	20
	Acetone	100	110	ug/Kg	110	31	*	40	171	20
	Carbon disulfide	20	20.9	ug/Kg	104	0		59	130	20
	Methyl tert-butyl Ether	20	17.4	ug/Kg	87	15		77	129	20
	Methyl Acetate	20	15.3	ug/Kg	77	21	*	69	149	20
	Methylene Chloride	20	19.1	ug/Kg	96	5		72	131	20
	trans-1,2-Dichloroethene	20	21.1	ug/Kg	106	0		80	123	20
	1,1-Dichloroethane	20	20.9	ug/Kg	104	2		82	123	20
	Cyclohexane	20	19.6	ug/Kg	98	0		76	122	20
	2-Butanone	100	94.8	ug/Kg	95	31	*	69	131	20
	Carbon Tetrachloride	20	21.9	ug/Kg	110	1		76	129	20
	cis-1,2-Dichloroethene	20	20.6	ug/Kg	103	1		82	123	20
	Bromochloromethane	20	19.9	ug/Kg	100	8		80	127	20
	Chloroform	20	21.0	ug/Kg	105	3		82	125	20
	1,1,1-Trichloroethane	20	21.5	ug/Kg	108	2		80	126	20
	Methylcyclohexane	20	19.8	ug/Kg	99	1		77	123	20
	Benzene	20	21.5	ug/Kg	108	0		84	121	20
	1,2-Dichloroethane	20	19.9	ug/Kg	100	6		81	126	20
	Trichloroethene	20	21.2	ug/Kg	106	1		83	122	20
	1,2-Dichloropropane	20	21.0	ug/Kg	105	2		83	122	20
	Bromodichloromethane	20	21.2	ug/Kg	106	4		82	123	20
	4-Methyl-2-Pentanone	100	81.3	ug/Kg	81	30	*	70	135	20
	Toluene	20	21.3	ug/Kg	106	3		83	122	20
	t-1,3-Dichloropropene	20	19.3	ug/Kg	97	7		78	124	20
	cis-1,3-Dichloropropene	20	20.3	ug/Kg	102	4		81	122	20
	1,1,2-Trichloroethane	20	19.8	ug/Kg	99	11		82	125	20
	2-Hexanone	100	90.6	ug/Kg	91	27	*	66	138	20
	Dibromochloromethane	20	20.2	ug/Kg	101	9		79	125	20
	1,2-Dibromoethane	20	19.0	ug/Kg	95	13		80	125	20
	Tetrachloroethene	20	21.4	ug/Kg	107	4		83	125	20
	Chlorobenzene	20	21.1	ug/Kg	106	1		84	122	20
	Ethyl Benzene	20	20.8	ug/Kg	104	2		82	124	20
	m/p-Xylenes	40	43.1	ug/Kg	108	3		83	124	20
	o-Xylene	20	20.5	ug/Kg	103	1		83	123	20
	Styrene	20	21.1	ug/Kg	106	1		82	124	20
	Bromoform	20	19.5	ug/Kg	98	11		75	127	20
	Isopropylbenzene	20	20.9	ug/Kg	104	6		82	124	20
	1,1,2,2-Tetrachloroethane	20	18.6	ug/Kg	93	9		77	127	20
	1,3-Dichlorobenzene	20	21.0	ug/Kg	105	4		83	122	20
	1,4-Dichlorobenzene	20	21.0	ug/Kg	105	4		84	121	20
	1,2-Dichlorobenzene	20	20.5	ug/Kg	103	1		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3363</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>AECOM</u>	Datafile :	<u>VY023484.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1016SBSD01	1,2-Dibromo-3-Chloropropane	20	17.1	ug/Kg	86	18		66	134	20
	1,2,4-Trichlorobenzene	20	18.3	ug/Kg	92	3		78	127	20
	1,2,3-Trichlorobenzene	20	17.9	ug/Kg	90	3		70	137	20