

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3363
 Lab File ID: VW032378.D Date Analyzed: 10/20/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	276051	7.97	500417	8.86	466876	11.63
UPPER LIMIT	552102	8.465	1000830	9.355	933752	12.129
LOWER LIMIT	138026	7.465	250209	8.355	233438	11.129
EPA SAMPLE NO.						
COMP01-20251015	173392	7.96	385425	8.85	401644	11.63
VW1020SBL01	172114	7.96	392797	8.85	400571	11.63
VW1020SBS01	280998	7.96	517149	8.85	480746	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.