

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3363
 Lab File ID: VY023481.D Date Analyzed: 10/16/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:12
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	326705	13.346				
UPPER LIMIT	653410	13.846				
LOWER LIMIT	163353	12.846				
EPA SAMPLE NO.						
PR132-S01-085011-20251015	376450	13.35				
PR132-S01-039085-20251015	423616	13.35				
PR132-S03-028010-20251015	475817	13.35				
PR132-S03-010013-20251015	468564	13.35				
VY1016SBL01	465964	13.35				
VY1016SBS01	360043	13.35				
VY1016SBSD01	331485	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.