

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab File ID: VW032312.D

BFB Injection Date: 10/01/2025

Instrument ID: MSVOA_W

BFB Injection Time: 08:12

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.3
75	30.0 - 60.0% of mass 95	51.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 100.0% of mass 95	66.2
175	5.0 - 9.0% of mass 174	5.3 (8) 1
176	95.0 - 101.0% of mass 174	63.3 (95.6) 1
177	5.0 - 9.0% of mass 176	4.2 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VW032313.D	10/01/2025	13:12
VSTDIC010	VSTDIC010	VW032314.D	10/01/2025	13:34
VSTDIC020	VSTDIC020	VW032315.D	10/01/2025	14:15
VSTDIC050	VSTDIC050	VW032316.D	10/01/2025	14:37
VSTDIC100	VSTDIC100	VW032317.D	10/01/2025	14:59
VSTDIC150	VSTDIC150	VW032318.D	10/01/2025	15:20