

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137567

Review By	rubina	Review On	10/21/2025 12:15:52 PM
Supervise By	Iwona	Supervise On	10/21/2025 12:17:45 PM
SubDirectory	LB137567	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP115135,WP115136,WP115137,WP115138,WP115139,WP115140,WP115141		
ICV Standard	WP115142		
CCV Standard	WP115236,WP115242		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115237,WP115243		
Chk Standard	WP115143,WP115144		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	10/09/25 10:46	All standards, samples, and	RM/IZ	OK
2	STD2	STD2	CAL2	10/09/25 11:08	QC are filtered through	RM/IZ	OK
3	STD3	STD3	CAL3	10/09/25 11:29	0.45um, filter lot W3160	RM/IZ	OK
4	STD4	STD4	CAL4	10/09/25 11:51		RM/IZ	OK
5	STD5	STD5	CAL5	10/09/25 12:12		RM/IZ	OK
6	STD6	STD6	CAL6	10/09/25 12:34		RM/IZ	OK
7	STD7	STD7	CAL7	10/09/25 12:55		RM/IZ	OK
8	ICV1	ICV1	ICV	10/09/25 13:16		RM/IZ	OK
9	ICB1	ICB1	ICB	10/09/25 13:38		RM/IZ	OK
10	CCV1	CCV1	CCV	10/16/25 10:11		RM/IZ	OK
11	CCB1	CCB1	CCB	10/16/25 10:33		RM/IZ	OK
12	LB137567BLW	LB137567BLW	MB	10/16/25 10:54		RM/IZ	OK
13	LB137567BSW	LB137567BSW	LCS	10/16/25 11:16		RM/IZ	OK
14	Q3365-01	MW25	SAM	10/16/25 14:17	Cl is high, Need dilution	RM/IZ	Dilution
15	Q3365-02	MW2D	SAM	10/16/25 14:39	Cl is high, Need dilution	RM/IZ	Dilution
16	Q3365-03	SUPPLY	SAM	10/16/25 15:01	Cl is high, Need dilution	RM/IZ	Dilution
17	Q3365-04	17 HARVEY	SAM	10/16/25 15:22	Cl is high, Need dilution	RM/IZ	Dilution
18	Q3365-05	19 HARVEY	SAM	10/16/25 15:44	Cl is high, Need dilution	RM/IZ	Dilution

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137567

Review By	rubina	Review On	10/21/2025 12:15:52 PM
Supervise By	Iwona	Supervise On	10/21/2025 12:17:45 PM
SubDirectory	LB137567	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP115135,WP115136,WP115137,WP115138,WP115139,WP115140,WP115141		
ICV Standard	WP115142		
CCV Standard	WP115236,WP115242		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115237,WP115243		
Chk Standard	WP115143,WP115144		

19	Q3365-05MS	19 HARVEYMS	MS	10/16/25 16:05	9.5ml of sample, 0.5mL W3096	RM/IZ	OK
20	Q3365-05MSD	19 HARVEYMSD	MSD	10/16/25 16:27	9.5ml of sample, 0.5mL W3096	RM/IZ	OK
21	Q3365-01DL	MW25DL	SAM	10/16/25 17:01	100X For CI still high	RM/IZ	Dilution
22	CCV2	CCV2	CCV	10/16/25 17:23		RM/IZ	OK
23	CCB2	CCB2	CCB	10/16/25 17:45		RM/IZ	OK
24	Q3365-02DL	MW2DDL	SAM	10/16/25 19:11	100X For CI still high	RM/IZ	Dilution
25	Q3365-03DL	SUPPLYDL	SAM	10/16/25 19:54	100X For CI	RM/IZ	Confirms
26	Q3365-04DL	17 HARVEYDL	SAM	10/16/25 20:15	100X For CI	RM/IZ	Confirms
27	Q3365-05DL	19 HARVEYDL	SAM	10/16/25 20:37	100X For CI	RM/IZ	Confirms
28	CCV3	CCV3	CCV	10/16/25 20:58		RM/IZ	OK
29	CCB3	CCB3	CCB	10/16/25 21:20		RM/IZ	OK
30	CCV4	CCV4	CCV	10/17/25 10:15		RM/IZ	OK
31	CCB4	CCB4	CCB	10/17/25 10:36		RM/IZ	OK
32	LB137567BLW2	LB137567BLW2	MB	10/17/25 10:58		RM/IZ	OK
33	LB137567BSW2	LB137567BSW2	LCS	10/17/25 11:19		RM/IZ	OK
34	Q3365-01DL2	MW25DL2	SAM	10/17/25 11:41	1000X For CI	RM/IZ	Confirms
35	Q3365-02DL2	MW2DDL2	SAM	10/17/25 12:02	1000X For CI	RM/IZ	Confirms
36	CCV5	CCV5	CCV	10/17/25 12:24		RM/IZ	OK
37	CCB5	CCB5	CCB	10/17/25 12:45		RM/IZ	OK