

Duplicate Sample Summary

Client: G Environmental	SDG No.: Q3365
Project: Banker	Sample ID: Q3365-05
Client ID: 19 HARVEYMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-20	2.10		2.10		1	0		10/16/2025
Nitrate	mg/L	+/-20	4.10		4.10		1	0		10/16/2025
Nitrite	mg/L	+/-20	2.90		2.90		1	0		10/16/2025
Chloride	mg/L	+/-20	471	OR	472	OR	1	0		10/16/2025
Sulfate	mg/L	+/-20	43.9	OR	44.0	OR	1	0		10/16/2025
Bromide	mg/L	+/-20	10.0		9.90		1	1		10/16/2025
Orthophosphate as P	mg/L	+/-20	4.70		5.00		1	6		10/16/2025