

LAB CHRONICLE

OrderID:	Q3365	OrderDate:	10/16/2025 1:32:00 PM
Client:	G Environmental	Project:	Banker
Contact:	Gary Landis	Location:	J13

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3365-01	MW25	Water	Metals Group4	6010D	10/16/25	10/21/25	10/23/25	10/16/25
Q3365-02	MW2D	Water	Metals Group4	6010D	10/16/25	10/21/25	10/23/25	10/16/25
Q3365-03	SUPPLY	Water	Metals Group4	6010D	10/16/25	10/21/25	10/21/25	10/16/25
Q3365-04	17 HARVEY	Water	Metals Group4	6010D	10/16/25	10/21/25	10/21/25	10/16/25
Q3365-05	19 HARVEY	Water	Metals Group4	6010D	10/16/25	10/21/25	10/21/25	10/16/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3365 **Order ID:** Q3365
Client: G Environmental **Project ID:** Banker

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : MW25 Q3365-01	MW25	Water	Sodium	2570000	D	2170	5000	ug/L
Client ID : MW2D Q3365-02	MW2D	Water	Sodium	2000000	D	2170	5000	ug/L
Client ID : SUPPLY Q3365-03	SUPPLY	Water	Sodium	47400		434	1000	ug/L
Client ID : 17 HARVEY Q3365-04	17 HARVEY	Water	Sodium	126000		434	1000	ug/L
Client ID : 19 HARVEY Q3365-05	19 HARVEY	Water	Sodium	159000		434	1000	ug/L



SAMPLE DATA

Report of Analysis

Client: G Environmental
Project: Banker
Client Sample ID: MW25
Lab Sample ID: Q3365-01
Level (low/med): low

Date Collected: 10/16/25
Date Received: 10/16/25
SDG No.: Q3365
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	2570000	D	5	2170	5000	ug/L	10/21/25 16:10	10/23/25 14:28	6010D	

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: G Environmental
Project: Banker
Client Sample ID: MW2D
Lab Sample ID: Q3365-02
Level (low/med): low

Date Collected: 10/16/25
Date Received: 10/16/25
SDG No.: Q3365
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	2000000	D	5	2170	5000	ug/L	10/21/25 16:10	10/23/25 14:32	6010D	

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: G Environmental
Project: Banker
Client Sample ID: SUPPLY
Lab Sample ID: Q3365-03
Level (low/med): low

Date Collected: 10/16/25
Date Received: 10/16/25
SDG No.: Q3365
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	47400	1	434		1000	ug/L	10/21/25 16:10	10/21/25 17:53	6010D	

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: G Environmental
Project: Banker
Client Sample ID: 17 HARVEY
Lab Sample ID: Q3365-04
Level (low/med): low

Date Collected: 10/16/25
Date Received: 10/16/25
SDG No.: Q3365
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	126000	1	434		1000	ug/L	10/21/25 16:10	10/21/25 17:58	6010D	

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: G Environmental
Project: Banker
Client Sample ID: 19 HARVEY
Lab Sample ID: Q3365-05
Level (low/med): low

Date Collected: 10/16/25
Date Received: 10/16/25
SDG No.: Q3365
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	159000	1	434		1000	ug/L	10/21/25 16:10	10/21/25 18:02	6010D	

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	18300	20000	92	90 - 110	P	10/21/2025	12:40	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	1790	2000	90	80 - 120	P	10/21/2025	12:44	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

Contract: GENV01

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

SDG No.: Q3365

Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	23900	25000	96	90 - 110	P	10/21/2025	13:45	LB137613
CCV02	Sodium	24800	25000	99	90 - 110	P	10/21/2025	15:03	LB137613
CCV03	Sodium	25300	25000	101	90 - 110	P	10/21/2025	15:56	LB137613
CCV04	Sodium	24000	25000	96	90 - 110	P	10/21/2025	16:48	LB137613
CCV05	Sodium	23100	25000	92	90 - 110	P	10/21/2025	17:37	LB137613
CCV06	Sodium	24400	25000	98	90 - 110	P	10/21/2025	18:27	LB137613
CCV07	Sodium	22700	25000	91	90 - 110	P	10/21/2025	19:18	LB137613
CCV08	Sodium	23300	25000	93	90 - 110	P	10/21/2025	20:03	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	18900	20000	94	90 - 110	P	10/23/2025	13:05	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	1890	2000	95	80 - 120	P	10/23/2025	13:14	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental
Contract: GENV01
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

SDG No.: Q3365
Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	24300	25000	97	90 - 110	P	10/23/2025	13:52	LB137630
CCV02	Sodium	25400	25000	102	90 - 110	P	10/23/2025	15:04	LB137630
CCV03	Sodium	25100	25000	100	90 - 110	P	10/23/2025	15:56	LB137630
CCV04	Sodium	22800	25000	91	90 - 110	P	10/23/2025	16:43	LB137630
CCV05	Sodium	25100	25000	101	90 - 110	P	10/23/2025	17:21	LB137630
CCV06	Sodium	23800	25000	95	90 - 110	P	10/23/2025	18:11	LB137630
CCV07	Sodium	22700	25000	91	90 - 110	P	10/23/2025	18:59	LB137630
CCV08	Sodium	23100	25000	92	90 - 110	P	10/23/2025	19:48	LB137630
CCV09	Sodium	25700	25000	103	90 - 110	P	10/23/2025	20:39	LB137630
CCV10	Sodium	23700	25000	95	90 - 110	P	10/23/2025	21:30	LB137630
CCV11	Sodium	32800	25000	131	90 - 110	P	10/23/2025	22:20	LB137630
CCV12	Sodium	30000	25000	120	90 - 110	P	10/23/2025	22:33	LB137630



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Sodium	1870	2000	93	65 - 135	P	10/21/2025	12:54	LB137613
CRI01	Sodium	1820	2000	91	65 - 135	P	10/23/2025	13:24	LB137630



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Sodium	868	+/-1000	U	2000	P	10/21/2025	12:50	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Sodium	868	+/-1000	U	2000	P	10/21/2025	13:54	LB137613
CCB02	Sodium	868	+/-1000	U	2000	P	10/21/2025	15:07	LB137613
CCB03	Sodium	868	+/-1000	U	2000	P	10/21/2025	16:00	LB137613
CCB04	Sodium	868	+/-1000	U	2000	P	10/21/2025	16:52	LB137613
CCB05	Sodium	868	+/-1000	U	2000	P	10/21/2025	17:41	LB137613
CCB06	Sodium	1000	+/-1000	J	2000	P	10/21/2025	18:31	LB137613
CCB07	Sodium	868	+/-1000	U	2000	P	10/21/2025	19:22	LB137613
CCB08	Sodium	868	+/-1000	U	2000	P	10/21/2025	20:07	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Sodium	868	+/-1000	U	2000	P	10/23/2025	13:20	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Sodium	868	+/-1000	U	2000	P	10/23/2025	13:57	LB137630
CCB02	Sodium	868	+/-1000	U	2000	P	10/23/2025	15:08	LB137630
CCB03	Sodium	868	+/-1000	U	2000	P	10/23/2025	16:11	LB137630
CCB04	Sodium	868	+/-1000	U	2000	P	10/23/2025	16:47	LB137630
CCB05	Sodium	868	+/-1000	U	2000	P	10/23/2025	17:25	LB137630
CCB06	Sodium	868	+/-1000	U	2000	P	10/23/2025	18:15	LB137630
CCB07	Sodium	868	+/-1000	U	2000	P	10/23/2025	19:04	LB137630
CCB08	Sodium	868	+/-1000	U	2000	P	10/23/2025	19:52	LB137630
CCB09	Sodium	868	+/-1000	U	2000	P	10/23/2025	20:43	LB137630
CCB10	Sodium	1020	+/-1000	J*	2000	P	10/23/2025	21:34	LB137630
CCB11	Sodium	10400	+/-1000	*	2000	P	10/23/2025	22:24	LB137630
CCB12	Sodium	7480	+/-1000	*	2000	P	10/23/2025	22:37	LB137630

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3365

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170192BL		WATER		Batch Number:	PB170192		Prep Date:	10/21/2025	
	Sodium	434	<500	U	1000	P	10/21/2025	17:28	LB137613

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Sodium	51.5			0	0	10/21/2025	12:59	LB137613
ICSAB01	Sodium	95.7			0	0	10/21/2025	13:03	LB137613
ICSA01	Sodium	130			0	0	10/23/2025	13:28	LB137630
ICSAB01	Sodium	46.8			0	0	10/23/2025	13:39	LB137630



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client:	G Environmental	level:	low	sdg no.:	Q3365
contract:	GENV01			lab code:	ACE
matrix:	Water	sample id:	Q3370-01	client id:	WATER-TREATMENT DISCHARGEMS
Percent Solids for Sample:	NA	Spiked ID:	Q3370-01MS	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	102000		101000		1500	37		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q3365</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3370-01</u>	client id: <u>WATER-TREATMENT DISCHARGEMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3370-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	102000		101000		1500	18		P

Metals
- 5b -

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Matrix: _____

Level: LOW

Client ID: _____

Sample ID: _____

Spiked ID: _____

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

A
B
C
D
E
F
G
H

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3365</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3370-01</u>	Client ID:	<u>WATER-TREATMENT DISCHARGEDU</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3370-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Sodium	ug/L	20	101000		97600		3		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3365</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3370-01MS</u>	Client ID:	<u>WATER-TREATMENT DISCHARGEMS</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3370-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Sodium	ug/L	20	102000		102000		0		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170192BS Sodium	ug/L	1500	1400		93	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WATER-TREATMENT DISCHARGEL

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **Lb No.:** lb137613 **Lab Sample ID :** Q3370-01L **SDG No.:** Q3365
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Sodium	101000	102000	1		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

METAL
PREPARATION &
ANALYICAL
SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: G Environmental

SDG No.: Q3365

Contract: GENV01

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB170192							
PB170192BL	PB170192BL	MB	WATER	10/21/2025	50.0	25.0	
PB170192BS	PB170192BS	LCS	WATER	10/21/2025	50.0	25.0	
Q3365-01	MW25	SAM	WATER	10/21/2025	50.0	25.0	
Q3365-02	MW2D	SAM	WATER	10/21/2025	50.0	25.0	
Q3365-03	SUPPLY	SAM	WATER	10/21/2025	50.0	25.0	
Q3365-04	17 HARVEY	SAM	WATER	10/21/2025	50.0	25.0	
Q3365-05	19 HARVEY	SAM	WATER	10/21/2025	50.0	25.0	
Q3370-01DUP	WATER-TREATMENT DISCHARGEDUP	DUP	WATER	10/21/2025	50.0	25.0	
Q3370-01MS	WATER-TREATMENT DISCHARGEMS	MS	WATER	10/21/2025	50.0	25.0	
Q3370-01MSD	WATER-TREATMENT DISCHARGEMSD	MSD	WATER	10/21/2025	50.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: G Environmental

Contract: GENV01

Lab code: ACE

Sdg no.: Q3365

Instrument id number: _____

Method: _____

Run number: LB137613

Start date: 10/21/2025

End date: 10/21/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1203	Na
S1	S1	1	1208	Na
S2	S2	1	1212	Na
S3	S3	1	1216	Na
S4	S4	1	1221	Na
S5	S5	1	1225	Na
ICV01	ICV01	1	1240	Na
LLICV01	LLICV01	1	1244	Na
ICB01	ICB01	1	1250	Na
CRI01	CRI01	1	1254	Na
ICSA01	ICSA01	1	1259	Na
ICSAB01	ICSAB01	1	1303	Na
ICSA	ICSA	20	1307	Na
ICSAB	ICSAB	20	1311	Na
CCV01	CCV01	1	1345	Na
CCB01	CCB01	1	1354	Na
CCV02	CCV02	1	1503	Na
CCB02	CCB02	1	1507	Na
CCV03	CCV03	1	1556	Na
CCB03	CCB03	1	1600	Na
CCV04	CCV04	1	1648	Na
CCB04	CCB04	1	1652	Na
PB170192BL	PB170192BL	1	1728	Na
PB170192BS	PB170192BS	1	1733	Na
CCV05	CCV05	1	1737	Na
CCB05	CCB05	1	1741	Na
Q3365-03	SUPPLY	1	1753	Na
Q3365-04	17 HARVEY	1	1758	Na
Q3365-05	19 HARVEY	1	1802	Na
Q3370-01DUP	WATER-TREATMENT DISCH#	1	1811	Na
Q3370-01L	WATER-TREATMENT DISCH#	5	1815	Na
Q3370-01MS	WATER-TREATMENT DISCH#	1	1819	Na
Q3370-01MSD	WATER-TREATMENT DISCH#	1	1823	Na
CCV06	CCV06	1	1827	Na
CCB06	CCB06	1	1831	Na
CCV07	CCV07	1	1918	Na
CCB07	CCB07	1	1922	Na
CCV08	CCV08	1	2003	Na
CCB08	CCB08	1	2007	Na

metals
- 14 -
ANALYSIS RUN LOG

Client: G Environmental

Contract: GENV01

Lab code: ACE

Sdg no.: Q3365

Instrument id number: _____

Method: _____

Run number: LB137630

Start date: 10/23/2025

End date: 10/23/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1240	Na
S1	S1	1	1245	Na
S2	S2	1	1249	Na
S3	S3	1	1253	Na
S4	S4	1	1257	Na
S5	S5	1	1301	Na
ICV01	ICV01	1	1305	Na
LLICV01	LLICV01	1	1314	Na
ICB01	ICB01	1	1320	Na
CRI01	CRI01	1	1324	Na
ICSA01	ICSA01	1	1328	Na
ICSAB01	ICSAB01	1	1339	Na
ICSA	ICSA01	20	1343	Na
ICSAB	ICSAB01	20	1347	Na
CCV01	CCV01	1	1352	Na
CCB01	CCB01	1	1357	Na
Q3365-01	MW25	5	1428	Na
Q3365-02	MW2D	5	1432	Na
CCV02	CCV02	1	1504	Na
CCB02	CCB02	1	1508	Na
CCV03	CCV03	1	1556	Na
CCB03	CCB03	1	1611	Na
CCV04	CCV04	1	1643	Na
CCB04	CCB04	1	1647	Na
CCV05	CCV05	1	1721	Na
CCB05	CCB05	1	1725	Na
CCV06	CCV06	1	1811	Na
CCB06	CCB06	1	1815	Na
CCV07	CCV07	1	1859	Na
CCB07	CCB07	1	1904	Na
CCV08	CCV08	1	1948	Na
CCB08	CCB08	1	1952	Na
CCV09	CCV09	1	2039	Na
CCB09	CCB09	1	2043	Na
CCV10	CCV10	1	2130	Na
CCB10	CCB10	1	2134	Na
CCV11	CCV11	1	2220	Na
CCB11	CCB11	1	2224	Na
CCV12	CCV12	1	2233	Na
CCB12	CCB12	1	2237	Na