

Initial and Continuing Calibration Verification

Client: Aramark Uniforms

SDG No.: Q3368

Project: Monthly 2025

RunNo.: LB137602

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.093	0.099	94	85-115	10/21/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	10/21/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	10/21/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	10/21/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	10/21/2025

Initial and Continuing Calibration Verification

Client: Aramark Uniforms

SDG No.: Q3368

Project: Monthly 2025

RunNo.: LB137603

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Corrosivity		pH	6.98	7	100	90-110	10/21/2025
Sample ID:	CCV1						
Corrosivity		pH	2.01	2.00	101	90-110	10/21/2025
Sample ID:	CCV2						
Corrosivity		pH	12.02	12.00	100	90-110	10/21/2025
Sample ID:	CCV3						
Corrosivity		pH	2.01	2.00	101	90-110	10/21/2025

Initial and Continuing Calibration Verification

Client: Aramark Uniforms

SDG No.: Q3368

Project: Monthly 2025

RunNo.: LB137698

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	0.98	1	98	90-110	10/29/2025
Sample ID: CCV1 ASTM Ammonia	mg/L	0.96	1	96	90-110	10/29/2025
Sample ID: CCV2 ASTM Ammonia	mg/L	1	1	100	90-110	10/29/2025

Initial and Continuing Calibration Verification

Client: Aramark Uniforms

SDG No.: Q3368

Project: Monthly 2025

RunNo.: LB137721

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV ASTM COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 ASTM COD	mg/L	48.150	50	96	95-105	10/30/2025
Sample ID: CCV2 ASTM COD	mg/L	50.173	50	100	95-105	10/30/2025

Initial and Continuing Calibration Verification

Client: Aramark Uniforms

SDG No.: Q3368

Project: Monthly 2025

RunNo.: LB137757

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	11/04/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	11/04/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	11/04/2025