

### Initial and Continuing Calibration Blank Summary

**Client:** Aramark Uniforms

**SDG No.:** Q3368

**Project:** Monthly 2025

**RunNo.:** LB137602

| Analyte                                    | Units | Result   | Acceptance Limits | Conc Qual | MDL     | RDL   | Analysis Date |
|--------------------------------------------|-------|----------|-------------------|-----------|---------|-------|---------------|
| Sample ID: <b>ICB1</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 10/21/2025    |
| Sample ID: <b>CCB1</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 10/21/2025    |
| Sample ID: <b>CCB2</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 10/21/2025    |
| Sample ID: <b>CCB3</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 10/21/2025    |
| Sample ID: <b>CCB4</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 10/21/2025    |

### Initial and Continuing Calibration Blank Summary

**Client:** Aramark Uniforms

**SDG No.:** Q3368

**Project:** Monthly 2025

**RunNo.:** LB137698

| Analyte                                | Units | Result   | Acceptance<br>Limits | Conc<br>Qual | MDL   | RDL | Analysis<br>Date |
|----------------------------------------|-------|----------|----------------------|--------------|-------|-----|------------------|
| Sample ID: <b>ICB1</b><br>ASTM Ammonia | mg/L  | < 0.0500 | 0.0500               | U            | 0.030 | 0.1 | 10/29/2025       |
| Sample ID: <b>CCB1</b><br>ASTM Ammonia | mg/L  | < 0.0500 | 0.0500               | U            | 0.030 | 0.1 | 10/29/2025       |
| Sample ID: <b>CCB2</b><br>ASTM Ammonia | mg/L  | 0.034    | 0.0500               | J            | 0.030 | 0.1 | 10/29/2025       |

### Initial and Continuing Calibration Blank Summary

**Client:** Aramark Uniforms

**SDG No.:** Q3368

**Project:** Monthly 2025

**RunNo.:** LB137721

| Analyte                            | Units | Result   | Acceptance Limits | Conc Qual | MDL  | RDL | Analysis Date |
|------------------------------------|-------|----------|-------------------|-----------|------|-----|---------------|
| Sample ID: <b>ICB</b><br>ASTM COD  | mg/L  | < 5.0000 | 5.0000            | U         | 1.50 | 10  | 09/08/2025    |
| Sample ID: <b>CCB1</b><br>ASTM COD | mg/L  | < 5.0000 | 5.0000            | U         | 1.50 | 10  | 10/30/2025    |
| Sample ID: <b>CCB2</b><br>ASTM COD | mg/L  | < 5.0000 | 5.0000            | U         | 1.50 | 10  | 10/30/2025    |