

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3368

Project: Monthly 2025

Run No.: LB137692

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137692BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	10/29/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3368

Project: Monthly 2025

Run No.: LB137695

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137695BS							
ASTM Oil and Grease	mg/L	20.0	16.7		84	1	78-114	10/29/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3368

Project: Monthly 2025

Run No.: LB137721

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137721BS							
ASTM COD	mg/L	50	51.2		102	1	90-110	10/30/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3368

Project: Monthly 2025

Run No.: LB137698

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170314BS							
ASTM Ammonia	mg/L	1	1.00		100	1	90-110	10/29/2025