

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV84	Mercury	3.99	4.0	100	90 - 110	CV	10/23/2025	08:15	LB137620

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV75	Mercury	5.03	5.0	101	90 - 110	CV	10/23/2025	08:22	LB137620
CCV76	Mercury	5.05	5.0	101	90 - 110	CV	10/23/2025	08:50	LB137620
CCV77	Mercury	4.90	5.0	98	90 - 110	CV	10/23/2025	09:18	LB137620
CCV78	Mercury	5.05	5.0	101	90 - 110	CV	10/23/2025	09:36	LB137620

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3870	4000	97	90 - 110	P	10/23/2025	13:05	LB137630
	Barium	7690	8000	96	90 - 110	P	10/23/2025	13:05	LB137630
	Cadmium	2050	2000	103	90 - 110	P	10/23/2025	13:05	LB137630
	Chromium	800	800	100	90 - 110	P	10/23/2025	13:05	LB137630
	Copper	995	1000	100	90 - 110	P	10/23/2025	13:05	LB137630
	Lead	3900	4000	98	90 - 110	P	10/23/2025	13:05	LB137630
	Nickel	1940	2000	97	90 - 110	P	10/23/2025	13:05	LB137630
	Selenium	3910	4000	98	90 - 110	P	10/23/2025	13:05	LB137630
	Silver	1010	1000	101	90 - 110	P	10/23/2025	13:05	LB137630
	Zinc	1970	2000	99	90 - 110	P	10/23/2025	13:05	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	22.1	20.0	110	80 - 120	P	10/23/2025	13:14	LB137630
	Barium	98.7	100	99	80 - 120	P	10/23/2025	13:14	LB137630
	Cadmium	5.95	6.0	99	80 - 120	P	10/23/2025	13:14	LB137630
	Chromium	11.0	10.0	110	80 - 120	P	10/23/2025	13:14	LB137630
	Copper	22.5	20.0	113	80 - 120	P	10/23/2025	13:14	LB137630
	Lead	13.2	12.0	110	80 - 120	P	10/23/2025	13:14	LB137630
	Nickel	39.8	40.0	100	80 - 120	P	10/23/2025	13:14	LB137630
	Selenium	23.6	20.0	118	80 - 120	P	10/23/2025	13:14	LB137630
	Silver	10.7	10.0	106	80 - 120	P	10/23/2025	13:14	LB137630
	Zinc	43.5	40.0	109	80 - 120	P	10/23/2025	13:14	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5250	5000	105	90 - 110	P	10/23/2025	13:52	LB137630
	Barium	9900	10000	99	90 - 110	P	10/23/2025	13:52	LB137630
	Cadmium	2580	2500	103	90 - 110	P	10/23/2025	13:52	LB137630
	Chromium	1060	1000	106	90 - 110	P	10/23/2025	13:52	LB137630
	Copper	1300	1250	104	90 - 110	P	10/23/2025	13:52	LB137630
	Lead	5160	5000	103	90 - 110	P	10/23/2025	13:52	LB137630
	Nickel	2580	2500	103	90 - 110	P	10/23/2025	13:52	LB137630
	Selenium	5330	5000	107	90 - 110	P	10/23/2025	13:52	LB137630
	Silver	1300	1250	104	90 - 110	P	10/23/2025	13:52	LB137630
	Zinc	2620	2500	105	90 - 110	P	10/23/2025	13:52	LB137630
CCV02	Arsenic	5240	5000	105	90 - 110	P	10/23/2025	15:04	LB137630
	Barium	10600	10000	106	90 - 110	P	10/23/2025	15:04	LB137630
	Cadmium	2630	2500	105	90 - 110	P	10/23/2025	15:04	LB137630
	Chromium	1050	1000	105	90 - 110	P	10/23/2025	15:04	LB137630
	Copper	1310	1250	105	90 - 110	P	10/23/2025	15:04	LB137630
	Lead	5230	5000	105	90 - 110	P	10/23/2025	15:04	LB137630
	Nickel	2610	2500	104	90 - 110	P	10/23/2025	15:04	LB137630
	Selenium	5210	5000	104	90 - 110	P	10/23/2025	15:04	LB137630
	Silver	1300	1250	104	90 - 110	P	10/23/2025	15:04	LB137630
	Zinc	2610	2500	104	90 - 110	P	10/23/2025	15:04	LB137630
CCV03	Arsenic	4960	5000	99	90 - 110	P	10/23/2025	15:56	LB137630
	Barium	9590	10000	96	90 - 110	P	10/23/2025	15:56	LB137630
	Cadmium	2520	2500	101	90 - 110	P	10/23/2025	15:56	LB137630
	Chromium	1030	1000	103	90 - 110	P	10/23/2025	15:56	LB137630
	Copper	1240	1250	99	90 - 110	P	10/23/2025	15:56	LB137630
	Lead	5020	5000	100	90 - 110	P	10/23/2025	15:56	LB137630
	Nickel	2490	2500	100	90 - 110	P	10/23/2025	15:56	LB137630
	Selenium	4960	5000	99	90 - 110	P	10/23/2025	15:56	LB137630
	Silver	1230	1250	99	90 - 110	P	10/23/2025	15:56	LB137630
	Zinc	2480	2500	99	90 - 110	P	10/23/2025	15:56	LB137630
CCV04	Arsenic	4960	5000	99	90 - 110	P	10/23/2025	16:43	LB137630
	Barium	9850	10000	98	90 - 110	P	10/23/2025	16:43	LB137630
	Cadmium	2520	2500	101	90 - 110	P	10/23/2025	16:43	LB137630
	Chromium	1000	1000	100	90 - 110	P	10/23/2025	16:43	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Copper	1240	1250	99	90 - 110	P	10/23/2025	16:43	LB137630
	Lead	5020	5000	100	90 - 110	P	10/23/2025	16:43	LB137630
	Nickel	2490	2500	100	90 - 110	P	10/23/2025	16:43	LB137630
	Selenium	4920	5000	98	90 - 110	P	10/23/2025	16:43	LB137630
	Silver	1220	1250	98	90 - 110	P	10/23/2025	16:43	LB137630
	Zinc	2450	2500	98	90 - 110	P	10/23/2025	16:43	LB137630
CCV05	Arsenic	5210	5000	104	90 - 110	P	10/23/2025	17:21	LB137630
	Barium	10300	10000	103	90 - 110	P	10/23/2025	17:21	LB137630
	Cadmium	2640	2500	105	90 - 110	P	10/23/2025	17:21	LB137630
	Chromium	967	1000	97	90 - 110	P	10/23/2025	17:21	LB137630
	Copper	1300	1250	104	90 - 110	P	10/23/2025	17:21	LB137630
	Lead	5250	5000	105	90 - 110	P	10/23/2025	17:21	LB137630
	Nickel	2610	2500	104	90 - 110	P	10/23/2025	17:21	LB137630
	Selenium	5180	5000	104	90 - 110	P	10/23/2025	17:21	LB137630
	Silver	1290	1250	103	90 - 110	P	10/23/2025	17:21	LB137630
	Zinc	2600	2500	104	90 - 110	P	10/23/2025	17:21	LB137630
CCV06	Arsenic	5210	5000	104	90 - 110	P	10/23/2025	18:11	LB137630
	Barium	10000	10000	100	90 - 110	P	10/23/2025	18:11	LB137630
	Cadmium	2670	2500	107	90 - 110	P	10/23/2025	18:11	LB137630
	Chromium	987	1000	99	90 - 110	P	10/23/2025	18:11	LB137630
	Copper	1310	1250	105	90 - 110	P	10/23/2025	18:11	LB137630
	Lead	5320	5000	106	90 - 110	P	10/23/2025	18:11	LB137630
	Nickel	2630	2500	105	90 - 110	P	10/23/2025	18:11	LB137630
	Selenium	5190	5000	104	90 - 110	P	10/23/2025	18:11	LB137630
	Silver	1320	1250	105	90 - 110	P	10/23/2025	18:11	LB137630
	Zinc	2630	2500	105	90 - 110	P	10/23/2025	18:11	LB137630
CCV07	Arsenic	5030	5000	101	90 - 110	P	10/23/2025	18:59	LB137630
	Barium	9590	10000	96	90 - 110	P	10/23/2025	18:59	LB137630
	Cadmium	2530	2500	101	90 - 110	P	10/23/2025	18:59	LB137630
	Chromium	952	1000	95	90 - 110	P	10/23/2025	18:59	LB137630
	Copper	1250	1250	100	90 - 110	P	10/23/2025	18:59	LB137630
	Lead	5060	5000	101	90 - 110	P	10/23/2025	18:59	LB137630
	Nickel	2490	2500	100	90 - 110	P	10/23/2025	18:59	LB137630
	Selenium	5090	5000	102	90 - 110	P	10/23/2025	18:59	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Silver	1280	1250	102	90 - 110	P	10/23/2025	18:59	LB137630
	Zinc	2550	2500	102	90 - 110	P	10/23/2025	18:59	LB137630
CCV08	Arsenic	5200	5000	104	90 - 110	P	10/23/2025	19:48	LB137630
	Barium	9920	10000	99	90 - 110	P	10/23/2025	19:48	LB137630
	Cadmium	2670	2500	107	90 - 110	P	10/23/2025	19:48	LB137630
	Chromium	999	1000	100	90 - 110	P	10/23/2025	19:48	LB137630
	Copper	1300	1250	104	90 - 110	P	10/23/2025	19:48	LB137630
	Lead	5320	5000	106	90 - 110	P	10/23/2025	19:48	LB137630
	Nickel	2620	2500	105	90 - 110	P	10/23/2025	19:48	LB137630
	Selenium	5190	5000	104	90 - 110	P	10/23/2025	19:48	LB137630
	Silver	1300	1250	104	90 - 110	P	10/23/2025	19:48	LB137630
	Zinc	2610	2500	104	90 - 110	P	10/23/2025	19:48	LB137630
CCV09	Arsenic	5320	5000	106	90 - 110	P	10/23/2025	20:39	LB137630
	Barium	9990	10000	100	90 - 110	P	10/23/2025	20:39	LB137630
	Cadmium	2690	2500	108	90 - 110	P	10/23/2025	20:39	LB137630
	Chromium	1000	1000	100	90 - 110	P	10/23/2025	20:39	LB137630
	Copper	1320	1250	105	90 - 110	P	10/23/2025	20:39	LB137630
	Lead	5380	5000	108	90 - 110	P	10/23/2025	20:39	LB137630
	Nickel	2650	2500	106	90 - 110	P	10/23/2025	20:39	LB137630
	Selenium	5360	5000	107	90 - 110	P	10/23/2025	20:39	LB137630
	Silver	1320	1250	106	90 - 110	P	10/23/2025	20:39	LB137630
	Zinc	2650	2500	106	90 - 110	P	10/23/2025	20:39	LB137630
CCV10	Arsenic	5010	5000	100	90 - 110	P	10/23/2025	21:30	LB137630
	Barium	9490	10000	95	90 - 110	P	10/23/2025	21:30	LB137630
	Cadmium	2570	2500	103	90 - 110	P	10/23/2025	21:30	LB137630
	Chromium	962	1000	96	90 - 110	P	10/23/2025	21:30	LB137630
	Copper	1250	1250	100	90 - 110	P	10/23/2025	21:30	LB137630
	Lead	5110	5000	102	90 - 110	P	10/23/2025	21:30	LB137630
	Nickel	2510	2500	100	90 - 110	P	10/23/2025	21:30	LB137630
	Selenium	5040	5000	101	90 - 110	P	10/23/2025	21:30	LB137630
	Silver	1280	1250	102	90 - 110	P	10/23/2025	21:30	LB137630
	Zinc	2530	2500	101	90 - 110	P	10/23/2025	21:30	LB137630
CCV11	Arsenic	4980	5000	100	90 - 110	P	10/23/2025	22:20	LB137630
	Barium	9350	10000	94	90 - 110	P	10/23/2025	22:20	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Cadmium	2590	2500	104	90 - 110	P	10/23/2025	22:20	LB137630
	Chromium	951	1000	95	90 - 110	P	10/23/2025	22:20	LB137630
	Copper	1250	1250	100	90 - 110	P	10/23/2025	22:20	LB137630
	Lead	5160	5000	103	90 - 110	P	10/23/2025	22:20	LB137630
	Nickel	2530	2500	101	90 - 110	P	10/23/2025	22:20	LB137630
	Selenium	4940	5000	99	90 - 110	P	10/23/2025	22:20	LB137630
	Silver	1230	1250	99	90 - 110	P	10/23/2025	22:20	LB137630
	Zinc	2450	2500	98	90 - 110	P	10/23/2025	22:20	LB137630
CCV12	Arsenic	4920	5000	98	90 - 110	P	10/23/2025	22:33	LB137630
	Barium	9070	10000	91	90 - 110	P	10/23/2025	22:33	LB137630
	Cadmium	2540	2500	102	90 - 110	P	10/23/2025	22:33	LB137630
	Chromium	932	1000	93	90 - 110	P	10/23/2025	22:33	LB137630
	Copper	1230	1250	98	90 - 110	P	10/23/2025	22:33	LB137630
	Lead	5070	5000	101	90 - 110	P	10/23/2025	22:33	LB137630
	Nickel	2470	2500	99	90 - 110	P	10/23/2025	22:33	LB137630
	Selenium	4930	5000	99	90 - 110	P	10/23/2025	22:33	LB137630
	Silver	1220	1250	98	90 - 110	P	10/23/2025	22:33	LB137630
	Zinc	2410	2500	96	90 - 110	P	10/23/2025	22:33	LB137630