



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Metals

- 3a -

#### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| ICB84     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 10/23/2025       | 08:18            | LB137620      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Aramark Uniforms

**SDG No.:** Q3368

**Contract:** ARAM01

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| CCB75     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 10/23/2025       | 08:24            | LB137620      |
| CCB76     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 10/23/2025       | 08:52            | LB137620      |
| CCB77     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 10/23/2025       | 09:20            | LB137620      |
| CCB78     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 10/23/2025       | 09:38            | LB137620      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Aramark Uniforms

**SDG No.:** Q3368

**Contract:** ARAM01

**Lab Code:** ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 13:20            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 13:20            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 13:20            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 13:20            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 13:20            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 13:20            | LB137630      |
|           | Nickel   | 6.19           | +/-20               | J            | 40.0 | P | 10/23/2025       | 13:20            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 13:20            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 13:20            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 13:20            | LB137630      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 13:57            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 13:57            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 13:57            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 13:57            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 13:57            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 13:57            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 13:57            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 13:57            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 13:57            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 13:57            | LB137630      |
| CCB02     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 15:08            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 15:08            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 15:08            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 15:08            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 15:08            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 15:08            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 15:08            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 15:08            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 15:08            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 15:08            | LB137630      |
| CCB03     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 16:11            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 16:11            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 16:11            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 16:11            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 16:11            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 16:11            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 16:11            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 16:11            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 16:11            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 16:11            | LB137630      |
| CCB04     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 16:47            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 16:47            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 16:47            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 16:47            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 16:47            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 16:47            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 16:47            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 16:47            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 16:47            | LB137630      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB04     | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 16:47            | LB137630      |
| CCB05     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 17:25            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 17:25            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 17:25            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 17:25            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 17:25            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 17:25            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 17:25            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 17:25            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 17:25            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 17:25            | LB137630      |
| CCB06     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 18:15            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 18:15            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 18:15            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 18:15            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 18:15            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 18:15            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 18:15            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 18:15            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 18:15            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 18:15            | LB137630      |
| CCB07     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 19:04            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 19:04            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 19:04            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 19:04            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 19:04            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 19:04            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 19:04            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 19:04            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 19:04            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 19:04            | LB137630      |
| CCB08     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 19:52            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 19:52            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 19:52            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 19:52            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 19:52            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 19:52            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 19:52            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 19:52            | LB137630      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB08     | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 19:52            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 19:52            | LB137630      |
| CCB09     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 20:43            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 20:43            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 20:43            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 20:43            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 20:43            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 20:43            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 20:43            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 20:43            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 20:43            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 20:43            | LB137630      |
| CCB10     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 21:34            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 21:34            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 21:34            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 21:34            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 21:34            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 21:34            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 21:34            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 21:34            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 21:34            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 21:34            | LB137630      |
| CCB11     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 22:24            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 22:24            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 22:24            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 22:24            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 22:24            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 22:24            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 22:24            | LB137630      |
|           | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 22:24            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 22:24            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 22:24            | LB137630      |
| CCB12     | Arsenic  | 5.12           | +/-10               | U            | 20.0 | P | 10/23/2025       | 22:37            | LB137630      |
|           | Barium   | 14.6           | +/-50               | U            | 100  | P | 10/23/2025       | 22:37            | LB137630      |
|           | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 10/23/2025       | 22:37            | LB137630      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 10/23/2025       | 22:37            | LB137630      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 10/23/2025       | 22:37            | LB137630      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 10/23/2025       | 22:37            | LB137630      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 10/23/2025       | 22:37            | LB137630      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Aramark Uniforms

**SDG No.:** Q3368

**Contract:** ARAM01

**Lab Code:** ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB12     | Selenium | 9.64           | +/-10               | U            | 20.0 | P | 10/23/2025       | 22:37            | LB137630      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 10/23/2025       | 22:37            | LB137630      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 10/23/2025       | 22:37            | LB137630      |