

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Aramark Uniforms

SDG No.: Q3368

Contract: ARAM01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	8.83			-20	20	10/23/2025	13:28	LB137630
	Barium	2.45	6.0	41	-94	106	10/23/2025	13:28	LB137630
	Cadmium	-0.071	1.0	7	-5	7	10/23/2025	13:28	LB137630
	Chromium	48.5	52.0	93	42	62	10/23/2025	13:28	LB137630
	Copper	12.3	2.0	615	-18	22	10/23/2025	13:28	LB137630
	Lead	-2.09			-12	12	10/23/2025	13:28	LB137630
	Nickel	3.96	2.0	198	-38	42	10/23/2025	13:28	LB137630
	Selenium	4.51			-20	20	10/23/2025	13:28	LB137630
	Silver	-2.45			-10	10	10/23/2025	13:28	LB137630
	Zinc	3.55			-40	40	10/23/2025	13:28	LB137630
ICSAB01	Arsenic	107	104	103	88.4	120	10/23/2025	13:39	LB137630
	Barium	483	537	90	437	637	10/23/2025	13:39	LB137630
	Cadmium	994	972	102	826	1120	10/23/2025	13:39	LB137630
	Chromium	550	542	102	460	624	10/23/2025	13:39	LB137630
	Copper	485	511	95	434	588	10/23/2025	13:39	LB137630
	Lead	46.0	49.0	94	37	61	10/23/2025	13:39	LB137630
	Nickel	993	954	104	810	1100	10/23/2025	13:39	LB137630
	Selenium	55.7	46.0	121	26	66	10/23/2025	13:39	LB137630
	Silver	206	201	102	170	232	10/23/2025	13:39	LB137630
	Zinc	1030	952	108	809	1095	10/23/2025	13:39	LB137630