

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q3368</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3412-04</u>	client id: <u>TP-6MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3412-04MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	4030		100	U	4000	101		P
Barium	ug/L	75 - 125	2020		515		1000	151	N	P
Cadmium	ug/L	75 - 125	1000		30.0	U	1000	100		P
Chromium	ug/L	75 - 125	1880		50.0	U	2000	94		P
Copper	ug/L	75 - 125	1530		29.1	J	1500	100		P
Lead	ug/L	75 - 125	4870		67.3		5000	96		P
Mercury	ug/L	75 - 125	38.2		2.00	U	40.0	96		CV
Nickel	ug/L	75 - 125	2500		200	U	2500	100		P
Selenium	ug/L	75 - 125	10000		100	U	10000	100		P
Silver	ug/L	75 - 125	346		50.0	U	380	91		P
Zinc	ug/L	75 - 125	1280		124	J	1000	116		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q3368</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3412-04</u>	client id: <u>TP-6MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3412-04MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	3940		100	U	4000	98		P
Barium	ug/L	75 - 125	1950		515		1000	144	N	P
Cadmium	ug/L	75 - 125	974		30.0	U	1000	97		P
Chromium	ug/L	75 - 125	1830		50.0	U	2000	92		P
Copper	ug/L	75 - 125	1480		29.1	J	1500	97		P
Lead	ug/L	75 - 125	4720		67.3		5000	93		P
Mercury	ug/L	75 - 125	38.1		2.00	U	40.0	95		CV
Nickel	ug/L	75 - 125	2420		200	U	2500	97		P
Selenium	ug/L	75 - 125	9890		100	U	10000	99		P
Silver	ug/L	75 - 125	340		50.0	U	380	90		P
Zinc	ug/L	75 - 125	1250		124	J	1000	113		P