

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ARAM01

Lab Code: ACE

SDG NO.: Q3368

Continuing Calib Date: 10/21/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 09:42

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.50	5.49	5.39	5.59	-0.01
Aroclor-1242-2 (2)	5.52	5.52	5.42	5.62	0.00
Aroclor-1242-3 (3)	5.58	5.58	5.48	5.68	0.00
Aroclor-1242-4 (4)	5.68	5.67	5.57	5.77	-0.01
Aroclor-1242-5 (5)	6.40	6.40	6.30	6.50	0.00
Tetrachloro-m-xylene	4.36	4.36	4.26	4.46	0.00
Decachlorobiphenyl	9.95	9.93	9.83	10.03	-0.02

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ARAM01

Lab Code: ACE

SDG NO.: Q3368

Continuing Calib Date: 10/21/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 09:42

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.74	4.74	4.64	4.84	0.00
Aroclor-1242-2 (2)	4.76	4.76	4.66	4.86	0.00
Aroclor-1242-3 (3)	4.93	4.94	4.84	5.04	0.01
Aroclor-1242-4 (4)	5.02	5.02	4.92	5.12	0.00
Aroclor-1242-5 (5)	5.54	5.54	5.44	5.64	0.00
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.66	8.67	8.57	8.77	0.01