

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ARAM01

Lab Code: ACE

SDG NO.: Q3368

Continuing Calib Date: 10/21/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 10:18

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.34       | 6.33      | 6.23      | 6.43  | -0.01      |
| Aroclor-1254-2       | (2) | 6.55       | 6.55      | 6.45      | 6.65  | 0.00       |
| Aroclor-1254-3       | (3) | 6.91       | 6.91      | 6.81      | 7.01  | 0.00       |
| Aroclor-1254-4       | (4) | 7.19       | 7.19      | 7.09      | 7.29  | 0.00       |
| Aroclor-1254-5       | (5) | 7.61       | 7.61      | 7.51      | 7.71  | 0.00       |
| Tetrachloro-m-xylene |     | 4.36       | 4.36      | 4.26      | 4.46  | 0.00       |
| Decachlorobiphenyl   |     | 9.94       | 9.93      | 9.83      | 10.03 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ARAM01

Lab Code: ACE

SDG NO.: Q3368

Continuing Calib Date: 10/21/2025

Initial Calibration Date(s): 10/09/2025 10/09/2025

Continuing Calib Time: 10:18

Initial Calibration Time(s): 14:14 22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.54       | 5.54      | 5.44      | 5.64 | 0.00       |
| Aroclor-1254-2 (2)   | 5.69       | 5.69      | 5.59      | 5.79 | 0.01       |
| Aroclor-1254-3 (3)   | 6.09       | 6.09      | 5.99      | 6.19 | 0.00       |
| Aroclor-1254-4 (4)   | 6.32       | 6.32      | 6.22      | 6.42 | 0.00       |
| Aroclor-1254-5 (5)   | 6.73       | 6.74      | 6.64      | 6.84 | 0.01       |
| Tetrachloro-m-xylene | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   | 8.66       | 8.67      | 8.57      | 8.77 | 0.01       |