

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114542.D **Time Analyzed:** 14:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.494	5.392	5.592	531.160	500.000	6.2
Aroclor-1016-2	5.516	5.414	5.614	491.630	500.000	-1.7
Aroclor-1016-3	5.578	5.476	5.676	502.920	500.000	0.6
Aroclor-1016-4	5.673	5.571	5.771	517.460	500.000	3.5
Aroclor-1016-5	5.965	5.863	6.063	506.450	500.000	1.3
Aroclor-1260-1	7.078	6.975	7.175	498.620	500.000	-0.3
Aroclor-1260-2	7.332	7.229	7.429	496.090	500.000	-0.8
Aroclor-1260-3	7.689	7.585	7.785	451.450	500.000	-9.7
Aroclor-1260-4	7.911	7.807	8.007	489.140	500.000	-2.2
Aroclor-1260-5	8.218	8.113	8.313	484.870	500.000	-3.0
Decachlorobiphenyl	9.941	9.834	10.034	49.470	50.000	-1.1
Tetrachloro-m-xylene	4.355	4.255	4.455	52.640	50.000	5.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114542.D **Time Analyzed:** 14:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.643	4.843	496.120	500.000	-0.8
Aroclor-1016-2	4.757	4.662	4.862	496.300	500.000	-0.7
Aroclor-1016-3	4.932	4.836	5.036	517.580	500.000	3.5
Aroclor-1016-4	4.974	4.879	5.079	535.770	500.000	7.2
Aroclor-1016-5	5.186	5.092	5.292	502.200	500.000	0.4
Aroclor-1260-1	6.218	6.125	6.325	552.360	500.000	10.5
Aroclor-1260-2	6.405	6.313	6.513	583.740	500.000	16.7
Aroclor-1260-3	6.558	6.466	6.666	533.250	500.000	6.7
Aroclor-1260-4	7.030	6.939	7.139	577.390	500.000	15.5
Aroclor-1260-5	7.271	7.179	7.379	561.760	500.000	12.4
Decachlorobiphenyl	8.658	8.571	8.771	48.340	50.000	-3.3
Tetrachloro-m-xylene	3.657	3.560	3.760	49.070	50.000	-1.9