

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071029.D **Time Analyzed:** 11:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	528.530	500.000	5.7
Aroclor-1016-2	4.499	4.399	4.599	530.130	500.000	6.0
Aroclor-1016-3	4.557	4.456	4.656	543.840	500.000	8.8
Aroclor-1016-4	4.646	4.546	4.746	538.840	500.000	7.8
Aroclor-1016-5	4.931	4.831	5.031	538.570	500.000	7.7
Aroclor-1260-1	6.021	5.919	6.119	525.500	500.000	5.1
Aroclor-1260-2	6.277	6.175	6.375	523.040	500.000	4.6
Aroclor-1260-3	6.628	6.526	6.726	550.320	500.000	10.1
Aroclor-1260-4	6.841	6.740	6.940	559.640	500.000	11.9
Aroclor-1260-5	7.146	7.045	7.245	545.310	500.000	9.1
Decachlorobiphenyl	8.496	8.396	8.596	54.800	50.000	9.6
Tetrachloro-m-xylene	3.400	3.300	3.500	54.570	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071029.D **Time Analyzed:** 11:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.777	3.678	3.878	447.820	500.000	-10.4
Aroclor-1016-2	3.792	3.693	3.893	454.090	500.000	-9.2
Aroclor-1016-3	3.952	3.853	4.053	469.870	500.000	-6.0
Aroclor-1016-4	4.001	3.902	4.102	479.910	500.000	-4.0
Aroclor-1016-5	4.196	4.097	4.297	467.650	500.000	-6.5
Aroclor-1260-1	5.188	5.089	5.289	465.160	500.000	-7.0
Aroclor-1260-2	5.377	5.278	5.478	457.420	500.000	-8.5
Aroclor-1260-3	5.519	5.419	5.619	464.910	500.000	-7.0
Aroclor-1260-4	5.979	5.879	6.079	468.670	500.000	-6.3
Aroclor-1260-5	6.223	6.124	6.324	462.640	500.000	-7.5
Decachlorobiphenyl	7.533	7.433	7.633	54.060	50.000	8.1
Tetrachloro-m-xylene	2.777	2.678	2.878	49.880	50.000	-0.2