

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071145.D **Time Analyzed:** 17:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.479	4.380	4.580	519.230	500.000	3.8
Aroclor-1016-2	4.498	4.399	4.599	522.060	500.000	4.4
Aroclor-1016-3	4.555	4.456	4.656	533.570	500.000	6.7
Aroclor-1016-4	4.645	4.546	4.746	534.850	500.000	7.0
Aroclor-1016-5	4.929	4.831	5.031	531.200	500.000	6.2
Aroclor-1260-1	6.019	5.919	6.119	511.710	500.000	2.3
Aroclor-1260-2	6.274	6.175	6.375	517.130	500.000	3.4
Aroclor-1260-3	6.626	6.526	6.726	555.580	500.000	11.1
Aroclor-1260-4	6.840	6.740	6.940	559.920	500.000	12.0
Aroclor-1260-5	7.145	7.045	7.245	554.240	500.000	10.8
Decachlorobiphenyl	8.496	8.396	8.596	55.690	50.000	11.4
Tetrachloro-m-xylene	3.399	3.300	3.500	52.260	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071145.D **Time Analyzed:** 17:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.775	3.678	3.878	425.620	500.000	-14.9
Aroclor-1016-2	3.791	3.693	3.893	426.460	500.000	-14.7
Aroclor-1016-3	3.951	3.853	4.053	444.290	500.000	-11.1
Aroclor-1016-4	4.000	3.902	4.102	443.950	500.000	-11.2
Aroclor-1016-5	4.194	4.097	4.297	433.880	500.000	-13.2
Aroclor-1260-1	5.187	5.089	5.289	462.090	500.000	-7.6
Aroclor-1260-2	5.377	5.278	5.478	444.270	500.000	-11.1
Aroclor-1260-3	5.517	5.419	5.619	427.930	500.000	-14.4
Aroclor-1260-4	5.977	5.879	6.079	449.520	500.000	-10.1
Aroclor-1260-5	6.222	6.124	6.324	426.790	500.000	-14.6
Decachlorobiphenyl	7.531	7.433	7.633	52.860	50.000	5.7
Tetrachloro-m-xylene	2.776	2.678	2.878	45.740	50.000	-8.5