

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ARAM01

Lab Code: ACE

SDG NO.: Q3368

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 11/03/2025

11/03/2025

Continuing Calib Time: 11:39

Initial Calibration Time(s): 09:40

11:17

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
2,4-DCAA	7.28	7.28	7.18	7.38	0.00
2,4-D	8.41	8.41	8.31	8.51	0.00
2,4,5-TP(Silvex)	9.28	9.28	9.18	9.38	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ARAM01

Lab Code: ACE

SDG NO.: Q3368

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 11/03/2025 11/03/2025

Continuing Calib Time: 11:39

Initial Calibration Time(s): 09:40 11:17

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.61	7.62	7.52	7.72	0.01
2,4-D	8.86	8.86	8.76	8.96	0.01
2,4,5-TP(Silvex)	9.74	9.74	9.64	9.84	0.00