

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/04/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032267.D **Time Analyzed:** 09:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.277	9.180	9.380	732.160	712.500	2.8
2,4-D	8.410	8.314	8.514	716.400	705.000	1.6
2,4-DCAA	7.282	7.183	7.383	764.650	750.000	2.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/04/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032267.D **Time Analyzed:** 09:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.743	9.644	9.844	743.530	712.500	4.4
2,4-D	8.861	8.764	8.964	745.930	705.000	5.8
2,4-DCAA	7.615	7.517	7.717	790.290	750.000	5.4