

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/06/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032279.D **Time Analyzed:** 11:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.278	9.180	9.380	677.735	712.500	-4.9
2,4-D	8.408	8.314	8.514	680.697	705.000	-3.4
2,4-DCAA	7.280	7.183	7.383	716.394	750.000	-4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/06/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032279.D **Time Analyzed:** 11:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.739	9.644	9.844	652.874	712.500	-8.4
2,4-D	8.855	8.764	8.964	665.608	705.000	-5.6
2,4-DCAA	7.610	7.517	7.717	732.155	750.000	-2.4