

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/06/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032288.D **Time Analyzed:** 15:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.275	9.180	9.380	666.437	712.500	-6.5
2,4-D	8.407	8.314	8.514	666.769	705.000	-5.4
2,4-DCAA	7.280	7.183	7.383	700.130	750.000	-6.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/06/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032288.D **Time Analyzed:** 15:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.741	9.644	9.844	655.916	712.500	-7.9
2,4-D	8.858	8.764	8.964	660.388	705.000	-6.3
2,4-DCAA	7.613	7.517	7.717	726.288	750.000	-3.2