

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110625\
 Data File : PS032279.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 11:39
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/07/2025
 Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 01:05:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 04 04:59:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.280	7.610	3246.2E6	2669.1E6	716.394	732.155
Target Compounds							
1) T	Dalapon	2.656	2.598	4623.8E6	4358.1E6	644.099	626.706
2) T	3,5-DICHL...	6.447	6.577	4336.9E6	3860.1E6	670.696	642.396
3) T	4-Nitroph...	7.091	7.166	963.6E6	1469.4E6	717.790m	647.338
5) T	DICAMBA	7.465	7.803	12603.0E6	11351.6E6	668.022	657.208
6) T	MCPD	7.643	7.901	746.7E6	907.6E6	68.569	68.239
7) T	MCPA	7.792	8.146	905.4E6	1027.9E6	67.437	63.045
8) T	DICHLORPROP	8.172	8.516	2808.4E6	3073.9E6	631.283	644.545
9) T	2,4-D	8.408	8.855	3071.2E6	3491.6E6	680.697	665.608
10) T	Pentachlo...	8.706	9.353	45098.6E6	41004.7E6	718.827	698.515
11) T	2,4,5-TP ...	9.278	9.739	17628.9E6	17532.2E6	677.735	652.874
12) T	2,4,5-T	9.576	10.165	16807.8E6	16645.3E6	705.079	701.801
13) T	2,4-DB	10.151	10.730	1995.8E6	2607.5E6	757.387	691.403
14) T	DINOSEB	11.346	11.101	13530.2E6	11594.6E6	650.254	655.531
15) T	Picloram	11.178	12.216	17237.2E6	23581.6E6	812.095	634.219
16) T	DCPA	11.646	12.136	22818.3E6	21229.1E6	680.087	666.313m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110625\
Data File : PS032279.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 11:39
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 01:05:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
Quant Title : 8080.M
QLast Update : Tue Nov 04 04:59:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

