

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: Aramark Uniforms Date Collected: 11/06/25 Project: Monthly 2025 Date Received: 11/06/25 Client Sample ID: PIBLK-PS032287.D SDG No.: Q3368 Lab Sample ID: **TCLP** I.BLK-PS032287.D Matrix: Analytical Method: 8151A % Solid: 0

Sample Wt/Vol: 1000 mL Final Vol: 10000 uL Test: TCLP Herbicide

Prep Method: 8151A Prep Date:

CAS Number	r Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
94-75-7	2,4-D	0.92	U	1	0.92	2.00	ug/L	11/06/25	pS110625
93-72-1	2,4,5-TP (Silvex)	0.78	U	1	0.78	2.00	ug/L	11/06/25	pS110625
SURROGATES									
19719-28-9	2,4-DCAA	520			61 - 136	104%	SPK: 500)	

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit