

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/04/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090937.D **Time Analyzed:** 09:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.063	8.964	9.164	45.240	50.000	-9.5
Endrin	6.567	6.468	6.668	47.020	50.000	-6.0
gamma-BHC (Lindane)	4.324	4.225	4.425	45.360	50.000	-9.3
Heptachlor	4.922	4.823	5.023	56.000	50.000	12.0
Heptachlor epoxide	5.683	5.584	5.784	46.040	50.000	-7.9
Methoxychlor	7.487	7.388	7.588	53.360	50.000	6.7
Tetrachloro-m-xylene	3.544	3.445	3.645	43.890	50.000	-12.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/04/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090937.D **Time Analyzed:** 09:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.059	7.962	8.162	55.160	50.000	10.3
Endrin	5.779	5.681	5.881	49.420	50.000	-1.2
gamma-BHC (Lindane)	3.721	3.623	3.823	47.830	50.000	-4.3
Heptachlor	4.073	3.975	4.175	50.260	50.000	0.5
Heptachlor epoxide	4.862	4.764	4.964	47.000	50.000	-6.0
Methoxychlor	6.742	6.645	6.845	53.310	50.000	6.6
Tetrachloro-m-xylene	2.874	2.775	2.975	46.870	50.000	-6.3