

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ARAM01
Lab Code: ACE
SDG NO.: Q3368
GC Column: ZB-MR1
ID: 0.32 (mm)
 Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No.: CCAL02
Date Analyzed: 11/04/2025
Lab Sample No.: PSTDCCC050
Data File : PD090949.D
Time Analyzed: 13:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.074	8.964	9.164	43.990	50.000	-12.0
Endrin	6.575	6.468	6.668	49.110	50.000	-1.8
gamma-BHC (Lindane)	4.332	4.225	4.425	54.710	50.000	9.4
Heptachlor	4.929	4.823	5.023	66.730	50.000	33.5
Heptachlor epoxide	5.690	5.584	5.784	49.950	50.000	-0.1
Methoxychlor	7.495	7.388	7.588	50.020	50.000	0.0
Tetrachloro-m-xylene	3.551	3.445	3.645	51.620	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/04/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090949.D **Time Analyzed:** 13:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.062	7.962	8.162	37.880	50.000	-24.2
Endrin	5.780	5.681	5.881	47.270	50.000	-5.5
gamma-BHC (Lindane)	3.722	3.623	3.823	53.140	50.000	6.3
Heptachlor	4.074	3.975	4.175	55.600	50.000	11.2
Heptachlor epoxide	4.864	4.764	4.964	51.470	50.000	2.9
Methoxychlor	6.745	6.645	6.845	43.330	50.000	-13.3
Tetrachloro-m-xylene	2.874	2.775	2.975	53.000	50.000	6.0