

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/04/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090957.D **Time Analyzed:** 16:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.064	8.964	9.164	48.210	50.000	-3.6
Endrin	6.567	6.468	6.668	54.530	50.000	9.1
gamma-BHC (Lindane)	4.324	4.225	4.425	55.080	50.000	10.2
Heptachlor	4.922	4.823	5.023	66.530	50.000	33.1
Heptachlor epoxide	5.683	5.584	5.784	54.960	50.000	9.9
Methoxychlor	7.487	7.388	7.588	52.810	50.000	5.6
Tetrachloro-m-xylene	3.544	3.445	3.645	52.920	50.000	5.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/04/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090957.D **Time Analyzed:** 16:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.060	7.962	8.162	46.000	50.000	-8.0
Endrin	5.778	5.681	5.881	53.220	50.000	6.4
gamma-BHC (Lindane)	3.722	3.623	3.823	53.100	50.000	6.2
Heptachlor	4.073	3.975	4.175	55.570	50.000	11.1
Heptachlor epoxide	4.862	4.764	4.964	51.570	50.000	3.1
Methoxychlor	6.742	6.645	6.845	51.400	50.000	2.8
Tetrachloro-m-xylene	2.874	2.775	2.975	52.660	50.000	5.3