

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/05/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090961.D **Time Analyzed:** 09:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.064	8.964	9.164	48.070	50.000	-3.9
Endrin	6.567	6.468	6.668	56.110	50.000	12.2
gamma-BHC (Lindane)	4.324	4.225	4.425	55.220	50.000	10.4
Heptachlor	4.922	4.823	5.023	67.540	50.000	35.1
Heptachlor epoxide	5.683	5.584	5.784	57.550	50.000	15.1
Methoxychlor	7.487	7.388	7.588	57.330	50.000	14.7
Tetrachloro-m-xylene	3.544	3.445	3.645	52.690	50.000	5.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/05/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090961.D **Time Analyzed:** 09:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.059	7.962	8.162	50.830	50.000	1.7
Endrin	5.778	5.681	5.881	54.460	50.000	8.9
gamma-BHC (Lindane)	3.722	3.623	3.823	54.740	50.000	9.5
Heptachlor	4.073	3.975	4.175	57.000	50.000	14.0
Heptachlor epoxide	4.862	4.764	4.964	53.510	50.000	7.0
Methoxychlor	6.743	6.645	6.845	55.570	50.000	11.1
Tetrachloro-m-xylene	2.874	2.775	2.975	54.140	50.000	8.3