

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ARAM01
Lab Code: ACE
SDG NO.: Q3368
GC Column: ZB-MR1
ID: 0.32 (mm)
 Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No.: CCAL05
Date Analyzed: 11/05/2025
Lab Sample No.: PSTDCCC050
Data File : PD090969.D
Time Analyzed: 11:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.065	8.964	9.164	48.090	50.000	-3.8
Endrin	6.567	6.468	6.668	55.650	50.000	11.3
gamma-BHC (Lindane)	4.324	4.225	4.425	55.720	50.000	11.4
Heptachlor	4.922	4.823	5.023	67.980	50.000	36.0
Heptachlor epoxide	5.683	5.584	5.784	56.010	50.000	12.0
Methoxychlor	7.488	7.388	7.588	58.730	50.000	17.5
Tetrachloro-m-xylene	3.544	3.445	3.645	53.450	50.000	6.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ARAM01
Lab Code: ACE **SDG NO.:** Q3368
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/05/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090969.D **Time Analyzed:** 11:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.059	7.962	8.162	50.210	50.000	0.4
Endrin	5.779	5.681	5.881	53.940	50.000	7.9
gamma-BHC (Lindane)	3.722	3.623	3.823	55.090	50.000	10.2
Heptachlor	4.073	3.975	4.175	57.640	50.000	15.3
Heptachlor epoxide	4.862	4.764	4.964	53.060	50.000	6.1
Methoxychlor	6.743	6.645	6.845	54.820	50.000	9.6
Tetrachloro-m-xylene	2.875	2.775	2.975	54.620	50.000	9.2