



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

A
B
C
D

Hit Summary Sheet
SW-846

SDG No.: Q3368
Client: Aramark Uniforms

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :	PRESS-SLUDGE							
Q3368-07	PRESS-SLUDGE	TCLP	3+4-Methylphenols	110.000	11		100	ug/L
			Total Svoc :		110.00			
			Total Concentration:		110.00			



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Aramark Uniforms		Date Collected:	10/22/25
Project:	Monthly 2025		Date Received:	10/22/25
Client Sample ID:	PB170172TB		SDG No.:	Q3368
Lab Sample ID:	PB170172TB		Matrix:	TCLP
Analytical Method:	8270E	Level : LOW	% Solid:	0
Sample Wt/Vol:	100 mL	Final Vol: 1000 uL	Test:	TCLP BNA Group1
Prep Method :	SW3541	Prep Date: 10/22/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
110-86-1	Pyridine	12.8	U	1	12.8	50.0	ug/L	10/22/25 18:36	PB170216
106-46-7	1,4-Dichlorobenzene	5.30	U	1	5.30	50.0	ug/L	10/22/25 18:36	PB170216
95-48-7	2-Methylphenol	11.2	U	1	11.2	50.0	ug/L	10/22/25 18:36	PB170216
65794-96-9	3+4-Methylphenols	11.0	U	1	11.0	100	ug/L	10/22/25 18:36	PB170216
67-72-1	Hexachloroethane	6.50	U	1	6.50	50.0	ug/L	10/22/25 18:36	PB170216
98-95-3	Nitrobenzene	7.60	U	1	7.60	50.0	ug/L	10/22/25 18:36	PB170216
88-06-2	2,4,6-Trichlorophenol	5.10	U	1	5.10	50.0	ug/L	10/22/25 18:36	PB170216
95-95-4	2,4,5-Trichlorophenol	6.20	U	1	6.20	50.0	ug/L	10/22/25 18:36	PB170216
121-14-2	2,4-Dinitrotoluene	12.2	U	1	12.2	50.0	ug/L	10/22/25 18:36	PB170216
118-74-1	Hexachlorobenzene	5.20	U	1	5.20	50.0	ug/L	10/22/25 18:36	PB170216
87-86-5	Pentachlorophenol	15.8	U	1	15.8	100	ug/L	10/22/25 18:36	PB170216
SURROGATES									
367-12-4	2-Fluorophenol	146			23 - 138	98%	SPK: 150		
13127-88-3	Phenol-d6	136			10 - 134	91%	SPK: 150		
4165-60-0	Nitrobenzene-d5	80.4			67 - 132	80%	SPK: 100		
321-60-8	2-Fluorobiphenyl	77.1			52 - 132	77%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	129			44 - 137	86%	SPK: 150		
1718-51-0	Terphenyl-d14	89.0			42 - 152	89%	SPK: 100		
INTERNAL STANDARDS									
		Area Count							
3855-82-1	1,4-Dichlorobenzene-d4	158000							
1146-65-2	Naphthalene-d8	655000							
15067-26-2	Acenaphthene-d10	478000							
1517-22-2	Phenanthrene-d10	1020000							
1719-03-5	Chrysene-d12	810000							
1520-96-3	Perylene-d12	659000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Aramark Uniforms		Date Collected:	10/16/25
Project:	Monthly 2025		Date Received:	10/16/25
Client Sample ID:	OUTSIDE-DUMPSTER		SDG No.:	Q3368
Lab Sample ID:	Q3368-06		Matrix:	TCLP
Analytical Method:	8270E	Level : LOW	% Solid:	0
Sample Wt/Vol:	100 mL	Final Vol: 1000 uL	Test:	TCLP BNA Group1
Prep Method :	SW3541	Prep Date: 10/22/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
110-86-1	Pyridine	12.8	U	1	12.8	50.0	ug/L	10/22/25 19:17	PB170216
106-46-7	1,4-Dichlorobenzene	5.30	U	1	5.30	50.0	ug/L	10/22/25 19:17	PB170216
95-48-7	2-Methylphenol	11.2	U	1	11.2	50.0	ug/L	10/22/25 19:17	PB170216
65794-96-9	3+4-Methylphenols	11.0	U	1	11.0	100	ug/L	10/22/25 19:17	PB170216
67-72-1	Hexachloroethane	6.50	U	1	6.50	50.0	ug/L	10/22/25 19:17	PB170216
98-95-3	Nitrobenzene	7.60	U	1	7.60	50.0	ug/L	10/22/25 19:17	PB170216
88-06-2	2,4,6-Trichlorophenol	5.10	U	1	5.10	50.0	ug/L	10/22/25 19:17	PB170216
95-95-4	2,4,5-Trichlorophenol	6.20	U	1	6.20	50.0	ug/L	10/22/25 19:17	PB170216
121-14-2	2,4-Dinitrotoluene	12.2	U	1	12.2	50.0	ug/L	10/22/25 19:17	PB170216
118-74-1	Hexachlorobenzene	5.20	U	1	5.20	50.0	ug/L	10/22/25 19:17	PB170216
87-86-5	Pentachlorophenol	15.8	U	1	15.8	100	ug/L	10/22/25 19:17	PB170216
SURROGATES									
367-12-4	2-Fluorophenol	130			23 - 138	87%	SPK: 150		
13127-88-3	Phenol-d6	122			10 - 134	81%	SPK: 150		
4165-60-0	Nitrobenzene-d5	74.6			67 - 132	75%	SPK: 100		
321-60-8	2-Fluorobiphenyl	68.9			52 - 132	69%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	129			44 - 137	86%	SPK: 150		
1718-51-0	Terphenyl-d14	55.3			42 - 152	55%	SPK: 100		
INTERNAL STANDARDS									
		Area Count							
3855-82-1	1,4-Dichlorobenzene-d4	151000							
1146-65-2	Naphthalene-d8	649000							
15067-26-2	Acenaphthene-d10	478000							
1517-22-2	Phenanthrene-d10	1050000							
1719-03-5	Chrysene-d12	1090000							
1520-96-3	Perylene-d12	970000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Aramark Uniforms		Date Collected:	10/16/25
Project:	Monthly 2025		Date Received:	10/16/25
Client Sample ID:	PRESS-SLUDGE		SDG No.:	Q3368
Lab Sample ID:	Q3368-07		Matrix:	TCLP
Analytical Method:	8270E	Level : LOW	% Solid:	0
Sample Wt/Vol:	100 mL	Final Vol: 1000 uL	Test:	TCLP BNA Group1
Prep Method :	SW3541	Prep Date: 10/22/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
110-86-1	Pyridine	12.8	U	1	12.8	50.0	ug/L	10/23/25 02:46	PB170216
106-46-7	1,4-Dichlorobenzene	5.30	U	1	5.30	50.0	ug/L	10/23/25 02:46	PB170216
95-48-7	2-Methylphenol	11.2	U	1	11.2	50.0	ug/L	10/23/25 02:46	PB170216
65794-96-9	3+4-Methylphenols	110		1	11.0	100	ug/L	10/23/25 02:46	PB170216
67-72-1	Hexachloroethane	6.50	U	1	6.50	50.0	ug/L	10/23/25 02:46	PB170216
98-95-3	Nitrobenzene	7.60	U	1	7.60	50.0	ug/L	10/23/25 02:46	PB170216
88-06-2	2,4,6-Trichlorophenol	5.10	U	1	5.10	50.0	ug/L	10/23/25 02:46	PB170216
95-95-4	2,4,5-Trichlorophenol	6.20	U	1	6.20	50.0	ug/L	10/23/25 02:46	PB170216
121-14-2	2,4-Dinitrotoluene	12.2	U	1	12.2	50.0	ug/L	10/23/25 02:46	PB170216
118-74-1	Hexachlorobenzene	5.20	U	1	5.20	50.0	ug/L	10/23/25 02:46	PB170216
87-86-5	Pentachlorophenol	15.8	U	1	15.8	100	ug/L	10/23/25 02:46	PB170216
SURROGATES									
367-12-4	2-Fluorophenol	129			23 - 138	86%	SPK: 150		
13127-88-3	Phenol-d6	119			10 - 134	79%	SPK: 150		
4165-60-0	Nitrobenzene-d5	75.7			67 - 132	76%	SPK: 100		
321-60-8	2-Fluorobiphenyl	73.5			52 - 132	73%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	94.4			44 - 137	63%	SPK: 150		
1718-51-0	Terphenyl-d14	58.0			42 - 152	58%	SPK: 100		
INTERNAL STANDARDS									
		Area Count							
3855-82-1	1,4-Dichlorobenzene-d4	175000							
1146-65-2	Naphthalene-d8	752000							
15067-26-2	Acenaphthene-d10	463000							
1517-22-2	Phenanthrene-d10	721000							
1719-03-5	Chrysene-d12	673000							
1520-96-3	Perylene-d12	811000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q3368	OrderDate:	10/16/2025 2:05:00 PM
Client:	Aramark Uniforms	Project:	Monthly 2025
Contact:	Tom Dikos	Location:	D41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3368-06	OUTSIDE-DUMPSTER	TCLP	TCLP BNA Group1	8270E	10/16/25	10/22/25	10/22/25	10/16/25
Q3368-07	PRESS-SLUDGE	TCLP	TCLP BNA Group1	8270E	10/16/25	10/22/25	10/23/25	10/16/25