

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3368

Analytical Method: SW8270E

Client: Aramark Uniforms

DataFile: BP026006.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3368-06MS		Client Sample ID: OUTSIDE-DUMPSTERMS									
Pyridine	500	0	370	ug/L	74				10	109	
1,4-Dichlorobenzene	500	0	400	ug/L	80				55	125	
2-Methylphenol	500	0	530	ug/L	106				60	131	
3+4-Methylphenols	500	0	520	ug/L	104				54	136	
Hexachloroethane	500	0	390	ug/L	78				19	146	
Nitrobenzene	500	0	470	ug/L	94				62	112	
2,4,6-Trichlorophenol	500	0	510	ug/L	102				78	112	
2,4,5-Trichlorophenol	500	0	530	ug/L	106				71	111	
2,4-Dinitrotoluene	500	0	530	ug/L	106				74	137	
Hexachlorobenzene	500	0	460	ug/L	92				72	115	
Pentachlorophenol	1000	0	1100	ug/L	110				52	162	