

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3368

Analytical Method: SW8270E

Client: Aramark Uniforms

DataFile: BP026007.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3368-06MSD	Client Sample ID:	OUTSIDE-DUMPSTERMSD								
Pyridine	500	0	340	ug/L	68	8			10	109	20
1,4-Dichlorobenzene	500	0	390	ug/L	78	3			55	125	20
2-Methylphenol	500	0	510	ug/L	102	4			60	131	20
3+4-Methylphenols	500	0	490	ug/L	98	6			54	136	20
Hexachloroethane	500	0	390	ug/L	78	0			19	146	20
Nitrobenzene	500	0	460	ug/L	92	2			62	112	20
2,4,6-Trichlorophenol	500	0	500	ug/L	100	2			78	112	20
2,4,5-Trichlorophenol	500	0	500	ug/L	100	6			71	111	20
2,4-Dinitrotoluene	500	0	490	ug/L	98	8			74	137	20
Hexachlorobenzene	500	0	470	ug/L	94	2			72	115	20
Pentachlorophenol	1000	0	980	ug/L	98	12			52	162	20