

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3368

Analytical Method: 8270E

Client: Aramark Uniforms

DataFile: BP026003.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170216BS	Pyridine	50	35.7	ug/L	71				29	97	
	1,4-Dichlorobenzene	50	42.6	ug/L	85				76	103	
	2-Methylphenol	50	50.9	ug/L	102				69	109	
	3+4-Methylphenols	50	49.5	ug/L	99				67	106	
	Hexachloroethane	50	42.5	ug/L	85				76	118	
	Nitrobenzene	50	45.3	ug/L	91				58	106	
	2,4,6-Trichlorophenol	50	45.6	ug/L	91				61	110	
	2,4,5-Trichlorophenol	50	47.0	ug/L	94				70	106	
	2,4-Dinitrotoluene	50	47.2	ug/L	94				60	115	
	Hexachlorobenzene	50	42.2	ug/L	84				73	106	
	Pentachlorophenol	100	94.1	ug/L	94				47	114	