

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3368

Client ID : SSTDCCC040

Date Analyzed: 10/22/2025

Lab File ID: BP026001.D

Time Analyzed: 16:33

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	168370	7.696	761230	10.47	536314	14.32
UPPER LIMIT	336740	8.196	1522460	10.966	1072630	14.819
LOWER LIMIT	84185	7.196	380615	9.966	268157	13.819
EPA SAMPLE NO.						
01 PB170216BL	155512	7.70	659805	10.48	488964	14.33
02 PB170216BS	157158	7.69	711577	10.46	515285	14.32
03 PB170172TB	158418	7.70	655423	10.47	478330	14.33
04 OUTSIDE-DUMPSTER	151222	7.70	648521	10.47	478185	14.32
05 OUTSIDE-DUMPSTERMS	151873	7.70	669626	10.45	496531	14.32
06 OUTSIDE-DUMPSTERMSD	161283	7.70	704129	10.45	487784	14.32
07 PRESS-SLUDGE	175466	7.70	751992	10.46	463470	14.31

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.