

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ARAM01
Lab Code: ACE SDG NO.: Q3368
Lab File ID: VX048280.D Date Analyzed: 10/22/2025
Instrument ID: MSVOA_X Time Analyzed: 08:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	112883	5.53	171616	6.74	151325	10.04
UPPER LIMIT	225766	6.031	343232	7.239	302650	10.537
LOWER LIMIT	56441.5	5.031	85808	6.239	75662.5	9.537
EPA SAMPLE NO.						
OUTSIDE-DUMPSTER	101261	5.54	194156	6.75	190163	10.04
PRESS-SLUDGE	96420	5.54	193055	6.75	220019	10.04
VX1022WBL01	97887	5.54	185129	6.75	199062	10.04
VX1022WBS01	115938	5.54	185153	6.75	172449	10.04
VX1022WBSD01	120104	5.53	195961	6.75	180624	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.