

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ARAM01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3368</u>
Lab File ID:	<u>VX048280.D</u>	Date Analyzed:	<u>10/22/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>08:37</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	76595	12				
UPPER LIMIT	153190	12.5				
LOWER LIMIT	38297.5	11.5				
EPA SAMPLE NO.						
OUTSIDE-DUMPSTER	98385	12.01				
PRESS-SLUDGE	121001	12.01				
VX1022WBL01	99844	12.01				
VX1022WBS01	91086	12.01				
VX1022WBSD01	92297	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.