

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/21/25 12:03		Jaswal	OK
2	S1	S1	CAL2	10/21/25 12:08		Jaswal	OK
3	S2	S2	CAL3	10/21/25 12:12		Jaswal	OK
4	S3	S3	CAL4	10/21/25 12:16		Jaswal	OK
5	S4	S4	CAL5	10/21/25 12:21		Jaswal	OK
6	S5	S5	CAL6	10/21/25 12:25		Jaswal	OK
7	ICV01	ICV01	ICV	10/21/25 12:40		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/21/25 12:44		Jaswal	OK
9	ICB01	ICB01	ICB	10/21/25 12:50		Jaswal	OK
10	CRI01	CRI01	CRDL	10/21/25 12:54		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/21/25 12:59		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/21/25 13:03		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/21/25 13:07		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/21/25 13:11		Jaswal	OK
15	CCV01	CCV01	CCV	10/21/25 13:45		Jaswal	OK
16	CCB01	CCB01	CCB	10/21/25 13:54		Jaswal	OK
17	LR-1	LR-1	HIGH STD	10/21/25 13:57		Jaswal	OK
18	LR-2	LR-2	HIGH STD	10/21/25 14:02		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	PB170187BL	PB170187BL	MB	10/21/25 14:11		Jaswal	OK
20	PB170187BS	PB170187BS	LCS	10/21/25 14:16		Jaswal	OK
21	PB170183BL	PB170183BL	MB	10/21/25 14:25		Jaswal	OK
22	PB170183BS	PB170183BS	LCS	10/21/25 14:29		Jaswal	OK
23	LR-3	LR-3	HIGH STD	10/21/25 14:33		Jaswal	OK
24	Q3381-02	VNJ-222	SAM	10/21/25 14:39		Jaswal	OK
25	Q3381-04	VNJ-245	SAM	10/21/25 14:43		Jaswal	OK
26	Q3381-06	RBR200059	SAM	10/21/25 14:47		Jaswal	OK
27	CCV02	CCV02	CCV	10/21/25 15:03		Jaswal	OK
28	CCB02	CCB02	CCB	10/21/25 15:07		Jaswal	OK
29	Q3381-08	VNJ-255	SAM	10/21/25 15:11		Jaswal	OK
30	Q3369-04	MH-22	SAM	10/21/25 15:16		Jaswal	OK
31	Q3369-08	MH-21	SAM	10/21/25 15:20		Jaswal	OK
32	Q3374-04	TP-4-WC	SAM	10/21/25 15:25		Jaswal	OK
33	Q3374-08	TP-5-WC	SAM	10/21/25 15:29		Jaswal	OK
34	Q3388-02	461	SAM	10/21/25 15:34		Jaswal	OK
35	Q3390-02	4174	SAM	10/21/25 15:38		Jaswal	OK
36	Q3392-04	MH-20	SAM	10/21/25 15:42		Jaswal	OK
37	Q3392-08	MH-19	SAM	10/21/25 15:47		Jaswal	OK
38	Q3392-08DUP	MH-19DUP	DUP	10/21/25 15:51		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

39	CCV03	CCV03	CCV	10/21/25 15:56		Jaswal	OK
40	CCB03	CCB03	CCB	10/21/25 16:00		Jaswal	OK
41	Q3392-08L	MH-19L	SD	10/21/25 16:04		Jaswal	OK
42	Q3392-08MS	MH-19MS	MS	10/21/25 16:09		Jaswal	OK
43	Q3392-08MSD	MH-19MSD	MSD	10/21/25 16:13		Jaswal	OK
44	Q3392-08A	MH-19A	PS	10/21/25 16:17	0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
45	PB170147TB	PB170147TB	MB	10/21/25 16:21		Jaswal	OK
46	Q3411-01	Q4-WASTEWATER	SAM	10/21/25 16:26		Jaswal	OK
47	Q3401-01	41-CONCRETE	SAM	10/21/25 16:30		Jaswal	OK
48	Q3402-01	41-LA	SAM	10/21/25 16:34		Jaswal	OK
49	Q3402-04	41-RA	SAM	10/21/25 16:40		Jaswal	OK
50	Q3402-07	41-LB	SAM	10/21/25 16:44		Jaswal	OK
51	CCV04	CCV04	CCV	10/21/25 16:48		Jaswal	OK
52	CCB04	CCB04	CCB	10/21/25 16:52		Jaswal	OK
53	Q3402-10	41-RB	SAM	10/21/25 16:56		Jaswal	OK
54	Q3403-01	MH-18	SAM	10/21/25 17:00		Jaswal	OK
55	Q3397-01	TP-7-WC	SAM	10/21/25 17:04		Jaswal	OK
56	Q3397-01DUP	TP-7-WCDUP	DUP	10/21/25 17:08		Jaswal	OK
57	Q3397-01L	TP-7-WCL	SD	10/21/25 17:12		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

58	Q3397-01MS	TP-7-WCMS	MS	10/21/25 17:16		Jaswal	OK
59	Q3397-01MSD	TP-7-WCMSD	MSD	10/21/25 17:20		Jaswal	OK
60	Q3397-01A	TP-7-WCA	PS	10/21/25 17:24	0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
61	PB170192BL	PB170192BL	MB	10/21/25 17:28		Jaswal	OK
62	PB170192BS	PB170192BS	LCS	10/21/25 17:33		Jaswal	OK
63	CCV05	CCV05	CCV	10/21/25 17:37		Jaswal	OK
64	CCB05	CCB05	CCB	10/21/25 17:41		Jaswal	OK
65	Q3365-01	MW25	SAM	10/21/25 17:45	Na is high, Need dilution	Jaswal	Dilution
66	Q3365-02	MW2D	SAM	10/21/25 17:49	Na is high, Need dilution	Jaswal	Dilution
67	Q3365-03	SUPPLY	SAM	10/21/25 17:53		Jaswal	OK
68	Q3365-04	17 HARVEY	SAM	10/21/25 17:58		Jaswal	OK
69	Q3365-05	19 HARVEY	SAM	10/21/25 18:02		Jaswal	OK
70	Q3370-01	WATER-TREATMENT	SAM	10/21/25 18:06		Jaswal	OK
71	Q3370-01DUP	WATER-TREATMENT	DUP	10/21/25 18:11		Jaswal	OK
72	Q3370-01L	WATER-TREATMENT	SD	10/21/25 18:15		Jaswal	OK
73	Q3370-01MS	WATER-TREATMENT	MS	10/21/25 18:19		Jaswal	OK
74	Q3370-01MSD	WATER-TREATMENT	MSD	10/21/25 18:23		Jaswal	OK
75	CCV06	CCV06	CCV	10/21/25 18:27		Jaswal	OK
76	CCB06	CCB06	CCB	10/21/25 18:31		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

77	Q3370-01A	WATER-TREATMENT	PS	10/21/25 18:36	0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
78	Q3376-07	PR132-FB01-202510	SAM	10/21/25 18:40		Jaswal	OK
79	Q3400-01	FRAC-TANK-0760450	SAM	10/21/25 18:44		Jaswal	OK
80	Q3398-01	A748	SAM	10/21/25 18:48	BAD Injection	Jaswal	Not Ok
81	Q3399-01	COMP-ROLLOFF	SAM	10/21/25 19:01		Jaswal	OK
82	Q3405-01	PAD-10-20-25-A	SAM	10/21/25 19:05		Jaswal	OK
83	Q3405-02	PAD-10-20-25-B	SAM	10/21/25 19:09		Jaswal	OK
84	Q3405-03	PAD-10-20-25-C	SAM	10/21/25 19:14		Jaswal	OK
85	CCV07	CCV07	CCV	10/21/25 19:18		Jaswal	OK
86	CCB07	CCB07	CCB	10/21/25 19:22		Jaswal	OK
87	Q3405-04	SW-10-20-25	SAM	10/21/25 19:26		Jaswal	OK
88	Q3409-01	EXCAVATION-AREA-	SAM	10/21/25 19:31		Jaswal	OK
89	Q3409-02	EXCAVATION-AREA-	SAM	10/21/25 19:35		Jaswal	OK
90	Q3412-01	TP-6	SAM	10/21/25 19:39	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
91	Q3412-01DUP	TP-6DUP	DUP	10/21/25 19:43	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
92	Q3412-01L	TP-6L	SD	10/21/25 19:47	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
93	Q3412-01MS	TP-6MS	MS	10/21/25 19:51	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
94	Q3412-01MSD	TP-6MSD	MSD	10/21/25 19:55	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

95	Q3412-01A	TP-6A	PS	10/21/25 19:59	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
96	CCV08	CCV08	CCV	10/21/25 20:03		Jaswal	OK
97	CCB08	CCB08	CCB	10/21/25 20:07		Jaswal	OK