

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3370

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	784	800	98	90 - 110	P	10/21/2025	12:40	LB137613
	Copper	976	1000	98	90 - 110	P	10/21/2025	12:40	LB137613
	Nickel	1910	2000	95	90 - 110	P	10/21/2025	12:40	LB137613
	Zinc	1920	2000	96	90 - 110	P	10/21/2025	12:40	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3370

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	9.98	10.0	100	80 - 120	P	10/21/2025	12:44	LB137613
	Copper	21.5	20.0	107	80 - 120	P	10/21/2025	12:44	LB137613
	Nickel	38.9	40.0	97	80 - 120	P	10/21/2025	12:44	LB137613
	Zinc	41.7	40.0	104	80 - 120	P	10/21/2025	12:44	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3370

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	976	1000	98	90 - 110	P	10/21/2025	13:45	LB137613
	Copper	1220	1250	98	90 - 110	P	10/21/2025	13:45	LB137613
	Nickel	2410	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
	Zinc	2430	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
CCV02	Chromium	934	1000	93	90 - 110	P	10/21/2025	15:03	LB137613
	Copper	1210	1250	96	90 - 110	P	10/21/2025	15:03	LB137613
	Nickel	2330	2500	93	90 - 110	P	10/21/2025	15:03	LB137613
	Zinc	2370	2500	95	90 - 110	P	10/21/2025	15:03	LB137613
CCV03	Chromium	941	1000	94	90 - 110	P	10/21/2025	15:56	LB137613
	Copper	1210	1250	96	90 - 110	P	10/21/2025	15:56	LB137613
	Nickel	2350	2500	94	90 - 110	P	10/21/2025	15:56	LB137613
	Zinc	2390	2500	96	90 - 110	P	10/21/2025	15:56	LB137613
CCV04	Chromium	934	1000	93	90 - 110	P	10/21/2025	16:48	LB137613
	Copper	1190	1250	96	90 - 110	P	10/21/2025	16:48	LB137613
	Nickel	2330	2500	93	90 - 110	P	10/21/2025	16:48	LB137613
	Zinc	2400	2500	96	90 - 110	P	10/21/2025	16:48	LB137613
CCV05	Chromium	964	1000	96	90 - 110	P	10/21/2025	17:37	LB137613
	Copper	1220	1250	98	90 - 110	P	10/21/2025	17:37	LB137613
	Nickel	2390	2500	96	90 - 110	P	10/21/2025	17:37	LB137613
	Zinc	2430	2500	97	90 - 110	P	10/21/2025	17:37	LB137613
CCV06	Chromium	905	1000	90	90 - 110	P	10/21/2025	18:27	LB137613
	Copper	1180	1250	95	90 - 110	P	10/21/2025	18:27	LB137613
	Nickel	2300	2500	92	90 - 110	P	10/21/2025	18:27	LB137613
	Zinc	2320	2500	93	90 - 110	P	10/21/2025	18:27	LB137613
CCV07	Chromium	945	1000	94	90 - 110	P	10/21/2025	19:18	LB137613
	Copper	1200	1250	96	90 - 110	P	10/21/2025	19:18	LB137613
	Nickel	2350	2500	94	90 - 110	P	10/21/2025	19:18	LB137613
	Zinc	2370	2500	95	90 - 110	P	10/21/2025	19:18	LB137613
CCV08	Chromium	915	1000	92	90 - 110	P	10/21/2025	20:03	LB137613
	Copper	1210	1250	97	90 - 110	P	10/21/2025	20:03	LB137613
	Nickel	2350	2500	94	90 - 110	P	10/21/2025	20:03	LB137613
	Zinc	2340	2500	94	90 - 110	P	10/21/2025	20:03	LB137613