



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 2b -**

**CRDL STANDARD FOR AA & ICP**

**Client:** VERINA CONSULTING GROUP, LLC

**SDG No.:** Q3370

**Contract:** VERI01

**Lab Code:** ACE

**Initial Calibration Source:** \_\_\_\_\_

**Continuing Calibration Source:** \_\_\_\_\_

| Sample ID | Analyte  | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|--------------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CRI01     | Chromium | 9.94           | 10.0               | 99            | 65 - 135                  | P | 10/21/2025       | 12:54            | LB137613      |
|           | Copper   | 21.6           | 20.0               | 108           | 65 - 135                  | P | 10/21/2025       | 12:54            | LB137613      |
|           | Nickel   | 38.6           | 40.0               | 96            | 65 - 135                  | P | 10/21/2025       | 12:54            | LB137613      |
|           | Zinc     | 41.9           | 40.0               | 105           | 65 - 135                  | P | 10/21/2025       | 12:54            | LB137613      |