



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3370

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Chromium	2.12	+/-5	U	10.0	P	10/21/2025	12:50	LB137613
	Copper	4.60	+/-10	U	20.0	P	10/21/2025	12:50	LB137613
	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	12:50	LB137613
	Zinc	16.7	+/-20	U	40.0	P	10/21/2025	12:50	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3370

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Chromium	2.12	+/-5	U	10.0	P	10/21/2025	13:54	LB137613
	Copper	4.60	+/-10	U	20.0	P	10/21/2025	13:54	LB137613
	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	13:54	LB137613
	Zinc	16.7	+/-20	U	40.0	P	10/21/2025	13:54	LB137613
CCB02	Chromium	2.12	+/-5	U	10.0	P	10/21/2025	15:07	LB137613
	Copper	4.60	+/-10	U	20.0	P	10/21/2025	15:07	LB137613
	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	15:07	LB137613
	Zinc	16.7	+/-20	U	40.0	P	10/21/2025	15:07	LB137613
CCB03	Chromium	2.83	+/-5	J	10.0	P	10/21/2025	16:00	LB137613
	Copper	4.60	+/-10	U	20.0	P	10/21/2025	16:00	LB137613
	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	16:00	LB137613
	Zinc	16.7	+/-20	U	40.0	P	10/21/2025	16:00	LB137613
CCB04	Chromium	2.12	+/-5	U	10.0	P	10/21/2025	16:52	LB137613
	Copper	4.60	+/-10	U	20.0	P	10/21/2025	16:52	LB137613
	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	16:52	LB137613
	Zinc	16.7	+/-20	U	40.0	P	10/21/2025	16:52	LB137613
CCB05	Chromium	2.12	+/-5	U	10.0	P	10/21/2025	17:41	LB137613
	Copper	4.60	+/-10	U	20.0	P	10/21/2025	17:41	LB137613
	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	17:41	LB137613
	Zinc	16.7	+/-20	U	40.0	P	10/21/2025	17:41	LB137613
CCB06	Chromium	2.12	+/-5	U	10.0	P	10/21/2025	18:31	LB137613
	Copper	4.60	+/-10	U	20.0	P	10/21/2025	18:31	LB137613
	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	18:31	LB137613
	Zinc	16.7	+/-20	U	40.0	P	10/21/2025	18:31	LB137613
CCB07	Chromium	2.12	+/-5	U	10.0	P	10/21/2025	19:22	LB137613
	Copper	4.60	+/-10	U	20.0	P	10/21/2025	19:22	LB137613
	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	19:22	LB137613
	Zinc	16.7	+/-20	U	40.0	P	10/21/2025	19:22	LB137613
CCB08	Chromium	2.12	+/-5	U	10.0	P	10/21/2025	20:07	LB137613
	Copper	4.60	+/-10	U	20.0	P	10/21/2025	20:07	LB137613
	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	20:07	LB137613
	Zinc	16.7	+/-20	U	40.0	P	10/21/2025	20:07	LB137613